

7471B


LB 96708

INSTRUMENT ID: CV2

Linear

 μ Abs.:

48968



A= 0.0000e+000

B= 2.0493e-004

C= -1.3912e-002

Rho= 0.9999471

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	13	0.000	13				
0.2	0.200	0.231	0.031	1197	0.0 %	1197				
2.5	2.500	2.523	0.023	12380	0.0 %	12380				
5.0	5.000	4.920	-0.080	24078	0.0 %	24078				
7.5	7.500	7.515	0.015	36738	0.0 %	36738				
10.0	10.000	10.021	0.021	48968	0.0 %	48968				

LB96708 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jul 2018 18:31:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7471B	PPB	17 Jul 2018 18:34:19	S	Std
0.2 - 1	50.2	1197	-	0.2000	7471B	PPB	17 Jul 2018 18:36:36	S	Std
2.5 - 1	52.5	12380	-	2.5000	7471B	PPB	17 Jul 2018 18:38:54	S	Std
5.0 - 1	55	24078	-	5.0000	7471B	PPB	17 Jul 2018 18:41:11	S	Std
7.5 - 1	57.5	36738	-	7.5000	7471B	PPB	17 Jul 2018 18:46:01	S	Std
10.0 - 1	51.0	48968	-	10.0000	7471B	PPB	17 Jul 2018 18:50:52	S	Std
ICV56 - 1	ICV56	17997	3.6743	-	7471B	PPB	17 Jul 2018 18:53:33	U	SMPL
ICB56 - 1	ICB56	-167	-0.0481	-	7471B	PPB	17 Jul 2018 18:55:49	U	SMPL
CCV16 - 1	CCV16	24256	4.9570	-	7471B	PPB	17 Jul 2018 18:58:06	U	SMPL
CCB16 - 1	CCB16	-172	-0.0492	-	7471B	PPB	17 Jul 2018 19:00:21	U	SMPL
CRA - 1	CRA	1152	0.2222	-	7471B	PPB	17 Jul 2018 19:02:38	U	SMPL
HighStd - 1	HighStd	48757	9.9780	-	7471B	PPB	17 Jul 2018 19:04:54	U	SMPL
ChkStd - 1	ChkStd	31529	6.4474	-	7471B	PPB	17 Jul 2018 19:07:10	U	SMPL
PB111200BL - 1	PBS	-198	-0.0545	-	7471B	PPB	17 Jul 2018 19:09:28	U	SMPL
PB111200BS - 1	LCSS	17119	3.4943	-	7471B	PPB	17 Jul 2018 19:11:46	U	SMPL
J3826-03 - 1	OK-01-071618	3164	0.6345	-	7471B	PPB	17 Jul 2018 19:14:03	U	SMPL
J3832-03 - 1	TR-04-071618	4403	0.8884	-	7471B	PPB	17 Jul 2018 19:16:20	U	SMPL
J4011-03 - 1	20180413-CORE-3	716	0.1328	-	7471B	PPB	17 Jul 2018 19:18:38	U	SMPL
J4011-04 - 1	20180415-COMPOSITE-1	4031	0.8122	-	7471B	PPB	17 Jul 2018 19:20:55	U	SMPL
J4012-01 - 1	FS-RT-4917	5748	1.1640	-	7471B	PPB	17 Jul 2018 19:23:12	U	SMPL
CCV17 - 1	CCV17	23906	4.8852	-	7471B	PPB	17 Jul 2018 19:25:27	U	SMPL
CCB17 - 1	CCB17	-174	-0.0496	-	7471B	PPB	17 Jul 2018 19:27:43	U	SMPL
J4012-01DUP - 1	FS-RT-4917DUP	5522	1.1177	-	7471B	PPB	17 Jul 2018 19:30:00	U	SMPL
J4012-01MS - 1	FS-RT-4917MS	24293	4.9645	-	7471B	PPB	17 Jul 2018 19:32:16	U	SMPL
J4012-01MSD - 1	FS-RT-4917MSD	23451	4.7920	-	7471B	PPB	17 Jul 2018 19:34:31	U	SMPL
J4013-01 - 1	TP-I-S	803	0.1506	-	7471B	PPB	17 Jul 2018 19:36:48	U	SMPL
J4013-05 - 1	TP-I-D	545	0.0978	-	7471B	PPB	17 Jul 2018 19:39:06	U	SMPL
J4016-02 - 1	SB2-U-23-13	12168	2.4797	-	7471B	PPB	17 Jul 2018 19:41:23	U	SMPL
J4016-05 - 1	SB2-L-12-7	22585	4.6145	-	7471B	PPB	17 Jul 2018 19:43:40	U	SMPL
J4016-08 - 1	SB3-U-23-13	12913	2.6324	-	7471B	PPB	17 Jul 2018 19:45:58	U	SMPL
J4016-11 - 1	SB3-L-12-7	38047	7.7832	-	7471B	PPB	17 Jul 2018 19:48:17	U	SMPL
J4016-14 - 1	SB4-U-23-13	20221	4.1300	-	7471B	PPB	17 Jul 2018 19:50:33	U	SMPL
CCV18 - 1	CCV18	23866	4.8770	-	7471B	PPB	17 Jul 2018 19:52:51	U	SMPL
CCB18 - 1	CCB18	-175	-0.0498	-	7471B	PPB	17 Jul 2018 19:55:07	U	SMPL
J4016-17 - 1	SB4-L-12-7	20686	4.2253	-	7471B	PPB	17 Jul 2018 19:57:24	U	SMPL
J4016-20 - 1	SB5-U-23-13	13037	2.6578	-	7471B	PPB	17 Jul 2018 19:59:41	U	SMPL
J4016-23 - 1	SB5-L-12-7	22423	4.5813	-	7471B	PPB	17 Jul 2018 20:01:58	U	SMPL
J4012-01LX5 - 1		914	0.1734	-	7471B	PPB	17 Jul 2018 20:04:15	U	SMPL
J4012-01A - 1		23295	4.7600	-	7471B	PPB	17 Jul 2018 20:06:32	U	SMPL
CCV19 - 1	CCV19	23766	4.8565	-	7471B	PPB	17 Jul 2018 20:08:49	U	SMPL
CCB19 - 1	CCB19	-181	-0.0510	-	7471B	PPB	17 Jul 2018 20:11:07	U	SMPL