



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/18/2018

TEMP1 IN: 103 07/18/2018 07:10 TEMP OUT: 103 07/18/2018 08:15

TEMP2 IN: 104 07/18/2018 09:00 TEMP OUT: 104 07/18/2018 10:05

TEMP3 IN: 105 07/18/2018 11:30 TEMP OUT: 105 07/18/2018 13:05

TEMP4 IN: 105 07/18/2018 13:45 TEMP OUT: 105 07/18/2018 15:20

Run Number: LB96711

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96711BL	LB96711BL	1.4031	1.4031	100	1.4031	1.4031	1.4031	0.0000	0
2	LB96711BS	LB96711BS	1.3958	1.3958	100	1.4470	1.4470	1.4470	0.0512	512
3	J4018-02	DRUM-3-4	1.3990	1.3990	100	1.4021	1.4021	1.4021	0.0031	31
4	J4033-01	071618-TIPC-E-D-M	1.3988	1.3988	200	1.4107	1.4107	1.4107	0.0119	59.5
5	J4033-01DUP	071618-TIPC-E-D-MDUP	1.3999	1.3999	200	1.4120	1.4120	1.4120	0.0121	60.5
6	J4033-02	071618-TIPC-E-U-M	1.3963	1.3963	400	1.4110	1.4110	1.4110	0.0147	36.8
7	J4033-03	071618-DBRI-F-D-B	1.3953	1.3953	400	1.4157	1.4157	1.4157	0.0204	51
8	J4033-04	071618-DBRI-F-D-M	1.3981	1.3981	400	1.4129	1.4129	1.4129	0.0148	37
9	J4033-05	071618-DBRI-F-D-S	1.4104	1.4104	400	1.4221	1.4221	1.4221	0.0117	29.3
10	J4033-06	071618-DBRI-F-U-B	1.3988	1.3988	300	1.4210	1.4210	1.4210	0.0222	74
11	J4033-07	071618-DBRI-F-U-M	1.3954	1.3954	400	1.4152	1.4152	1.4152	0.0198	49.5
12	J4033-08	071618-DBRI-F-U-S	1.4008	1.4008	400	1.4129	1.4129	1.4129	0.0121	30.3

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS071818

WorkList ID : 114585

Date : 7/18/2018 7:42:41 AM

21396711

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2018	Water	J4018-02	TSS	Cool 4 deg C	PSEG01	K62	DRUM-3-4	07/17/2018	SM2540 D
07/23/2018	Water	J4033-01	TSS	Cool 4 deg C	HDR103	K63	071618-TIPC-E-D-M	07/16/2018	SM2540 D
07/23/2018	Water	J4033-02	TSS	Cool 4 deg C	HDR103	K63	071618-TIPC-E-U-M	07/16/2018	SM2540 D
07/23/2018	Water	J4033-03	TSS	Cool 4 deg C	HDR103	K63	071618-DBRI-F-D-B	07/16/2018	SM2540 D
07/23/2018	Water	J4033-04	TSS	Cool 4 deg C	HDR103	K63	071618-DBRI-F-D-M	07/16/2018	SM2540 D
07/23/2018	Water	J4033-05	TSS	Cool 4 deg C	HDR103	K63	071618-DBRI-F-D-S	07/16/2018	SM2540 D
07/23/2018	Water	J4033-06	TSS	Cool 4 deg C	HDR103	K63	071618-DBRI-F-U-B	07/16/2018	SM2540 D
07/23/2018	Water	J4033-07	TSS	Cool 4 deg C	HDR103	K63	071618-DBRI-F-U-M	07/16/2018	SM2540 D
07/23/2018	Water	J4033-08	TSS	Cool 4 deg C	HDR103	K63	071618-DBRI-F-U-S	07/16/2018	SM2540 D

FP-ID - 9823888

Date/Time 07/18/18, 09:00
 Received by: AS
 Relinquished by: CS

Date/Time 07/18/18, 12:00
 Received by: CS
 Relinquished by: AS