



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96717

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	07/18/2018	09:42
2	J4018-03		1	5	>100.0	>212.0	>212.0	07/18/2018	10:15
3	J4018-03DUP		1	5	>100.0	>212.0	>212.0	07/18/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 96717

WorkList Name : FP71818

WorkList ID : 114614

Date : 7/18/2018 9:04:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2018	Water	J4018-03 C	Flash Point	Cool 4 deg C	PSEG01	K62	DRUM-3-4	07/17/2018	1010A

Date/Time 7/18/18 9:15AM
 Received by: KS
 Relinquished by: CP

Date/Time 7/18/18 10:55AM
 Received by: CP
 Relinquished by: KS