

LB96718


7471B

INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

44497



A= 0.0000e+000

B= 2.2562e-004

C= -2.3683e-002

Rho= 0.9995899

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	21	0.000	21				
0.2	0.200	0.215	0.015	1056	0.0 %	1056				
2.5	2.500	2.427	-0.073	10863	0.0 %	10863				
5.0	5.000	5.202	0.202	23163	0.0 %	23163				
7.5	7.500	7.359	-0.141	32721	0.0 %	32721				
10.0	10.000	10.016	0.016	44497	0.0 %	44497				

LB96718 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Jul 2018 11:41:59

Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SO	21	-	0.00007471B	PPB	18 Jul 2018 11:44:36	S	Std
0.2 - 1	SO.2	1056	-	0.20007471B	PPB	18 Jul 2018 11:46:54	S	Std
2.5 - 1	SO.5	10863	-	2.50007471B	PPB	18 Jul 2018 11:49:11	S	Std
5.0 - 1	SO.5	23163	-	5.00007471B	PPB	18 Jul 2018 11:51:29	S	Std
7.5 - 1	SO.5	32721	-	7.50007471B	PPB	18 Jul 2018 11:53:46	S	Std
10.0 - 1	SO.5	44497	-	10.00007471B	PPB	18 Jul 2018 11:58:40	S	Std
ICV58 - 1	ICV58	17529	3.9312	-7471B	PPB	18 Jul 2018 12:01:28	U	SMPL
ICB58 - 1	ICB58	-91	-0.0442	-7471B	PPB	18 Jul 2018 12:03:43	U	SMPL
CCV24 - 1	CCV24	23459	5.2692	-7471B	PPB	18 Jul 2018 12:05:59	U	SMPL
CCB24 - 1	CCB24	-72	-0.0399	-7471B	PPB	18 Jul 2018 12:08:15	U	SMPL
CRA - 1	CRA	1055	0.2143	-7471B	PPB	18 Jul 2018 12:10:31	U	SMPL
HighStd - 1	HighStd	45155	10.1643	-7471B	PPB	18 Jul 2018 12:12:47	U	SMPL
ChkStd - 1	ChkStd	32518	7.3131	-7471B	PPB	18 Jul 2018 12:15:03	U	SMPL
PB111217BL - 1	PBS	12	-0.0210	-7471B	PPB	18 Jul 2018 12:19:55	U	SMPL
PB111217BS - 1	LCSS	15965	3.5784	-7471B	PPB	18 Jul 2018 12:22:11	U	SMPL
J4019-01 - 1	HS-SOIL	1504	0.3157	-7471B	PPB	18 Jul 2018 12:24:28	U	SMPL
J4026-01 - 1	KY030LC068-S-180716	1818	0.3865	-7471B	PPB	18 Jul 2018 12:26:45	U	SMPL
J4026-02 - 1	KY030LC051-S-180716	645	0.1218	-7471B	PPB	18 Jul 2018 12:29:03	U	SMPL
J4026-03 - 1	KY030LC001-S-180716	33937	7.6333	-7471B	PPB	18 Jul 2018 12:31:20	U	SMPL
J4026-04 - 1	KY030OF002-S-180716	44376	9.9885	-7471B	PPB	18 Jul 2018 12:33:36	U	SMPL
CCV25 - 1	CCV25	23411	5.2584	-7471B	PPB	18 Jul 2018 12:35:53	U	SMPL
CCB25 - 1	CCB25	-115	-0.0496	-7471B	PPB	18 Jul 2018 12:38:11	U	SMPL
J4026-05 - 1	KY030LC006-S-180716	1429	0.2987	-7471B	PPB	18 Jul 2018 12:40:27	U	SMPL
J4026-05DUP - 1	KY030LC006-S-180716DUP	1447	0.3028	-7471B	PPB	18 Jul 2018 12:47:58	U	SMPL
J4026-06 - 1	J4026-05MS	19214	4.3114	-7471B	PPB	18 Jul 2018 12:50:14	U	SMPL
J4026-07 - 1	J4026-05MSD	18241	4.0919	-7471B	PPB	18 Jul 2018 12:52:30	U	SMPL
J4026-08 - 1	DUP-0492-180716	1515	0.3181	-7471B	PPB	18 Jul 2018 12:54:46	U	SMPL
J4030-02 - 1	SB1-12-0	6935	1.5410	-7471B	PPB	18 Jul 2018 12:57:03	U	SMPL
J4030-05 - 1	SB6-12-7	38900	8.7530	-7471B	PPB	18 Jul 2018 12:59:20	U	SMPL
J4030-08 - 1	SB7-12-7	26635	5.9858	-7471B	PPB	18 Jul 2018 13:01:38	U	SMPL
J4030-11 - 1	SB8-12-0	18978	4.2582	-7471B	PPB	18 Jul 2018 13:03:57	U	SMPL
J4030-14 - 1	SB9-12-0	28434	6.3917	-7471B	PPB	18 Jul 2018 13:06:14	U	SMPL
CCV26 - 1	CCV26	23617	5.3048	-7471B	PPB	18 Jul 2018 13:08:31	U	SMPL
CCB26 - 1	CCB26	-89	-0.0438	-7471B	PPB	18 Jul 2018 13:10:48	U	SMPL
J4035-01 - 1	MH-2	1506	0.3161	-7471B	PPB	18 Jul 2018 13:13:05	U	SMPL
J4035-07 - 1	MH-3	3979	0.8741	-7471B	PPB	18 Jul 2018 13:15:22	U	SMPL
J4035-13 - 1	5-B	1493	0.3132	-7471B	PPB	18 Jul 2018 13:17:38	U	SMPL
PB111218BL - 1	PBS	106	0.0002	-7471B	PPB	18 Jul 2018 13:19:54	U	SMPL
PB111218BS - 1	LCSS	16415	3.6799	-7471B	PPB	18 Jul 2018 13:22:10	U	SMPL
J4027-01 - 1	WALL-1-161	3418	0.7475	-7471B	PPB	18 Jul 2018 13:24:27	U	SMPL
J4027-02 - 1	WALL-2-161	37374	8.4087	-7471B	PPB	18 Jul 2018 13:26:45	U	SMPL
J4027-03 - 1	WALL-3-161	666	0.1266	-7471B	PPB	18 Jul 2018 13:29:02	U	SMPL
J4027-04 - 1	WALL-4-161	594	0.1103	-7471B	PPB	18 Jul 2018 13:31:21	U	SMPL
J4027-05 - 1	FLOOR-1-161	130963	29.5245	-7471B	PPB	18 Jul 2018 13:33:38	U	SMPL
CCV27 - 1	CCV27	22789	5.1180	-7471B	PPB	18 Jul 2018 13:35:54	U	SMPL
CCB27 - 1	CCB27	-189	-0.0663	-7471B	PPB	18 Jul 2018 13:38:15	U	SMPL
J4027-06 - 1	WALL-4-159	343578	HIGH	-7471B	PPB	18 Jul 2018 13:40:33	X	
J4027-07 - 1	WALL-3-159	326	0.0499	-7471B	PPB	18 Jul 2018 13:42:23	U	SMPL
J4027-08 - 1	WALL-2-159	1225	0.2527	-7471B	PPB	18 Jul 2018 13:44:49	U	SMPL
J4027-09 - 1	WALL-1-159	1131	0.2315	-7471B	PPB	18 Jul 2018 13:47:05	U	SMPL
J4031-01 - 1	PLANTER-DOC-1	16606	3.7230	-7471B	PPB	18 Jul 2018 13:49:22	U	SMPL
J4031-02 - 1	PLANTER-DOC-2	79769	17.9740	-7471B	PPB	18 Jul 2018 13:51:39	U	SMPL
J4031-03 - 1	PLANTER-DOC-3	5014	1.1076	-7471B	PPB	18 Jul 2018 13:53:56	U	SMPL
J4031-04 - 1	PLANTER-DOC-4	3801	0.8339	-7471B	PPB	18 Jul 2018 13:56:16	U	SMPL
J4031-05 - 1	PLANTER-DOC-5	18901	4.2408	-7471B	PPB	18 Jul 2018 13:58:34	U	SMPL
J4031-06 - 1	PLANTER-DOC-6	17457	3.9150	-7471B	PPB	18 Jul 2018 14:00:50	U	SMPL
CCV28 - 1	CCV28	23239	5.2195	-7471B	PPB	18 Jul 2018 14:03:07	U	SMPL
CCB28 - 1	CCB28	-111	-0.0487	-7471B	PPB	18 Jul 2018 14:05:23	U	SMPL
J4031-07 - 1	PLANTER-DOC-7	836	0.1649	-7471B	PPB	18 Jul 2018 14:07:39	U	SMPL
J4031-08 - 1	PLANTER-DOC-8	18128	4.0664	-7471B	PPB	18 Jul 2018 14:09:56	U	SMPL
J4031-09 - 1	PLANTER-DOC-9	3605	0.7897	-7471B	PPB	18 Jul 2018 14:12:12	U	SMPL
J4031-09DUP - 1	PLANTER-DOC-9DUP	3629	0.7951	-7471B	PPB	18 Jul 2018 14:14:28	U	SMPL
J4031-09MS - 1	PLANTER-DOC-9MS	21573	4.8437	-7471B	PPB	18 Jul 2018 14:19:27	U	SMPL
J4031-09MSD - 1	PLANTER-DOC-9MSD	20733	4.6541	-7471B	PPB	18 Jul 2018 14:21:43	U	SMPL
J4031-10 - 1	PLANTER-DOC-10	12161	2.7201	-7471B	PPB	18 Jul 2018 14:24:01	U	SMPL
CCV29 - 1	CCV29	23471	5.2719	-7471B	PPB	18 Jul 2018 14:26:18	U	SMPL
CCB29 - 1	CCB29	-99	-0.0460	-7471B	PPB	18 Jul 2018 14:28:36	U	SMPL
J4026-05LX5 - 1		249	0.0325	-7471B	PPB	18 Jul 2018 14:30:52	U	SMPL
J4026-05A - 1		19164	4.3001	-7471B	PPB	18 Jul 2018 14:33:09	U	SMPL
J4031-09LX5 - 1		537	0.0975	-7471B	PPB	18 Jul 2018 14:35:25	U	SMPL
J4031-09A - 1		21081	4.7327	-7471B	PPB	18 Jul 2018 14:37:43	U	SMPL
J4027-05DLX5 - 1		23867	5.3612	-7471B	PPB	18 Jul 2018 14:39:59	U	SMPL
J4027-06DLX10 - 1		44440	10.0030	-7471B	PPB	18 Jul 2018 14:42:16	U	SMPL
J4031-02DLX5 - 1		14978	3.3557	-7471B	PPB	18 Jul 2018 14:44:34	U	SMPL
J4027-06DLX2X25 - 1		19157	4.2986	-7471B	PPB	18 Jul 2018 14:46:52	U	SMPL
CCV30 - 1	CCV30	23527	5.2845	-7471B	PPB	18 Jul 2018 14:49:09	U	SMPL
CCB30 - 1	CCB30	-98	-0.0458	-7471B	PPB	18 Jul 2018 14:51:26	U	SMPL