



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96734

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/18/2018	17:40
2	CAL2	1	Water			21.6	7.00	07/18/2018	17:42
3	CAL3	1	Water			21.7	10.03	07/18/2018	17:44
4	ICV	1	Water			21.8	7.00	07/18/2018	17:46
5	CCV1	1	Water			21.7	2.00	07/18/2018	18:16
6	J4043-04	1	Water			25.7	6.18	07/18/2018	18:18
7	J4043-04DUP	1	Water			25.8	6.19	07/18/2018	18:20
8	CCV2	1	Water			21.5	12.00	07/18/2018	18:22

WORKLIST(Hardcopy Internal Chain)

LB 96734

WorkList Name : pH071818

WorkList ID : 114630

Date : 7/18/2018 5:13:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/18/2018	Water	J4043-03	pH	Cool 4 deg C	PSEG01	K63	SD-BASELINE-WATER	07/18/2018	9040C

Date/Time 07/18/18 5:30
 Received by: SP
 Relinquished by: SP

Date/Time 07/18/18 6:30
 Received by: SP
 Relinquished by: SP