

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/19/2018

TEMP1 IN: 103 07/19/2018 07:15 TEMP OUT: 104 07/19/2018 08:15

TEMP2 IN: 104 07/19/2018 09:00 TEMP OUT: 104 07/19/2018 10:05

TEMP3 IN: 105 07/19/2018 11:30 TEMP OUT: 105 07/19/2018 13:10

TEMP4 IN: 105 07/19/2018 14:00 TEMP OUT: 105 07/19/2018 15:35

Run Number: LB96740

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96740BL	LB96740BL	1.4015	1.4015	100	1.4015	1.4015	1.4015	0.0000	0
2	LB96740BS	LB96740BS	1.3937	1.3937	100	1.4440	1.4440	1.4440	0.0503	503
3	J4043-03	SD-BASELINE-WATER	1.4087	1.4087	50	1.4133	1.4133	1.4133	0.0046	92
4	J4043-03DUP	SD-BASELINE-WATERDUP	1.4117	1.4117	50	1.4164	1.4164	1.4164	0.0047	94
5	J4053-01	001M-WILLETS-PT-BLVD (JULY)	1.4058	1.4058	200	1.4187	1.4187	1.4187	0.0129	64.5
6	J4053-02	002M-35TH-AVE (JULY)	1.4070	1.4070	100	1.4236	1.4236	1.4236	0.0166	166
7	J4058-01	COMP	1.3970	1.3970	50	1.4098	1.4098	1.4098	0.0128	256
8	J4059-01	DSN001	1.4021	1.4021	200	1.4104	1.4104	1.4104	0.0083	41.5
9	J4061-01	071718-DBRI-E-D-B	1.4044	1.4044	200	1.4192	1.4192	1.4192	0.0148	74
10	J4061-02	071718-DBRI-E-D-M	1.4038	1.4038	300	1.4198	1.4198	1.4198	0.0160	53.3
11	J4061-03	071718-DBRI-E-D-S	1.4086	1.4086	300	1.4241	1.4241	1.4241	0.0155	51.7
12	J4061-04	071718-DBRI-E-U-B	1.3902	1.3902	200	1.4093	1.4093	1.4093	0.0191	95.5
13	J4061-05	071718-DBRI-E-U-M	1.4008	1.4008	300	1.4175	1.4175	1.4175	0.0167	55.7
14	J4061-06	071718-DBRI-E-U-S	1.3915	1.3915	300	1.4001	1.4001	1.4001	0.0086	28.7
15	J4061-07	071718-TIPC-F-D-B	1.4090	1.4090	200	1.4241	1.4241	1.4241	0.0151	75.5

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/19/2018

TEMP1 IN: 103 07/19/2018 07:15 TEMP OUT: 104 07/19/2018 08:15

TEMP2 IN: 104 07/19/2018 09:00 TEMP OUT: 104 07/19/2018 10:05

TEMP3 IN: 105 07/19/2018 11:30 TEMP OUT: 105 07/19/2018 13:10

TEMP4 IN: 105 07/19/2018 14:00 TEMP OUT: 105 07/19/2018 15:35

Run Number: LB96740

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4061-08	071718-TIPC-F-D-S	1.4128	1.4128	300	1.4268	1.4268	1.4268	0.0140	46.7
17	J4061-09	071718-TIPC-F-U-B	1.4022	1.4022	200	1.4260	1.4260	1.4260	0.0238	119
18	J4061-10	071718-TIPC-F-U-S	1.4020	1.4020	200	1.4208	1.4208	1.4208	0.0188	94
19	J4061-10DUP	071718-TIPC-F-U-SDUP	1.3960	1.3960	200	1.4153	1.4153	1.4153	0.0193	96.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS071918

WorkList ID : 114643

Date : 7/19/2018 7:55:26 AM

2096740

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2018	Water	J4043-03 C	TSS	Cool 4 deg C	PSEG01	K63	SD-BASELINE-WATER	07/18/2018	SM2540 D
07/20/2018	Water	J4053-01 ND	TSS	Cool 4 deg C	TULL01	L61	001M-WILLETS-PT-BLVD(JUL	07/17/2018	SM2540 D
07/20/2018	Water	J4053-02 ND	TSS	Cool 4 deg C	TULL01	L61	002M-35TH-AVE(JULY)	07/17/2018	SM2540 D
07/25/2018	Water	J4058-01 ND	TSS	Cool 4 deg C	ARAM01	L61	COMP	07/18/2018	SM2540 D
07/25/2018	Water	J4059-01 H	TSS	Cool 4 deg C	NAT104	L63	DSN001	07/18/2018	SM2540 D
07/24/2018	Water	J4061-01 C	TSS	Cool 4 deg C	HDR103	M13	071718-DBRI-E-D-B	07/17/2018	SM2540 D
07/24/2018	Water	J4061-02 C	TSS	Cool 4 deg C	HDR103	M13	071718-DBRI-E-D-M	07/17/2018	SM2540 D
07/24/2018	Water	J4061-03 C	TSS	Cool 4 deg C	HDR103	M13	071718-DBRI-E-D-S	07/17/2018	SM2540 D
07/24/2018	Water	J4061-04 C	TSS	Cool 4 deg C	HDR103	M13	071718-DBRI-E-U-B	07/17/2018	SM2540 D
07/24/2018	Water	J4061-05 C	TSS	Cool 4 deg C	HDR103	M13	071718-DBRI-E-U-M	07/17/2018	SM2540 D
07/24/2018	Water	J4061-06 C	TSS	Cool 4 deg C	HDR103	M13	071718-DBRI-E-U-S	07/17/2018	SM2540 D
07/20/2018	Water	J4061-07 C	TSS	Cool 4 deg C	HDR103	M13	071718-TIPC-F-D-B	07/17/2018	SM2540 D
07/20/2018	Water	J4061-08 C	TSS	Cool 4 deg C	HDR103	M13	071718-TIPC-F-D-S	07/17/2018	SM2540 D
07/20/2018	Water	J4061-09 C	TSS	Cool 4 deg C	HDR103	M13	071718-TIPC-F-U-B	07/17/2018	SM2540 D
07/20/2018	Water	J4061-10 C	TSS	Cool 4 deg C	HDR103	M13	071718-TIPC-F-U-S	07/17/2018	SM2540 D

FP-ID-9823888

Date/Time 07/19/18, 08:00
 Received by: AS
 Relinquished by: Cg

Date/Time 07/19/18, 12:00
 Received by: Cg
 Relinquished by: AS