


ISM02.4

LB96741
INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 2.0570e-004

C= -5.8949e-002

Rho= 0.9999814

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.004	0.004	307	0.000	307					-1
0.2	0.200	0.199	-0.001	1254	0.0 %	1254					-1
2.5	2.500	2.524	0.024	12559	0.0 %	12559					-1
5.0	5.000	4.971	-0.029	24455	0.0 %	24455					-1
7.5	7.500	7.472	-0.028	36612	0.0 %	36612					0
10.0	10.000	10.029	0.029	49041	0.0 %	49041					0

MB
7119118

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	307	-		0 ISM02.4	PPB	7/19/2018 8:32	Std
	0.2 S0.2	1254	-		0.2 ISM02.4	PPB	7/19/2018 8:34	Std
	2.5 S2.5	12559	-		2.5 ISM02.4	PPB	7/19/2018 8:37	Std
	5 S5	24455	-		5 ISM02.4	PPB	7/19/2018 8:39	Std
	7.5 S7.5	36612	-		7.5 ISM02.4	PPB	7/19/2018 8:41	Std
	10 S10.0	49041	-		10 ISM02.4	PPB	7/19/2018 8:43	Std
ICV	ICV	18990	3.8473	-	ISM02.4	PPB	7/19/2018 8:46	SMPL
ICB	ICB	166	-0.0248	-	ISM02.4	PPB	7/19/2018 8:48	SMPL
CCV033	CCV033	24590	4.9992	-	ISM02.4	PPB	7/19/2018 8:51	SMPL
CCB033	CCB033	179	-0.0221	-	ISM02.4	PPB	7/19/2018 8:53	SMPL
PB111262BL	PBS004	243	-0.009	-	ISM02.4	PPB	7/19/2018 8:55	SMPL
J3963-19	MX2D63	10468	2.0943	-	ISM02.4	PPB	7/19/2018 8:58	SMPL
J3963-20	MX2D63D	11127	2.2299	-	ISM02.4	PPB	7/19/2018 9:00	SMPL
J3963-21	MX2D63S	22401	4.5489	-	ISM02.4	PPB	7/19/2018 9:07	SMPL
CCV034	CCV034	24092	4.8968	-	ISM02.4	PPB	7/19/2018 9:14	SMPL
CCB034	CCB034	148	-0.0285	-	ISM02.4	PPB	7/19/2018 9:17	SMPL