



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96742

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	07/19/2018	09:30
2	CAL2	1	Water			21.9	7.00	07/19/2018	09:32
3	CAL3	1	Water			22.2	10.02	07/19/2018	09:34
4	ICV	1	Water			22.0	7.00	07/19/2018	09:36
5	CCV1	1	Water			22.1	2.00	07/19/2018	09:38
6	J4062-06	1	Solid	20.02	20	24.1	8.45	07/19/2018	09:40
7	J4062-06DUP	1	Solid	20.03	20	24.2	8.46	07/19/2018	09:42
8	J4062-12	1	Solid	20.02	20	24.0	7.99	07/19/2018	09:44
9	J4062-18	1	Solid	20.02	20	23.6	8.60	07/19/2018	09:46
10	J4062-24	1	Solid	20.04	20	23.8	8.66	07/19/2018	09:48
11	J4062-30	1	Solid	20.02	20	23.8	8.86	07/19/2018	09:50
12	J4062-36	1	Solid	20.02	20	23.7	6.92	07/19/2018	09:52
13	J4062-42	1	Solid	20.03	20	23.5	7.78	07/19/2018	09:54
14	CCV2	1	Water			21.9	12.00	07/19/2018	09:56

WORKLIST(Hardcopy Internal Chain)

LB96742

WorkList Name : corrosivity071918

WorkList ID : 114661

Date : 7/19/2018 9:09:04 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/18/2018	Solid	J4062-06	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-3	07/18/2018	9045C
07/18/2018	Solid	J4062-12	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-4	07/18/2018	9045C
07/18/2018	Solid	J4062-18	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-5	07/18/2018	9045C
07/18/2018	Solid	J4062-24	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-8	07/18/2018	9045C
07/18/2018	Solid	J4062-30	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-9	07/18/2018	9045C
07/18/2018	Solid	J4062-36	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-11	07/18/2018	9045C
07/18/2018	Solid	J4062-42	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-10	07/18/2018	9045C

Date/Time 07/19/18 9:15

Received by: SP

Relinquished by: SP

Date/Time 07/19/18 10:10

Received by: SP

Relinquished by: SP