



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96746

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	07/19/2018	09:45
2	J4043-04		1	7	>100.0	>212.0	>212.0	07/19/2018	10:20
3	J4043-04DUP		1	7	>100.0	>212.0	>212.0	07/19/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96746

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71918

WorkList ID : 114662

Date : 7/19/2018 9:09:51 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2018	Water	J4043-04	B Flash Point	Cool 4 deg C	PSEG01	K63	SD-SD-BASELINE-WATER	07/18/2018	1010A

Date/Time 7/19/18 9:15AM
Received by: KS
Relinquished by: CP

Date/Time 7/19/18 10:55AM
Received by: CP
Relinquished by: KS