

2096756

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/19/2018 15:01

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.759	0.0	0.093	
ICB1	-0.223	0.0	0.004	
CCV1	243.515	0.0	0.226	
CCB1	-0.289	0.0	0.004	
PB111277BL	-0.504	0.0	0.004	
PB111277BS	98.919	0.0	0.095	
LOWPB111277	9.621	0.0	0.013	
HIGHPB111277	484.786	0.0	0.447	
J3828-07	2.328	0.0	0.006	
J3828-07DUP	1.853	0.0	0.006	
J3828-07MS	38.968	0.0	0.040	
J3828-07MSD	38.505	0.0	0.039	
J3828-09	0.112	0.0	0.004	
J3828-11	0.076	0.0	0.004	
CCV2	245.897	0.0	0.229	
CCB2	-0.742	0.0	0.004	
J4041-01	-0.673	0.0	0.004	
J4051-03	0.005	0.0	0.004	
J4054-06	-0.105	0.0	0.004	
J4057-01	1.769	0.0	0.006	
J4057-05	1.284	0.0	0.005	
J4057-09	1.612	0.0	0.006	
J4057-13	0.605	0.0	0.005	
J4057-17	1.055	0.0	0.005	
CCV3	253.162	0.0	0.235	
CCB3	-0.203	0.0	0.004	

N	26
Mean	58.427
SD	118.7556
CV%	203.26

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/19/2018 9:50


Test Total CN

Accepted 7/19/2018 9:50

Factor 1096
Bias 0.004

Coeff. of det. 0.999849

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9430	0.0000	
2	5.0PPBCN	0.008	3.9105	5.0000	
3	10PPBCN	0.012	8.8775	10.0000	
4	50PPBCN	0.049	49.4897	50.0000	
5	100PPBCN	0.097	101.8007	100.0000	
6	250PPBCN	0.236	254.2808	250.0000	
7	500PPBCN	0.458	497.5838	500.0000	

Aquakem v. 7.2AQ1

✓AS

Results from time period:

Thu Jul 19 14:47:58 2018

Thu Jul 19 15:00:50 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.943	µg/l	7/19/2018 9:46:06	
5.OPPBCN	A	Total CN	P	3.9105	µg/l	7/19/2018 9:46:07	
10PPBCN	A	Total CN	P	8.8775	µg/l	7/19/2018 9:46:08	
50PPBCN	A	Total CN	P	49.4897	µg/l	7/19/2018 9:46:09	
100PPBCN	A	Total CN	P	101.8007	µg/l	7/19/2018 9:46:10	
250PPBCN	A	Total CN	P	254.2808	µg/l	7/19/2018 9:46:11	
500PPBCN	A	Total CN	P	497.5838	µg/l	7/19/2018 9:46:12	
ICV1	S	Total CN	P	97.7587	µg/l	7/19/2018 14:47:58	
ICB1	S	Total CN	P	-0.223	µg/l	7/19/2018 14:47:59	
CCV1	S	Total CN	P	243.5148	µg/l	7/19/2018 14:48:00	
CCB1	S	Total CN	P	-0.2886	µg/l	7/19/2018 14:48:01	
PB111277BL	S	Total CN	P	-0.5039	µg/l	7/19/2018 14:48:02	
PB111277BS	S	Total CN	P	98.9192	µg/l	7/19/2018 14:48:03	
LOWPB111277	S	Total CN	P	9.6209	µg/l	7/19/2018 14:48:04	
HIGHPB111277	S	Total CN	P	484.7856	µg/l	7/19/2018 14:48:05	
J3828-07	S	Total CN	P	2.3285	µg/l	7/19/2018 14:48:06	
J3828-07DUP	S	Total CN	P	1.8527	µg/l	7/19/2018 14:48:07	
J3828-07MS	S	Total CN	P	38.968	µg/l	7/19/2018 14:48:08	
J3828-07MSD	S	Total CN	P	38.505	µg/l	7/19/2018 14:55:30	
J3828-09	S	Total CN	P	0.1122	µg/l	7/19/2018 14:55:31	
J3828-11	S	Total CN	P	0.0763	µg/l	7/19/2018 14:55:32	
CCV2	S	Total CN	P	245.8975	µg/l	7/19/2018 14:55:33	
CCB2	S	Total CN	P	-0.7419	µg/l	7/19/2018 14:55:34	
J4041-01	S	Total CN	P	-0.673	µg/l	7/19/2018 14:55:35	
J4051-03	S	Total CN	P	0.0047	µg/l	7/19/2018 14:55:36	
J4054-06	S	Total CN	P	-0.1051	µg/l	7/19/2018 14:55:37	
J4057-01	S	Total CN	P	1.7693	µg/l	7/19/2018 14:55:38	
J4057-05	S	Total CN	P	1.2845	µg/l	7/19/2018 14:55:39	
J4057-09	S	Total CN	P	1.6122	µg/l	7/19/2018 14:55:40	
J4057-13	S	Total CN	P	0.6047	µg/l	7/19/2018 15:00:45	
J4057-17	S	Total CN	P	1.0553	µg/l	7/19/2018 15:00:46	
CCV3	S	Total CN	P	253.1623	µg/l	7/19/2018 15:00:47	
CCB3	S	Total CN	P	-0.2026	µg/l	7/19/2018 15:00:48	