


245.1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	95	0.000	95				
0.05	0.050	0.045	-0.005	268	0.0 %	268				
0.2	0.200	0.207	0.007	868	0.0 %	868				
2.5	2.500	2.510	0.010	9393	0.0 %	9393				
5.0	5.000	5.051	0.051	18799	0.0 %	18799				
7.5	7.500	7.376	-0.124	27404	0.0 %	27404				
10.0	10.000	10.062	0.062	37347	0.0 %	37347				

LB96759 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Jul 2018 16:10:46

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	95	-	0.0000	245.1	PPB	19 Jul 2018 16:13:01	S	Std
0.05 - 1	50.25	268	-	0.0500	245.1	PPB	19 Jul 2018 16:15:09	S	Std
0.2 - 1	50.2	868	-	0.2000	245.1	PPB	19 Jul 2018 16:17:18	S	Std
2.5 - 1	52.5	9393	-	2.5000	245.1	PPB	19 Jul 2018 16:19:26	S	Std
5.0 - 1	55	18799	-	5.0000	245.1	PPB	19 Jul 2018 16:21:34	S	Std
7.5 - 1	57.5	27404	-	7.5000	245.1	PPB	19 Jul 2018 16:23:41	S	Std
10.0 - 1	510	37347	-	10.0000	245.1	PPB	19 Jul 2018 16:41:28	S	Std
ICV60 - 1	ICV60	14936	4.0077	-245.1	PPB	PPB	19 Jul 2018 16:48:46	U	SMPL
ICB60 - 1	ICB60	-42	-0.0389	-245.1	PPB	PPB	19 Jul 2018 16:50:52	U	SMPL
CCV41 - 1	CCV41	18547	4.9833	-245.1	PPB	PPB	19 Jul 2018 16:52:59	U	SMPL
CCB41 - 1	CCB41	-48	-0.0405	-245.1	PPB	PPB	19 Jul 2018 16:55:06	U	SMPL
CRA - 1	CRA	306	0.0551	-245.1	PPB	PPB	19 Jul 2018 16:59:34	U	SMPL
HighStd - 1	HighStd	37839	10.1953	-245.1	PPB	PPB	19 Jul 2018 17:01:40	U	SMPL
ChkStd - 1	ChkStd	25595	6.8874	-245.1	PPB	PPB	19 Jul 2018 17:03:47	U	SMPL
PB111247BL - 1	PBW	-76	-0.0481	-245.1	PPB	PPB	19 Jul 2018 17:05:57	U	SMPL
PB111247BS - 1	LCSW	13739	3.6843	-245.1	PPB	PPB	19 Jul 2018 17:08:06	U	SMPL
J3762-10 - 1	LOD-MDL-WATER-04-QT3-2018	306	0.0551	-245.1	PPB	PPB	19 Jul 2018 17:15:04	U	SMPL
J3762-11 - 1	LOQ-WATER-05-QT3-2018	1435	0.3602	-245.1	PPB	PPB	19 Jul 2018 17:17:11	U	SMPL
J3762-12 - 1	MDL-WATER-09-QT3-2018	287	0.0500	-245.1	PPB	PPB	19 Jul 2018 17:19:19	U	SMPL
J4033-01 - 1	071618-TIPC-E-D-M	49	-0.0143	-245.1	PPB	PPB	19 Jul 2018 17:21:28	U	SMPL
J4033-02 - 1	071618-TIPC-E-U-M	34	-0.0183	-245.1	PPB	PPB	19 Jul 2018 17:23:35	U	SMPL
CCV42 - 1	CCV42	18479	4.9649	-245.1	PPB	PPB	19 Jul 2018 17:25:41	U	SMPL
CCB42 - 1	CCB42	-38	-0.0378	-245.1	PPB	PPB	19 Jul 2018 17:27:47	U	SMPL
J4033-03 - 1	071618-DBRI-F-D-B	4	-0.0264	-245.1	PPB	PPB	19 Jul 2018 17:29:55	U	SMPL
J4033-04 - 1	071618-DBRI-F-D-M	10	-0.0248	-245.1	PPB	PPB	19 Jul 2018 17:32:02	U	SMPL
J4033-05 - 1	071618-DBRI-F-D-S	116	0.0038	-245.1	PPB	PPB	19 Jul 2018 17:34:08	U	SMPL
J4033-06 - 1	071618-DBRI-F-U-B	172	0.0189	-245.1	PPB	PPB	19 Jul 2018 17:36:16	U	SMPL
J4033-07 - 1	071618-DBRI-F-U-M	90	-0.0032	-245.1	PPB	PPB	19 Jul 2018 17:38:23	U	SMPL
J4033-08 - 1	071618-DBRI-F-U-S	-4	-0.0286	-245.1	PPB	PPB	19 Jul 2018 17:40:30	U	SMPL
J4033-08DUP - 1	071618-DBRI-F-U-SDUP	-78	-0.0486	-245.1	PPB	PPB	19 Jul 2018 17:42:37	U	SMPL
J4033-08MS - 1	071618-DBRI-F-U-SMS	16088	4.3189	-245.1	PPB	PPB	19 Jul 2018 17:44:46	U	SMPL
J4033-08MSD - 1	071618-DBRI-F-U-SMSD	16555	4.4451	-245.1	PPB	PPB	19 Jul 2018 17:46:54	U	SMPL
CCV43 - 1	CCV43	18440	4.9543	-245.1	PPB	PPB	19 Jul 2018 17:49:01	U	SMPL
CCB43 - 1	CCB43	-59	-0.0435	-245.1	PPB	PPB	19 Jul 2018 17:51:08	U	SMPL
J4033-08LX5 - 1		48	-0.0146	-245.1	PPB	PPB	19 Jul 2018 17:53:16	U	SMPL
J4033-08A - 1		16229	4.3570	-245.1	PPB	PPB	19 Jul 2018 17:55:23	U	SMPL
CCV44 - 1	CCV44	18355	4.9314	-245.1	PPB	PPB	19 Jul 2018 17:57:30	U	SMPL
CCB44 - 1	CCB44	-52	-0.0416	-245.1	PPB	PPB	19 Jul 2018 17:59:36	U	SMPL