



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96764

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	07/19/2018	17:10
2	CAL2	1	Water			21.9	7.00	07/19/2018	17:12
3	CAL3	1	Water			21.8	10.03	07/19/2018	17:14
4	ICV	1	Water			22.0	7.00	07/19/2018	17:16
5	CCV1	1	Water			21.9	2.00	07/19/2018	17:18
6	J4090-01	1	Solid	20.02	20	23.7	5.66	07/19/2018	17:20
7	J4090-01DUP	1	Solid	20.02	20	23.8	5.67	07/19/2018	17:22
8	CCV2	1	Water			21.8	12.00	07/19/2018	17:24

WORKLIST(Hardcopy Internal Chain)

LB 96764

WorkList Name : pH071918

WorkList ID : 114689

Date : 7/19/2018 4:23:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2018	Solid	J4090-01	pH	Cool 4 deg C	TULL02	M31	FK-PS-01-034	07/19/2018	9045C

Date/Time 07/19/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 07/19/18 5:30
 Received by: CP
 Relinquished by: SP