



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96765

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	07/19/2018	15:45
2	J4054-05		1	7	70.0	158.0	158.3	07/19/2018	16:10
3	J4054-05DUP		1	7	70.0	158.0	158.3	07/19/2018	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96765

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71918-1

WorkList ID : 114681

Date : 7/19/2018 3:22:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2018	Solid	J4054-05	Flash Point	Cool 4 deg C	PSEG01	K63	WASTE-OIL	07/18/2018	1010A

Date/Time 7/19/18 3:25PM
Received by: KS
Relinquished by: CP

Date/Time 7/19/18 16:40
Received by: CP
Relinquished by: KS