

7471B


LB 96766

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42500



A= 0.0000e+000

B= 2.3175e-004

C= -2.2424e-002

Rho= 0.9994157

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	69	0.000	69				
0.05	0.050	0.050	0.000	313	0.0 %	313				
0.2	0.200	0.208	0.008	996	0.0 %	996				
2.5	2.500	2.471	-0.029	10760	0.0 %	10760				
5.0	5.000	4.924	-0.076	21344	0.0 %	21344				
7.5	7.500	7.776	0.276	33649	0.0 %	33649				
10.0	10.000	9.827	-0.173	42500	0.0 %	42500				

LB96766 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jul 2018 18:27:08

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	69	-	0.00007471B	PPB	19 Jul 2018 18:59:57	S	Std
0.05 - 1	50.5	313	-	0.05007471B	PPB	19 Jul 2018 19:02:14	S	Std
0.2 - 1	52.2	996	-	0.20007471B	PPB	19 Jul 2018 19:04:32	S	Std
2.5 - 1	55	10760	-	2.50007471B	PPB	19 Jul 2018 19:06:49	S	Std
5.0 - 1	57.5	21344	-	5.00007471B	PPB	19 Jul 2018 19:09:06	S	Std
7.5 - 1	510	33649	-	7.50007471B	PPB	19 Jul 2018 19:11:22	S	Std
10.0 - 1		42500	-	10.00007471B	PPB	19 Jul 2018 19:13:39	S	Std
ICV61 - 1	ICV61	16756	3.8608	-7471B	PPB	19 Jul 2018 19:16:31	U	SMPL
ICB61 - 1	ICB61	-5	-0.0236	-7471B	PPB	19 Jul 2018 19:18:46	U	SMPL
CCV45 - 1	CCV45	21421	4.9419	-7471B	PPB	19 Jul 2018 19:23:33	U	SMPL
CCB45 - 1	CCB45	-6	-0.0238	-7471B	PPB	19 Jul 2018 19:25:48	U	SMPL
CRA - 1	CRA	963	0.2008	-7471B	PPB	19 Jul 2018 19:28:05	U	SMPL
HighStd - 1	HighStd	43109	9.9681	-7471B	PPB	19 Jul 2018 19:30:21	U	SMPL
ChkStd - 1	ChkStd	31394	7.2531	-7471B	PPB	19 Jul 2018 19:32:37	U	SMPL
PB111260BL - 1	PBS	-80	-0.0410	-7471B	PPB	19 Jul 2018 19:34:55	U	SMPL
PB111260BS - 1	LCSS	16957	3.9074	-7471B	PPB	19 Jul 2018 19:37:13	U	SMPL
J3828-07 - 1	EO-02-071818-A	5971	1.3614	-7471B	PPB	19 Jul 2018 19:39:30	U	SMPL
J3828-07DUP - 1	EO-02-071818-ADUP	5565	1.2673	-7471B	PPB	19 Jul 2018 19:41:47	U	SMPL
J3828-07MS - 1	EO-02-071818-AMS	21925	5.0587	-7471B	PPB	19 Jul 2018 19:44:05	U	SMPL
J3828-07MSD - 1	EO-02-071818-AMSD	23597	5.4462	-7471B	PPB	19 Jul 2018 19:46:23	U	SMPL
J3828-09 - 1	EO-02-071818-B	3119	0.7004	-7471B	PPB	19 Jul 2018 19:48:38	U	SMPL
CCV46 - 1	CCV46	21191	4.8886	-7471B	PPB	19 Jul 2018 19:50:55	U	SMPL
CCB46 - 1	CCB46	-5	-0.0236	-7471B	PPB	19 Jul 2018 19:53:11	U	SMPL
J3828-11 - 1	EO-02-071818-C	7423	1.6979	-7471B	PPB	19 Jul 2018 19:55:27	U	SMPL
J4041-01 - 1	OILY-STONES	2557	0.5702	-7471B	PPB	19 Jul 2018 19:57:43	U	SMPL
J4043-01 - 1	SD-DEBRI-WIPE	178	0.0188	-7471B	PPB	19 Jul 2018 19:59:59	U	SMPL
J4051-01 - 1	NS-DEBRIS	460	0.0842	-7471B	PPB	19 Jul 2018 20:02:15	U	SMPL
J4051-03 - 1	NS-SOIL	538	0.1023	-7471B	PPB	19 Jul 2018 20:04:31	U	SMPL
J4054-01 - 1	159-RR-FLOOR	62808	14.5333	-7471B	PPB	19 Jul 2018 20:06:49	U	SMPL
J4054-03 - 1	161-NEW-WALL	-139	-0.0546	-7471B	PPB	19 Jul 2018 20:09:06	U	SMPL
J4054-06 - 1	161-TP-1	3355	0.7551	-7471B	PPB	19 Jul 2018 20:11:26	U	SMPL
PB111261BL - 1	PBS	14	-0.0192	-7471B	PPB	19 Jul 2018 20:13:44	U	SMPL
PB111261BS - 1	LCSS	17198	3.9632	-7471B	PPB	19 Jul 2018 20:16:00	U	SMPL
CCV47 - 1	CCV47	21519	4.9646	-7471B	PPB	19 Jul 2018 20:18:16	U	SMPL
CCB47 - 1	CCB47	-20	-0.0271	-7471B	PPB	19 Jul 2018 20:20:32	U	SMPL
J3762-01 - 1	LOD-MDL-SOIL-01-QT3-2018	286	0.0439	-7471B	PPB	19 Jul 2018 20:22:49	U	SMPL
J3762-02 - 1	LOQ-SOIL-02-QT3-2018	1822	0.3998	-7471B	PPB	19 Jul 2018 20:25:05	U	SMPL
J3762-03 - 1	MDL-SOIL-03-QT3-2018	361	0.0612	-7471B	PPB	19 Jul 2018 20:27:22	U	SMPL
J4057-01 - 1	LS-SOIL-1	26807	6.1901	-7471B	PPB	19 Jul 2018 20:29:38	U	SMPL
J4057-01DUP - 1	LS-SOIL-1DUP	29437	6.7996	-7471B	PPB	19 Jul 2018 20:31:54	U	SMPL
J4057-01MS - 1	LS-SOIL-1MS	41207	9.5273	-7471B	PPB	19 Jul 2018 20:36:52	U	SMPL
J4057-01MSD - 1	LS-SOIL-1MSD	41669	9.6344	-7471B	PPB	19 Jul 2018 20:39:11	U	SMPL
J4057-05 - 1	LS-SOIL-2	7491	1.7136	-7471B	PPB	19 Jul 2018 20:41:30	U	SMPL
J4057-09 - 1	LS-SOIL-3	5986	1.3648	-7471B	PPB	19 Jul 2018 20:43:49	U	SMPL
J4057-13 - 1	LS-SOIL-4	7369	1.6853	-7471B	PPB	19 Jul 2018 20:46:05	U	SMPL
CCV48 - 1	CCV48	21429	4.9437	-7471B	PPB	19 Jul 2018 20:48:21	U	SMPL
CCB48 - 1	CCB48	196	0.0230	-7471B	PPB	19 Jul 2018 20:50:38	U	SMPL
J4057-17 - 1	LS-SOIL-5	8988	2.0605	-7471B	PPB	19 Jul 2018 20:52:54	U	SMPL
J4062-01 - 1	TP-3	1697	0.3709	-7471B	PPB	19 Jul 2018 20:55:10	U	SMPL
J4062-07 - 1	TP-4	1344	0.2890	-7471B	PPB	19 Jul 2018 20:57:27	U	SMPL
J4062-13 - 1	TP-5	2582	0.5760	-7471B	PPB	19 Jul 2018 20:59:43	U	SMPL
J4062-19 - 1	TP-8	1862	0.4091	-7471B	PPB	19 Jul 2018 21:02:00	U	SMPL
J4062-25 - 1	TP-9	2944	0.6598	-7471B	PPB	19 Jul 2018 21:04:17	U	SMPL
J4062-31 - 1	TP-11	2374	0.5278	-7471B	PPB	19 Jul 2018 21:06:33	U	SMPL
J4062-37 - 1	TP-10	5329	1.2126	-7471B	PPB	19 Jul 2018 21:08:51	U	SMPL
CCV49 - 1	CCV49	21247	4.9016	-7471B	PPB	19 Jul 2018 21:11:09	U	SMPL
CCB49 - 1	CCB49	-30	-0.0294	-7471B	PPB	19 Jul 2018 21:13:24	U	SMPL
J3828-07LX5 - 1		1107	0.2341	-7471B	PPB	19 Jul 2018 21:15:41	U	SMPL
J3828-07A - 1		22693	5.2367	-7471B	PPB	19 Jul 2018 21:17:58	U	SMPL
J4057-01LX5 - 1		5024	1.1419	-7471B	PPB	19 Jul 2018 21:20:14	U	SMPL
J4057-01A - 1		42137	9.7428	-7471B	PPB	19 Jul 2018 21:22:32	U	SMPL
J4054-01DLX2 - 1		30598	7.0687	-7471B	PPB	19 Jul 2018 21:24:48	U	SMPL
CCV50 - 1	CCV50	20802	4.7984	-7471B	PPB	19 Jul 2018 21:27:06	U	SMPL
CCB50 - 1	CCB50	-19	-0.0268	-7471B	PPB	19 Jul 2018 21:29:23	U	SMPL