



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96774

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	07/19/2018	17:10
2	CAL2	1	Water			21.9	7.00	07/19/2018	17:12
3	CAL3	1	Water			21.8	10.03	07/19/2018	17:14
4	ICV	1	Water			22.0	7.00	07/19/2018	17:16
5	CCV1	1	Water			21.8	2.00	07/19/2018	18:05
6	J4066-01	1	Water			24.6	7.24	07/19/2018	18:07
7	J4066-01DUP	1	Water			24.7	7.25	07/19/2018	18:09
8	CCV2	1	Water			21.7	12.00	07/19/2018	18:11

WORKLIST(Hardcopy Internal Chain)

LB 96 774

WorkList Name : pH071918-1

WorkList ID : 114690

Date : 7/19/2018 4:23:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2018	Water	J4066-01	pH	Cool 4 deg C	PSEG01	K63	FADW4704L	07/19/2018	9040C

Date/Time 07/19/18 5:55
 Received by: SP
 Relinquished by: CP

Date/Time 07/19/18 6:30
 Received by: CP
 Relinquished by: SP