



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96775

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	07/19/2018	17:10
2	CAL2	1	Water			21.9	7.00	07/19/2018	17:12
3	CAL3	1	Water			21.8	10.03	07/19/2018	17:14
4	ICV	1	Water			22.0	7.00	07/19/2018	17:16
5	CCV1	1	Water			21.9	2.00	07/19/2018	18:15
6	J3820-06	1	Solid	20.02	20	24.6	7.12	07/19/2018	18:17
7	J3820-06DUP	1	Solid	20.03	20	24.7	7.13	07/19/2018	18:19
8	J3827-04	1	Solid	20.02	20	24.9	7.95	07/19/2018	18:21
9	J4081-03	1	Solid	20.02	20	24.7	7.17	07/19/2018	18:23
10	J4087-53	1	Solid	20.03	20	22.9	10.63	07/19/2018	18:25
11	CCV2	1	Water			21.7	12.00	07/19/2018	18:27

WORKLIST(Hardcopy Internal Chain)

UB96775

WorkList Name : corrosivity071918-1

WorkList ID : 114687

Date : 7/19/2018 4:23:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2018	Solid	J3820-06	Corrosivity	Cool 4 deg C	PSEG05		BU-01-071918	07/19/2018	9045C
07/19/2018	Solid	J3827-04	Corrosivity	Cool 4 deg C	PSEG05		OR-01-071918	07/19/2018	9045C
07/19/2018	Solid	J4081-03	Corrosivity	Cool 4 deg C	PSEG01	K63	RBR200029	07/19/2018	9045C
07/19/2018	Solid	J4087-53	Corrosivity	Cool 4 deg C	PSEG01		ES-PILLAR-3-4	07/19/2018	9045C

Date/Time 07/19/18 5:55
 Received by: SP
 Relinquished by: CP

Date/Time 07/19/18 6:30
 Received by: CP
 Relinquished by: SP