

7470A


LB96782

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39870



A= 0.0000e+000

B= 2.5048e-004

C= -1.7124e-002

Rho= 0.9999563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	10	0.000	10				
0.2	0.200	0.207	0.007	895	0.0 %	895				
2.5	2.500	2.464	-0.036	9905	0.0 %	9905				
5.0	5.000	5.068	0.068	20301	0.0 %	20301				
7.5	7.500	7.506	0.006	30037	0.0 %	30037				
10.0	10.000	9.969	-0.031	39870	0.0 %	39870				

LB96782 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jul 2018 11:29:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.00007470A	PPB		20 Jul 2018 11:32:10	S	Std
0.2 - 1	50.2	895	-	0.20007470A	PPB		20 Jul 2018 11:34:18	S	Std
2.5 - 1	52.5	9905	-	2.50007470A	PPB		20 Jul 2018 11:36:27	S	Std
5.0 - 1	55	20301	-	5.00007470A	PPB		20 Jul 2018 11:38:35	S	Std
7.5 - 1	57.5	30037	-	7.50007470A	PPB		20 Jul 2018 11:43:06	S	Std
10.0 - 1	510	39870	-	10.00007470A	PPB		20 Jul 2018 11:47:40	S	Std
ICV62 - 1	ICV62	15428	3.8472	-7470A	PPB		20 Jul 2018 11:52:33	U	SMPL
ICB62 - 1	ICB62	-94	-0.0407	-7470A	PPB		20 Jul 2018 11:54:40	U	SMPL
CCV51 - 1	CCV51	19684	4.9133	-7470A	PPB		20 Jul 2018 11:56:46	U	SMPL
CCB51 - 1	CCB51	-119	-0.0469	-7470A	PPB		20 Jul 2018 11:58:53	U	SMPL
CRA - 1	CRA	845	0.1945	-7470A	PPB		20 Jul 2018 12:01:00	U	SMPL
HighStd - 1	HighStd	39000	9.7515	-7470A	PPB		20 Jul 2018 12:03:07	U	SMPL
ChkStd - 1	ChkStd	26931	6.7285	-7470A	PPB		20 Jul 2018 12:05:14	U	SMPL
PB111290BL - 1	PBW	-230	-0.0747	-7470A	PPB		20 Jul 2018 12:07:22	U	SMPL
PB111290BS - 1	LCSW	13996	3.4886	-7470A	PPB		20 Jul 2018 12:09:31	U	SMPL
PB111290TB - 1	PB111250TB	-139	-0.0519	-7470A	PPB		20 Jul 2018 12:11:39	U	SMPL
J3828-08 - 1	EO-02-071818-A	-71	-0.0349	-7470A	PPB		20 Jul 2018 12:13:47	U	SMPL
J3828-10 - 1	EO-02-071818-B	-26	-0.0236	-7470A	PPB		20 Jul 2018 12:15:55	U	SMPL
J3828-12 - 1	EO-02-071818-C	-65	-0.0334	-7470A	PPB		20 Jul 2018 12:18:04	U	SMPL
J4027-12 - 1	WALL-1-161	242	0.0435	-7470A	PPB		20 Jul 2018 12:20:10	U	SMPL
CCV52 - 1	CCV52	19904	4.9684	-7470A	PPB		20 Jul 2018 12:22:17	U	SMPL
CCB52 - 1	CCB52	-102	-0.0427	-7470A	PPB		20 Jul 2018 12:24:24	U	SMPL
J4027-13 - 1	WALL-2-161	271	0.0508	-7470A	PPB		20 Jul 2018 12:26:31	U	SMPL
J4027-14 - 1	WALL-3-161	-22	-0.0226	-7470A	PPB		20 Jul 2018 12:28:38	U	SMPL
J4027-15 - 1	WALL-4-161	-17	-0.0214	-7470A	PPB		20 Jul 2018 12:30:45	U	SMPL
J4027-16 - 1	FLOOR-1-161	23	-0.0114	-7470A	PPB		20 Jul 2018 12:32:52	U	SMPL
J4027-17 - 1	WALL-4-159	886	0.2048	-7470A	PPB		20 Jul 2018 12:35:00	U	SMPL
J4027-18 - 1	WALL-3-159	32	-0.0091	-7470A	PPB		20 Jul 2018 12:37:07	U	SMPL
J4027-19 - 1	WALL-2-159	39	-0.0074	-7470A	PPB		20 Jul 2018 12:39:15	U	SMPL
J4027-20 - 1	WALL-3-159	102	0.0084	-7470A	PPB		20 Jul 2018 12:41:24	U	SMPL
J4035-06 - 1	MH-2	-27	-0.0239	-7470A	PPB		20 Jul 2018 12:43:32	U	SMPL
J4035-06DUP - 1	MH-2DUP	-12	-0.0201	-7470A	PPB		20 Jul 2018 12:45:39	U	SMPL
CCV53 - 1	CCV53	19795	4.9411	-7470A	PPB		20 Jul 2018 12:47:46	U	SMPL
CCB53 - 1	CCB53	-74	-0.0357	-7470A	PPB		20 Jul 2018 12:49:54	U	SMPL
J4035-06MS - 1	MH-2MS	13440	3.3493	-7470A	PPB		20 Jul 2018 12:52:01	U	SMPL
J4035-06MSD - 1	MH-2MSD	14255	3.5534	-7470A	PPB		20 Jul 2018 12:54:08	U	SMPL
J4035-12 - 1	MH-3	-123	-0.0479	-7470A	PPB		20 Jul 2018 12:56:15	U	SMPL
J4035-18 - 1	5-B	9	-0.0149	-7470A	PPB		20 Jul 2018 12:58:22	U	SMPL
J4041-02 - 1	OILY-STONES	0	-0.0171	-7470A	PPB		20 Jul 2018 13:00:30	U	SMPL
J4043-02 - 1	SD-DEBRI-WIPE	1	-0.0169	-7470A	PPB		20 Jul 2018 13:02:38	U	SMPL
J4051-02 - 1	NS-DEBRIS	-16	-0.0211	-7470A	PPB		20 Jul 2018 13:04:46	U	SMPL
J4051-06 - 1	NS-SOIL	-38	-0.0266	-7470A	PPB		20 Jul 2018 13:06:54	U	SMPL
PB111291BL - 1	PBW	-17	-0.0214	-7470A	PPB		20 Jul 2018 13:09:02	U	SMPL
PB111291BS - 1	LCSW	14794	3.6884	-7470A	PPB		20 Jul 2018 13:17:18	U	SMPL
CCV54 - 1	CCV54	19898	4.9669	-7470A	PPB		20 Jul 2018 13:19:24	U	SMPL
CCB54 - 1	CCB54	-81	-0.0374	-7470A	PPB		20 Jul 2018 13:21:32	U	SMPL
PB111291TB - 1	PB111259TB	-22	-0.0226	-7470A	PPB		20 Jul 2018 13:23:39	U	SMPL
J4054-02 - 1	159-RR-FLOOR	402	0.0836	-7470A	PPB		20 Jul 2018 13:25:47	U	SMPL
J4054-04 - 1	161-NEW-WALL	-33	-0.0254	-7470A	PPB		20 Jul 2018 13:27:54	U	SMPL
J4054-05 - 1	WASTE-OIL	12	-0.0141	-7470A	PPB		20 Jul 2018 13:30:01	U	SMPL
J4054-09 - 1	161-TP-1	-33	-0.0254	-7470A	PPB		20 Jul 2018 13:32:08	U	SMPL
J4057-04 - 1	LS-SOIL-1	-34	-0.0256	-7470A	PPB		20 Jul 2018 13:34:15	U	SMPL
J4057-08 - 1	LS-SOIL-2	13	-0.0139	-7470A	PPB		20 Jul 2018 13:36:23	U	SMPL
J4057-12 - 1	LS-SOIL-3	-32	-0.0251	-7470A	PPB		20 Jul 2018 13:38:31	U	SMPL
J4057-16 - 1	LS-SOIL-4	-11	-0.0199	-7470A	PPB		20 Jul 2018 13:40:40	U	SMPL
J4057-20 - 1	LS-SOIL-5	23	-0.0114	-7470A	PPB		20 Jul 2018 13:42:47	U	SMPL
CCV55 - 1	CCV55	19545	4.8785	-7470A	PPB		20 Jul 2018 13:44:54	U	SMPL
CCB55 - 1	CCB55	-78	-0.0367	-7470A	PPB		20 Jul 2018 13:47:01	U	SMPL
J4057-20DUP - 1	LS-SOIL-5DUP	-8	-0.0191	-7470A	PPB		20 Jul 2018 13:49:09	U	SMPL
J4057-20MS - 1	LS-SOIL-5MS	13561	3.3796	-7470A	PPB		20 Jul 2018 13:51:17	U	SMPL
J4057-20MSD - 1	LS-SOIL-5MSD	13364	3.3303	-7470A	PPB		20 Jul 2018 13:53:23	U	SMPL
PB111292BL - 1	PBW	-100	-0.0422	-7470A	PPB		20 Jul 2018 13:55:31	U	SMPL
PB111292BS - 1	LCSW	13403	3.3400	-7470A	PPB		20 Jul 2018 13:57:38	U	SMPL
PB111292TB - 1	PB111252TB	-77	-0.0364	-7470A	PPB		20 Jul 2018 13:59:46	U	SMPL
J4018-03 - 1	DRUM-3-4	69	0.0002	-7470A	PPB		20 Jul 2018 14:01:54	U	SMPL
J4018-03DUP - 1	DRUM-3-4DUP	82	0.0034	-7470A	PPB		20 Jul 2018 14:04:02	U	SMPL
J4018-03MS - 1	DRUM-3-4MS	13707	3.4162	-7470A	PPB		20 Jul 2018 14:06:11	U	SMPL
J4018-03MSD - 1	DRUM-3-4MSD	14396	3.5888	-7470A	PPB		20 Jul 2018 14:08:18	U	SMPL
CCV56 - 1	CCV56	20105	5.0187	-7470A	PPB		20 Jul 2018 14:10:26	U	SMPL
CCB56 - 1	CCB56	-99	-0.0419	-7470A	PPB		20 Jul 2018 14:12:33	U	SMPL
J4043-04 - 1	SD-BASELINE-WATER	-2	-0.0176	-7470A	PPB		20 Jul 2018 14:14:42	U	SMPL
PB111293BL - 1	PBW	-6	-0.0186	-7470A	PPB		20 Jul 2018 14:16:49	U	SMPL
PB111293BS - 1	LCSW	13885	3.4608	-7470A	PPB		20 Jul 2018 14:18:57	U	SMPL
J4043-03 - 1	SD-BASELINE-WATER	-141	-0.0524	-7470A	PPB		20 Jul 2018 14:21:04	U	SMPL
J4043-03DUP - 1	SD-BASELINE-WATERDUP	-37	-0.0264	-7470A	PPB		20 Jul 2018 14:23:11	U	SMPL
J4043-03MS - 1	SD-BASELINE-WATERMS	14477	3.6090	-7470A	PPB		20 Jul 2018 14:27:45	U	SMPL
J4043-03MSD - 1	SD-BASELINE-WATERMSD	13928	3.4715	-7470A	PPB		20 Jul 2018 14:29:53	U	SMPL
CCV57 - 1	CCV57	19346	4.8286	-7470A	PPB		20 Jul 2018 14:32:02	U	SMPL
CCB57 - 1	CCB57	-98	-0.0417	-7470A	PPB		20 Jul 2018 14:34:10	U	SMPL
J4035-06LX5 - 1		25	-0.0109	-7470A	PPB		20 Jul 2018 14:36:18	U	SMPL
J4035-06A - 1		14513	3.6181	-7470A	PPB		20 Jul 2018 14:38:26	U	SMPL

LB96782 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jul 2018 11:29:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J4057-20LX5 - 1		-71	-0.0349		-7470A	PPB	20 Jul 2018 14:40:34	U	SMPL
J4057-20A - 1		14813	3.6932		-7470A	PPB	20 Jul 2018 14:42:41	U	SMPL
J4018-03LX5 - 1		-66	-0.0337		-7470A	PPB	20 Jul 2018 14:44:49	U	SMPL
J4018-03A - 1		14692	3.6629		-7470A	PPB	20 Jul 2018 14:46:57	U	SMPL
J4043-03LX5 - 1		-103	-0.0429		-7470A	PPB	20 Jul 2018 14:49:05	U	SMPL
J4043-03A - 1		15105	3.7663		-7470A	PPB	20 Jul 2018 15:28:10	U	SMPL
CCV58 - 1	CCV58	19996	4.9914		-7470A	PPB	20 Jul 2018 15:30:17	U	SMPL
CCB58 - 1	CCB58	-102	-0.0427		-7470A	PPB	20 Jul 2018 15:32:26	U	SMPL