



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96783

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	07/20/2018	09:45
2	J4066-02		1	6	>100.0	>212.0	>212.0	07/20/2018	10:20
3	J4066-02DUP		1	6	>100.0	>212.0	>212.0	07/20/2018	10:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96783

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP72018

WorkList ID : 114720

Date : 7/20/2018 9:05:42 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Water	J4066-02 C	Flash Point	Cool 4 deg C	PSEG01	K63	FADW4704L	07/19/2018	1010A

Date/Time 7/20/18 9:15 AM
Received by: KS
Relinquished by: CP

Date/Time 7/20/18 11:05
Received by: CP
Relinquished by: KS