


245.1

LB96790
INSTRUMENT 10: CV2

Linear

 μ Abs.:

36997



A= 0.0000e+000

B= 2.6869e-004

C= -1.2441e-002

Rho= 0.9998191

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	145	0.000	145				
0.05	0.050	0.058	0.008	263	0.0 %	263				
0.2	0.200	0.220	0.020	864	0.0 %	864				
2.5	2.500	2.397	-0.103	8969	0.0 %	8969				
5.0	5.000	4.986	-0.014	18602	0.0 %	18602				
7.5	7.500	7.634	0.134	28457	0.0 %	28457				
10.0	10.000	9.928	-0.072	36997	0.0 %	36997				

LB96790 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Jul 2018 16:03:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	145	-	0.0000	245.1	PPB	20 Jul 2018 16:08:57	S	Std
0.05 - 1	50.05	263	-	0.0500	245.1	PPB	20 Jul 2018 16:14:45	S	Std
0.2 - 1	50.2	864	-	0.2000	245.1	PPB	20 Jul 2018 16:16:53	S	Std
2.5 - 1	52.5	8969	-	2.5000	245.1	PPB	20 Jul 2018 16:19:01	S	Std
5.0 - 1	55	18602	-	5.0000	245.1	PPB	20 Jul 2018 16:21:09	S	Std
7.5 - 1	57.5	28457	-	7.5000	245.1	PPB	20 Jul 2018 16:23:16	S	Std
10.0 - 1	510	36997	-	10.0000	245.1	PPB	20 Jul 2018 16:25:24	S	Std
ICV63 - 1	ICV63	15259	4.0876	-245.1		PPB	20 Jul 2018 16:33:30	U	SMPL
ICB63 - 1	ICB63	1	-0.0122	-245.1		PPB	20 Jul 2018 16:35:36	U	SMPL
CCV59 - 1	CCV59	18666	5.0030	-245.1		PPB	20 Jul 2018 16:37:43	U	SMPL
CCB59 - 1	CCB59	-21	-0.0181	-245.1		PPB	20 Jul 2018 16:39:50	U	SMPL
CRA - 1	CRA	219	0.0464	-245.1		PPB	20 Jul 2018 16:41:57	U	SMPL
HighStd - 1	HighStd	37213	9.9865	-245.1		PPB	20 Jul 2018 16:44:03	U	SMPL
ChkStd - 1	ChkStd	26053	6.9879	-245.1		PPB	20 Jul 2018 16:46:10	U	SMPL
PB111300BL - 1	PBW	-154	-0.0538	-245.1		PPB	20 Jul 2018 16:48:19	U	SMPL
PB111300BS - 1	LCSW	15085	4.0408	-245.1		PPB	20 Jul 2018 16:50:28	U	SMPL
J4061-01 - 1	071718-DBRI-E-D-B	-146	-0.0517	-245.1		PPB	20 Jul 2018 16:52:36	U	SMPL
J4061-02 - 1	071718-DBRI-E-D-M	-53	-0.0267	-245.1		PPB	20 Jul 2018 16:54:44	U	SMPL
J4061-03 - 1	071718-DBRI-E-D-S	-48	-0.0253	-245.1		PPB	20 Jul 2018 16:56:52	U	SMPL
J4061-04 - 1	071718-DBRI-E-U-B	-48	-0.0253	-245.1		PPB	20 Jul 2018 16:59:00	U	SMPL
J4061-05 - 1	071718-DBRI-E-U-M	-93	-0.0374	-245.1		PPB	20 Jul 2018 17:01:07	U	SMPL
CCV60 - 1	CCV60	18533	4.9673	-245.1		PPB	20 Jul 2018 17:03:14	U	SMPL
CCB60 - 1	CCB60	-12	-0.0157	-245.1		PPB	20 Jul 2018 17:05:20	U	SMPL
J4061-06 - 1	071718-DBRI-E-U-S	-98	-0.0388	-245.1		PPB	20 Jul 2018 17:07:28	U	SMPL
J4061-07 - 1	071718-TIPC-F-D-B	-66	-0.0302	-245.1		PPB	20 Jul 2018 17:09:34	U	SMPL
J4061-08 - 1	071718-TIPC-F-D-S	-52	-0.0264	-245.1		PPB	20 Jul 2018 17:11:41	U	SMPL
J4061-09 - 1	071718-TIPC-F-U-B	-22	-0.0184	-245.1		PPB	20 Jul 2018 17:13:48	U	SMPL
J4061-10 - 1	071718-TIPC-F-U-S	-26	-0.0194	-245.1		PPB	20 Jul 2018 17:15:55	U	SMPL
J4070-04 - 1	LC-SD-EB02-180718	-113	-0.0428	-245.1		PPB	20 Jul 2018 17:18:03	U	SMPL
J4094-01 - 1	071818-DBRI-E-D-B	-67	-0.0304	-245.1		PPB	20 Jul 2018 17:20:11	U	SMPL
J4094-02 - 1	071818-DBRI-E-D-M	-79	-0.0337	-245.1		PPB	20 Jul 2018 17:22:20	U	SMPL
J4094-03 - 1	071818-DBRI-E-D-S	-64	-0.0296	-245.1		PPB	20 Jul 2018 17:24:28	U	SMPL
J4094-04 - 1	071818-DBRI-E-U-B	-80	-0.0339	-245.1		PPB	20 Jul 2018 17:26:35	U	SMPL
CCV61 - 1	CCV61	18677	5.0060	-245.1		PPB	20 Jul 2018 17:28:42	U	SMPL
CCB61 - 1	CCB61	-17	-0.0170	-245.1		PPB	20 Jul 2018 17:30:49	U	SMPL
J4094-05 - 1	071818-DBRI-E-U-M	-72	-0.0318	-245.1		PPB	20 Jul 2018 17:32:57	U	SMPL
J4094-06 - 1	071818-DBRI-E-U-S	-98	-0.0388	-245.1		PPB	20 Jul 2018 17:35:04	U	SMPL
J4094-07 - 1	071818-TIPC-E-D-M	-62	-0.0291	-245.1		PPB	20 Jul 2018 17:37:11	U	SMPL
J4094-08 - 1	071818-TIPC-E-U-M	-28	-0.0200	-245.1		PPB	20 Jul 2018 17:39:19	U	SMPL
J4094-08DUP - 1	071818-TIPC-E-U-MDUP	-11	-0.0154	-245.1		PPB	20 Jul 2018 17:41:26	U	SMPL
J4094-08MS - 1	071818-TIPC-E-U-MMS	15997	4.2859	-245.1		PPB	20 Jul 2018 17:43:34	U	SMPL
J4094-08MSD - 1	071818-TIPC-E-U-MMSD	15058	4.0336	-245.1		PPB	20 Jul 2018 17:45:41	U	SMPL
J4094-08LX5 - 1		-59	-0.0283	-245.1		PPB	20 Jul 2018 17:47:49	U	SMPL
J4094-08A - 1		16296	4.3662	-245.1		PPB	20 Jul 2018 17:49:58	U	SMPL
CCV62 - 1	CCV62	18728	5.0197	-245.1		PPB	20 Jul 2018 17:52:05	U	SMPL
CCB62 - 1	CCB62	-32	-0.0210	-245.1		PPB	20 Jul 2018 17:54:12	U	SMPL