



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96794

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/20/2018	16:20
2	CAL2	1	Water			21.6	7.00	07/20/2018	16:22
3	CAL3	1	Water			21.6	10.03	07/20/2018	16:24
4	ICV	1	Water			21.5	7.00	07/20/2018	16:26
5	CCV1	1	Water			21.7	2.00	07/20/2018	16:28
6	J4101-02	1	Solid	20.02	20	22.9	7.16	07/20/2018	16:30
7	J4101-02DUP	1	Solid	20.02	20	23.0	7.17	07/20/2018	16:32
8	J4101-04	1	Solid	20.04	20	23.7	6.73	07/20/2018	16:34
9	J4102-01	1	Solid	20.02	20	22.6	8.38	07/20/2018	16:36
10	J4102-02	1	Solid	20.03	20	22.0	7.04	07/20/2018	16:38
11	CCV2	1	Water			21.6	12.00	07/20/2018	16:40

WORKLIST(Hardcopy Internal Chain)

LB 96794

WorkList Name : corrosivity072018

WorkList ID : 114733

Date : 7/20/2018 4:02:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2018	Solid	J4101-02	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-1	07/20/2018	9045C
07/20/2018	Solid	J4101-04	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-2	07/20/2018	9045C
07/20/2018	Solid	J4102-01	Corrosivity	Cool 4 deg C	PSEG01	K61	TA-DRUM-1-OILY-STONE	07/20/2018	9045C
07/20/2018	Solid	J4102-02	Corrosivity	Cool 4 deg C	PSEG01	K61	TA-DRUM-5-OILY-WIPE	07/20/2018	9045C

Date/Time 07/20/18 4:10
 Received by: SP
 Relinquished by: CP

Date/Time 07/20/18 5:10
 Received by: CP
 Relinquished by: SP