



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96795

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/20/2018	16:20
2	CAL2	1	Water			21.6	7.00	07/20/2018	16:22
3	CAL3	1	Water			21.6	10.03	07/20/2018	16:24
4	ICV	1	Water			21.5	7.00	07/20/2018	16:26
5	CCV1	1	Water			21.7	2.00	07/20/2018	16:45
6	J4102-03	1	Water			24.1	7.44	07/20/2018	16:47
7	J4102-03DUP	1	Water			24.2	7.45	07/20/2018	16:49
8	J4102-04	1	Water			23.3	7.37	07/20/2018	16:51
9	J4102-05	1	Water			24.5	6.09	07/20/2018	16:53
10	CCV2	1	Water			21.7	12.00	07/20/2018	16:55

WORKLIST(Hardcopy Internal Chain)

1396795

WorkList Name : corrosivity072018-1

WorkList ID : 114734

Date : 7/20/2018 4:02:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/20/2018	Water	J4102-03	Corrosivity	Cool 4 deg C	PSEG01	K61	TA-DRUM-2-OILY-WATER	07/20/2018	9040C
07/20/2018	Water	J4102-04	Corrosivity	Cool 4 deg C	PSEG01	K61	TA-DRUM-3-OILY-WATER	07/20/2018	9040C
07/20/2018	Water	J4102-05	Corrosivity	Cool 4 deg C	PSEG01	K61	TA-DRUM-4-OILY-WATER	07/20/2018	9040C

Date/Time 07/20/18 4:10
 Received by: SP
 Relinquished by: CP

Date/Time 07/20/18 5:10
 Received by: CP
 Relinquished by: SP