



PERCENT SOLID

Analyst Name: jignesh
Date: 7/23/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 07/20/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 07/21/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
BalanceID: M Sc-1

QC: LB96799

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J4054-09	161-TP-1	1	1.13	9.98	9.10	90.1
J4062-06	TP-3	2	1.15	9.90	8.79	87.3
J4062-12	TP-4	3	1.13	9.62	8.63	88.3
J4062-18	TP-5	4	1.14	9.70	8.72	88.6
J4062-24	TP-8	5	1.15	9.87	8.84	88.2
J4062-30	TP-9	6	1.14	9.86	8.71	86.8
J4062-36	TP-11	7	1.11	9.96	9.00	89.2
J4062-42	TP-10	8	1.14	9.90	8.60	85.2
J4081-05	RBR200029-GRO-B	9	1.13	9.86	9.00	90.1
J4087-01	ES-SIDE-1-PILLAR-6-TOP	10	1.00	2.00	2.00	100
J4087-03	ES-SIDE-1-PILLAR-6-BOTTO	11	1.00	2.00	2.00	100
J4087-05	ES-SIDE-1-PILLAR-7-TOP	12	1.00	2.00	2.00	100
J4087-07	ES-SIDE-1-PILLAR-7-BOTTO	13	1.00	2.00	2.00	100
J4087-09	ES-SIDE-1-PILLAR-8-TOP	14	1.00	2.00	2.00	100
J4087-11	ES-SIDE-1-PILLAR-8-BOTTO	15	1.00	2.00	2.00	100
J4087-13	ES-SIDE-1-PILLAR-9-TOP	16	1.00	2.00	2.00	100
J4087-15	ES-SIDE-1-PILLAR-9-BOTTO	17	1.00	2.00	2.00	100
J4087-17	ES-SIDE-2-PILLAR-1-TOP	18	1.00	2.00	2.00	100
J4087-19	ES-SIDE-2-PILLAR-1-BOTTO	19	1.00	2.00	2.00	100
J4087-21	ES-SIDE-2-PILLAR-2-TOP	20	1.00	2.00	2.00	100
J4087-23	ES-SIDE-2-PILLAR-2-BOTTO	21	1.00	2.00	2.00	100
J4087-25	ES-SIDE-2-PILLAR-3-TOP	22	1.00	2.00	2.00	100
J4087-27	ES-SIDE-2-PILLAR-3-BOTTO	23	1.00	2.00	2.00	100
J4087-29	ES-SIDE-2-PILLAR-4-TOP	24	1.00	2.00	2.00	100
J4087-31	ES-SIDE-2-PILLAR-4-BOTTO	25	1.00	2.00	2.00	100
J4087-33	ES-SIDE-2-PILLAR-5-TOP	26	1.00	2.00	2.00	100
J4087-35	ES-SIDE-2-PILLAR-5-BOTTO	27	1.00	2.00	2.00	100
J4087-37	ES-SIDE-2-PILLAR-6-TOP	28	1.00	2.00	2.00	100
J4087-39	ES-SIDE-2-PILLAR-6-BOTTO	29	1.00	2.00	2.00	100
J4087-41	ES-SIDE-2-PILLAR-7-TOP	30	1.00	2.00	2.00	100
J4087-43	ES-SIDE-2-PILLAR-7-BOTTO	31	1.00	2.00	2.00	100
J4087-45	ES-SIDE-2-PILLAR-8-TOP	32	1.00	2.00	2.00	100
J4087-47	ES-SIDE-2-PILLAR-8-BOTTO	33	1.00	2.00	2.00	100
J4087-49	ES-SIDE-2-PILLAR-9-TOP	34	1.00	2.00	2.00	100
J4087-51	ES-SIDE-2-PILLAR-9-BOTTO	35	1.00	2.00	2.00	100
J4087-54	NOTHING	36	1.00	2.00	2.00	100
J4087-56	ES-SIDE-1-PILLAR-1-TOP	37	1.00	2.00	2.00	100
J4087-58	ES-SIDE-1-PILLAR-1-BOTTO	38	1.00	2.00	2.00	100
J4087-60	ES-SIDE-1-PILLAR-2-TOP	39	1.00	2.00	2.00	100



PERCENT SOLID

Analyst Name: jignesh
Date: 7/23/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 07/20/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 07/21/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
BalanceID: M Sc-1

QC: LB96799

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J4087-62	ES-SIDE-1-PILLAR-2-BOTT	40	1.00	2.00	2.00	100
J4087-64	ES-SIDE-1-PILLAR-3-TOP	41	1.00	2.00	2.00	100
J4087-66	ES-SIDE-1-PILLAR-3-BOTT	42	1.00	2.00	2.00	100
J4087-68	ES-SIDE-1-PILLAR-4-TOP	43	1.00	2.00	2.00	100
J4087-70	ES-SIDE-1-PILLAR-4-BOTT	44	1.00	2.00	2.00	100
J4087-72	ES-SIDE-1-PILLAR-5-TOP	45	1.00	2.00	2.00	100
J4087-74	ES-SIDE-1-PILLAR-5-BOTT	46	1.00	2.00	2.00	100
J4095-01	CON-SB-06-0-2	47	1.12	9.96	9.46	94.3
J4095-02	CON-SB-07-0-2	48	1.16	9.69	9.09	93
J4095-03	CON-BALLARDS-0-2	49	1.17	9.88	9.08	90.8
J4095-04	CON-SB-04-7-9	50	1.13	9.96	9.05	89.7
J4095-05	CON-SB-06-7-9	51	1.11	9.63	8.37	85.2
J4095-06	CON-SB-05-7-9	52	1.16	9.70	8.44	85.2
J4095-07	CON-SB-07-9-11	53	1.18	9.97	9.25	91.8
J4095-08	CON-SB-08-9-11	54	1.16	9.94	9.35	93.3
J4095-09	CON-SB-09-7-9	55	1.17	9.85	9.11	91.5
J4095-10	CON-SB-04-05-06-07-0-11	56	1.15	9.95	9.18	91.3
J4095-11	J4095-10MS	57	1.15	9.95	9.18	91.3
J4095-12	J4095-10MSD	58	1.15	9.95	9.18	91.3
J4095-13	CON-SB-104-05-06-07-0-11	59	1.13	9.50	8.70	90.4
J4096-01	DW-3	60	1.14	9.64	7.16	70.8
J4096-02	DW-4	61	1.16	9.51	7.69	78.2
J4096-03	DW-5	62	1.20	9.97	6.04	55.2
J4096-04	DW-6	63	1.19	9.98	7.02	66.3
J4096-05	DW-7	64	1.10	9.96	7.82	75.8
J4096-06	DW-8	65	1.16	9.86	7.38	71.5
J4096-07	DW-9	66	1.18	9.73	5.19	46.9
J4096-08	DW-10	67	1.15	9.94	8.11	79.2
J4096-09	DW-11	68	1.17	9.74	7.03	68.4
J4096-10	DW-12	69	1.13	9.98	8.70	85.5
J4101-01	TP-1	70	1.19	9.82	8.93	89.7
J4101-02	TP-1	71	1.19	9.82	8.93	89.7
J4101-03	TP-2	72	1.13	9.57	8.25	84.4
J4101-04	TP-2	73	1.13	9.57	8.25	84.4
J4102-01	TA-DRUM-1-OILY-STONE	74	1.00	2.00	2.00	100
J4102-02	TA-DRUM-5-OILY-WIPE	75	1.00	2.00	2.00	100
J4103-01	DPC-81753-FB-08	76	1.00	2.00	2.00	100
J4103-02	DPC-81753-FB-10	77	1.00	2.00	2.00	100
J4103-03	DPC-81753-FB-11	78	1.00	2.00	2.00	100



PERCENT SOLID

Analyst Name: jignesh
Date: 7/23/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 07/20/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 07/21/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
BalanceID: M Sc-1

QC: LB96799

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J4103-04	DPC-81753-FB-12	79	1.00	2.00	2.00	100
J4104-01	LA119SW005-S-180719	80	1.15	9.85	6.66	63.3
J4104-02	KY030LC026-S-180719	81	1.10	9.98	6.75	63.6
J4104-03	DUP-0493-180719	82	1.16	9.55	6.98	69.4
J4106-01	MOLA-BOULEVARD	83	1.18	9.74	9.05	91.9
J4106-02	MARKET-ST	84	1.19	9.80	8.86	89.1

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

LB 96799

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-072018

WorkList ID : 114692

Date : 7/20/2018 7:26:34 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2018	Solid	J4054-09	Percent Solids	Cool 4 deg C	PSEG01	K63	161-TP-1	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-06	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-3	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-12	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-4	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-18	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-5	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-24	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-8	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-30	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-9	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-36	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-11	07/18/2018	Chemtech -SO
07/22/2018	Solid	J4062-42	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-10	07/18/2018	Chemtech -SO
07/24/2018	Solid	J4081-05	Percent Solids	Cool 4 deg C	PSEG01	K63	RBR200029-GRO-B	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-01	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-6-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-03	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-6-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-05	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-7-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-07	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-7-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-09	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-8-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-11	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-8-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-13	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-9-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-15	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-9-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-17	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-1-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-19	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-1-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-21	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-2-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-23	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-2-BOTTO	07/19/2018	Chemtech -SO

Date/Time 07/20/18 5:00 PM

Date/Time 07/20/18 5:35 PM

Received by: J

Received by: J

Relinquished by: J

Relinquished by: J

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-072018

WorkList ID : 114692

Date : 7/20/2018 7:26:34 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Solid	J4087-25	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-3-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-27	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-3-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-29	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-4-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-31	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-4-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-33	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-5-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-35	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-5-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-37	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-6-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-39	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-6-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-41	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-7-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-43	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-7-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-45	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-8-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-47	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-8-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-49	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-9-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-51	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-2-PILLAR-9-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-54	Percent Solids	Cool 4 deg C	PSEG01	K63	NOTHING	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-56	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-1-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-58	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-1-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-60	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-2-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-62	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-2-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-64	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-3-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-66	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-3-BOTTO	07/19/2018	Chemtech -SO

Date/Time 07/20/18 4:00 PM

Received by: SO

Relinquished by: CB

Date/Time 07/20/18 5:35 PM

Received by: SO

Relinquished by: SO

1-9 96799

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-072018

WorkList ID : 114692

Date : 7/20/2018 7:26:34 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Solid	J4087-68	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-4-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-70	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-4-BOTTO	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-72	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-5-TOP	07/19/2018	Chemtech -SO
07/24/2018	Solid	J4087-74	Percent Solids	Cool 4 deg C	PSEG01	K63	ES-SIDE-1-PILLAR-5-BOTTO	07/19/2018	Chemtech -SO
07/28/2018	Solid	J4095-01	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-06-0-2	07/18/2018	Chemtech -SO
07/28/2018	Solid	J4095-02	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-07-0-2	07/18/2018	Chemtech -SO
07/28/2018	Solid	J4095-03	Percent Solids	Cool 4 deg C	PARS15	M33	CON-BALLARDS-0-2	07/18/2018	Chemtech -SO
07/29/2018	Solid	J4095-04	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-04-7-9	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-05	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-06-7-9	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-06	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-05-7-9	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-07	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-07-9-11	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-08	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-08-9-11	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-09	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-09-7-9	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-10	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-04-05-06-07-0-11	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-11	Percent Solids	Cool 4 deg C	PARS15	M33	J4095-10MS	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-12	Percent Solids	Cool 4 deg C	PARS15	M33	J4095-10MSD	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4095-13	Percent Solids	Cool 4 deg C	PARS15	M33	CON-SB-104-05-06-07-0-11	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-01	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-3	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-02	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-4	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-03	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-5	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-04	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-6	07/19/2018	Chemtech -SO

Date/Time 07/20/18 6:10 PM

Received by: SP

Relinquished by: SP

Date/Time 07/20/18 5:35 PM

Received by: SP

Relinquished by: SP

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-072018

WorkList ID : 114692

Date : 7/20/2018 7:26:34 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2018	Solid	J4096-05	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-7	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-06	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-8	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-07	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-9	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-08	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-10	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-09	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-11	07/19/2018	Chemtech -SO
07/30/2018	Solid	J4096-10	Percent Solids	Cool 4 deg C	ATCE02	K62	DW-12	07/19/2018	Chemtech -SO
07/25/2018	Solid	J4101-01	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-1	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4101-02	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-1	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4101-03	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-2	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4101-04	Percent Solids	Cool 4 deg C	PSEG01	K63	TP-2	07/20/2018	Chemtech -SO
07/27/2018	Solid	J4102-01	Percent Solids	Cool 4 deg C	PSEG01	K61	TA-DRUM-1-OILY-STONE	07/20/2018	Chemtech -SO
07/27/2018	Solid	J4102-02	Percent Solids	Cool 4 deg C	PSEG01	K61	TA-DRUM-5-OILY-WIPE	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4103-01	Percent Solids	Cool 4 deg C	PSEG01	K63	DPC-81753-FB-08	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4103-02	Percent Solids	Cool 4 deg C	PSEG01	K63	DPC-81753-FB-10	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4103-03	Percent Solids	Cool 4 deg C	PSEG01	K63	DPC-81753-FB-11	07/20/2018	Chemtech -SO
07/25/2018	Solid	J4103-04	Percent Solids	Cool 4 deg C	PSEG01	K63	DPC-81753-FB-12	07/20/2018	Chemtech -SO
07/29/2018	Solid	J4104-01	Percent Solids	Cool 4 deg C	SHAW22	M42	LA119SW005-S-180719	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4104-02	Percent Solids	Cool 4 deg C	SHAW22	M42	KY030LC026-S-180719	07/19/2018	Chemtech -SO
07/29/2018	Solid	J4104-03	Percent Solids	Cool 4 deg C	SHAW22	M42	DUP-0493-180719	07/19/2018	Chemtech -SO
07/25/2018	Solid	J4106-01	Percent Solids	Cool 4 deg C	ZUCC01	M41	MOLA-BOULEVARD	07/18/2018	Chemtech -SO
07/25/2018	Solid	J4106-02	Percent Solids	Cool 4 deg C	ZUCC01	M41	MARKET-ST	07/18/2018	Chemtech -SO

Date/Time 07/20/18 11:00 AM

Received by: JP

Relinquished by: JP

Date/Time 07/20/18 5:35 PM

Received by: JP

Relinquished by: JP