


7471B

LB 96800
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46578



A= 0.0000e+000

B= 2.1555e-004

C= -5.4466e-002

Rho= 0.9999133

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	266	0.000	266				
0.2	0.200	0.180	-0.020	1088	0.0 %	1088				
2.5	2.500	2.562	0.062	12138	0.0 %	12138				
5.0	5.000	4.917	-0.083	23064	0.0 %	23064				
7.5	7.500	7.553	0.053	35294	0.0 %	35294				
10.0	10.000	9.985	-0.015	46578	0.0 %	46578				

LB96800 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jul 2018 18:06:27

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	266	-	0.00007471B	PPB		20 Jul 2018 18:07:22	S	Std
0.2 - 1	50.2	1088	-	0.20007471B	PPB		20 Jul 2018 18:09:39	S	Std
2.5 - 1	52.5	12138	-	2.50007471B	PPB		20 Jul 2018 18:11:57	S	Std
5.0 - 1	55	23064	-	5.00007471B	PPB		20 Jul 2018 18:14:15	S	Std
7.5 - 1	57.5	35294	-	7.50007471B	PPB		20 Jul 2018 18:19:36	S	Std
10.0 - 1	510	46578	-	10.00007471B	PPB		20 Jul 2018 18:24:25	S	Std
ICV64 - 1	ICV64	17459	3.7088	-7471B	PPB		20 Jul 2018 18:27:05	U	SMPL
ICB64 - 1	ICB64	106	-0.0316	-7471B	PPB		20 Jul 2018 18:29:20	U	SMPL
CCV63 - 1	CCV63	22975	4.8977	-7471B	PPB		20 Jul 2018 18:31:36	U	SMPL
CCB63 - 1	CCB63	79	-0.0374	-7471B	PPB		20 Jul 2018 18:33:52	U	SMPL
CRA - 1	CRA	1067	0.1755	-7471B	PPB		20 Jul 2018 18:36:08	U	SMPL
HighStd - 1	HighStd	46540	9.9771	-7471B	PPB		20 Jul 2018 18:38:24	U	SMPL
ChkStd - 1	ChkStd	30563	6.5333	-7471B	PPB		20 Jul 2018 18:40:40	U	SMPL
PB111314BS - 1	LCSS	17706	3.7620	-7471B	PPB		20 Jul 2018 18:45:16	U	SMPL
J4087-01 - 1	ES-SIDE-1-PILLAR-6-TOP	1408	0.2490	-7471B	PPB		20 Jul 2018 18:47:33	U	SMPL
J4087-01DUP - 1	ES-SIDE-1-PILLAR-6-TOPDUP	1350	0.2365	-7471B	PPB		20 Jul 2018 18:49:50	U	SMPL
J4087-01MS - 1	ES-SIDE-1-PILLAR-6-TOPMS	18283	3.8864	-7471B	PPB		20 Jul 2018 18:52:07	U	SMPL
J4087-01MSD - 1	ES-SIDE-1-PILLAR-6-TOPMSD	18250	3.8793	-7471B	PPB		20 Jul 2018 18:54:25	U	SMPL
J4087-03 - 1	ES-SIDE-1-PILLAR-6-BOTTOM	1184	0.2007	-7471B	PPB		20 Jul 2018 18:56:41	U	SMPL
CCV64 - 1	CCV64	22860	4.8729	-7471B	PPB		20 Jul 2018 18:58:58	U	SMPL
CCB64 - 1	CCB64	97	-0.0336	-7471B	PPB		20 Jul 2018 19:01:14	U	SMPL
J4087-05 - 1	ES-SIDE-1-PILLAR-7-TOP	1135	0.1902	-7471B	PPB		20 Jul 2018 19:03:30	U	SMPL
J4087-07 - 1	ES-SIDE-1-PILLAR-7-BOTTOM	1376	0.2421	-7471B	PPB		20 Jul 2018 19:05:46	U	SMPL
J4087-09 - 1	ES-SIDE-1-PILLAR-8-TOP	1480	0.2645	-7471B	PPB		20 Jul 2018 19:08:02	U	SMPL
J4087-11 - 1	ES-SIDE-1-PILLAR-8-BOTTOM	1328	0.2318	-7471B	PPB		20 Jul 2018 19:10:18	U	SMPL
J4087-13 - 1	ES-SIDE-1-PILLAR-9-TOP	1197	0.2035	-7471B	PPB		20 Jul 2018 19:12:34	U	SMPL
J4087-15 - 1	ES-SIDE-1-PILLAR-9-BOTTOM	961	0.1527	-7471B	PPB		20 Jul 2018 19:14:51	U	SMPL
J4087-17 - 1	ES-SIDE-2-PILLAR-1-TOP	1368	0.2404	-7471B	PPB		20 Jul 2018 19:17:08	U	SMPL
J4087-19 - 1	ES-SIDE-2-PILLAR-1-BOTTOM	1248	0.2145	-7471B	PPB		20 Jul 2018 19:19:26	U	SMPL
J4087-21 - 1	ES-SIDE-2-PILLAR-2-TOP	304	0.0111	-7471B	PPB		20 Jul 2018 19:21:43	U	SMPL
J4087-23 - 1	ES-SIDE-2-PILLAR-2-BOTTOM	248	-0.0010	-7471B	PPB		20 Jul 2018 19:23:59	U	SMPL
CCV65 - 1	CCV65	22891	4.8796	-7471B	PPB		20 Jul 2018 19:26:15	U	SMPL
CCB65 - 1	CCB65	108	-0.0312	-7471B	PPB		20 Jul 2018 19:28:32	U	SMPL
J4087-25 - 1	ES-SIDE-2-PILLAR-3-TOP	740	0.1050	-7471B	PPB		20 Jul 2018 19:30:48	U	SMPL
J4087-27 - 1	ES-SIDE-2-PILLAR-3-BOTTOM	1232	0.2111	-7471B	PPB		20 Jul 2018 19:33:05	U	SMPL
J4087-29 - 1	ES-SIDE-2-PILLAR-4-TOP	1227	0.2100	-7471B	PPB		20 Jul 2018 19:35:21	U	SMPL
J4087-31 - 1	ES-SIDE-2-PILLAR-4-BOTTOM	1204	0.2051	-7471B	PPB		20 Jul 2018 19:37:37	U	SMPL
J4087-33 - 1	ES-SIDE-2-PILLAR-5-TOP	224	-0.0062	-7471B	PPB		20 Jul 2018 19:39:54	U	SMPL
J4087-35 - 1	ES-SIDE-2-PILLAR-5-BOTTOM	1433	0.2544	-7471B	PPB		20 Jul 2018 19:42:11	U	SMPL
J4087-37 - 1	ES-SIDE-2-PILLAR-6-TOP	833	0.1251	-7471B	PPB		20 Jul 2018 19:44:29	U	SMPL
J4087-39 - 1	ES-SIDE-2-PILLAR-6-BOTTOM	1022	0.1658	-7471B	PPB		20 Jul 2018 19:46:46	U	SMPL
PB111315BL - 1	PB111315BL	-22	-0.0592	-7471B	PPB		20 Jul 2018 19:49:04	U	SMPL
PB111315BS - 1	PB111315BS	17118	3.6353	-7471B	PPB		20 Jul 2018 19:51:20	U	SMPL
CCV66 - 1	CCV66	22657	4.8292	-7471B	PPB		20 Jul 2018 19:53:37	U	SMPL
CCB66 - 1	CCB66	74	-0.0385	-7471B	PPB		20 Jul 2018 19:55:53	U	SMPL
J4087-41 - 1	ES-SIDE-2-PILLAR-7-TOP	796	0.1171	-7471B	PPB		20 Jul 2018 20:03:00	U	SMPL
J4087-41DUP - 1	ES-SIDE-2-PILLAR-7-TOPDUP	775	0.1126	-7471B	PPB		20 Jul 2018 20:05:16	U	SMPL
J4087-41MS - 1	ES-SIDE-2-PILLAR-7-TOPMS	17741	3.7696	-7471B	PPB		20 Jul 2018 20:07:32	U	SMPL
J4087-41MSD - 1	ES-SIDE-2-PILLAR-7-TOPMSD	16738	3.5534	-7471B	PPB		20 Jul 2018 20:09:48	U	SMPL
J4087-43 - 1	ES-SIDE-2-PILLAR-7-BOTTOM	1448	0.2576	-7471B	PPB		20 Jul 2018 20:12:05	U	SMPL
J4087-45 - 1	ES-SIDE-2-PILLAR-8-TOP	1476	0.2637	-7471B	PPB		20 Jul 2018 20:14:21	U	SMPL
J4087-47 - 1	ES-SIDE-2-PILLAR-8-BOTTOM	335	0.0177	-7471B	PPB		20 Jul 2018 20:16:39	U	SMPL
J4087-49 - 1	ES-SIDE-2-PILLAR-9-TOP	150	-0.0221	-7471B	PPB		20 Jul 2018 20:18:56	U	SMPL
J4087-51 - 1	ES-SIDE-2-PILLAR-9-BOTTOM	268	0.0033	-7471B	PPB		20 Jul 2018 20:21:14	U	SMPL
PB111314BL - 1	PBS	-19	-0.0586	-7471B	PPB		20 Jul 2018 20:23:30	U	SMPL
CCV67 - 1	CCV67	22954	4.8932	-7471B	PPB		20 Jul 2018 20:25:47	U	SMPL
CCB67 - 1	CCB67	98	-0.0333	-7471B	PPB		20 Jul 2018 20:28:03	U	SMPL
J4087-56 - 1	ES-SIDE-1-PILLAR-1-TOP	1505	0.2699	-7471B	PPB		20 Jul 2018 20:30:21	U	SMPL
J4087-58 - 1	ES-SIDE-1-PILLAR-1-BOTTOM	1149	0.1932	-7471B	PPB		20 Jul 2018 20:32:37	U	SMPL
J4087-60 - 1	ES-SIDE-1-PILLAR-2-TOP	1511	0.2712	-7471B	PPB		20 Jul 2018 20:34:53	U	SMPL
J4087-62 - 1	ES-SIDE-1-PILLAR-2-BOTTOM	1507	0.2704	-7471B	PPB		20 Jul 2018 20:37:09	U	SMPL
J4087-64 - 1	ES-SIDE-1-PILLAR-3-TOP	1191	0.2023	-7471B	PPB		20 Jul 2018 20:39:26	U	SMPL
J4087-66 - 1	ES-SIDE-1-PILLAR-3-BOTTOM	641	0.0837	-7471B	PPB		20 Jul 2018 20:41:43	U	SMPL
J4087-68 - 1	ES-SIDE-1-PILLAR-4-TOP	1902	0.3555	-7471B	PPB		20 Jul 2018 20:44:00	U	SMPL
J4087-70 - 1	ES-SIDE-1-PILLAR-4-BOTTOM	410	0.0339	-7471B	PPB		20 Jul 2018 20:46:18	U	SMPL
J4087-72 - 1	ES-SIDE-1-PILLAR-5-TOP	2119	0.4023	-7471B	PPB		20 Jul 2018 20:48:36	U	SMPL
J4087-74 - 1	ES-SIDE-1-PILLAR-5-BOTTOM	1545	0.2786	-7471B	PPB		20 Jul 2018 20:50:52	U	SMPL
CCV68 - 1	CCV68	22931	4.8882	-7471B	PPB		20 Jul 2018 20:53:09	U	SMPL
CCB68 - 1	CCB68	104	-0.0320	-7471B	PPB		20 Jul 2018 20:55:26	U	SMPL
PB111316BL - 1	PBS	38	-0.0463	-7471B	PPB		20 Jul 2018 20:57:43	U	SMPL
PB111316BS - 1	LCSS	15967	3.3872	-7471B	PPB		20 Jul 2018 21:00:00	U	SMPL
J3820-05 - 1	BU-01-071918	3467	0.6928	-7471B	PPB		20 Jul 2018 21:02:17	U	SMPL
J3827-03 - 1	OR-01-071918	786	0.1150	-7471B	PPB		20 Jul 2018 21:04:33	U	SMPL
J4070-01 - 1	KY030LC011-S-180718	2599	0.5057	-7471B	PPB		20 Jul 2018 21:06:50	U	SMPL
J4070-02 - 1	KY030LC015-S-180718	1484	0.2654	-7471B	PPB		20 Jul 2018 21:09:07	U	SMPL
J4070-03 - 1	KY030LC018-S-180718	539	0.0617	-7471B	PPB		20 Jul 2018 21:11:24	U	SMPL
J4071-01 - 1	GL-WC-S-04	261986	56.4158	-7471B	PPB		20 Jul 2018 21:13:42	U	SMPL
J4081-01 - 1	RBR200029	2938	0.5788	-7471B	PPB		20 Jul 2018 21:16:00	U	SMPL
J4090-01 - 1	FK-PS-01-034	456	0.0438	-7471B	PPB		20 Jul 2018 21:18:31	U	SMPL
CCV69 - 1	CCV69	23147	4.9348	-7471B	PPB		20 Jul 2018 21:20:48	U	SMPL
CCB69 - 1	CCB69	165	-0.0189	-7471B	PPB		20 Jul 2018 21:25:46	U	SMPL
J4095-03 - 1	CON-BALLARDS-0-2	17480	3.7133	-7471B	PPB		20 Jul 2018 21:28:03	U	SMPL
J4095-10 - 1	CON-SB-04-05-06-07-0-11	5705	1.1752	-7471B	PPB		20 Jul 2018 21:30:19	U	SMPL
J4095-10DUP - 1	CON-SB-04-05-06-07-0-11DUP	5225	1.0718	-7471B	PPB		20 Jul 2018 21:32:36	U	SMPL
J4095-11 - 1	J4095-10MS	25823	5.5116	-7471B	PPB		20 Jul 2018 21:39:49	U	SMPL
J4095-12 - 1	J4095-10MSD	23300	4.9678	-7471B	PPB		20 Jul 2018 21:42:06	U	SMPL
J4095-13 - 1	CON-SB-104-05-06-07-0-11	5319	1.0920	-7471B	PPB		20 Jul 2018 21:44:23	U	SMPL
CCV70 - 1	CCV70	23019	4.9072	-7471B	PPB		20 Jul 2018 21:46:40	U	SMPL
CCB70 - 1	CCB70	-130	-0.0825	-7471B	PPB		20 Jul 2018 21:48:58	U	SMPL
J4087-01LX5 - 1		288	0.0076	-7471B	PPB		20 Jul 2018 21:51:17	U	SMPL
J4087-01A - 1		19029	4.0472	-7471B	PPB		20 Jul 2018 21:53:34	U	SMPL
J4087-41LX5 - 1		76	-0.0381	-7471B	PPB		20 Jul 2018 21:55:51	U	SMPL
J4087-41A - 1		18185	3.8653	-7471B	PPB		20 Jul 2018 21:58:08	U	SMPL

LB96800 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jul 2018 18:06:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J4095-10LX5 - 1		1151	0.1936		-7471B	PPB	20 Jul 2018 22:00:25	U	SMPL
J4095-10A - 1		23305	4.9689		-7471B	PPB	20 Jul 2018 22:02:42	U	SMPL
J4071-01DLX10 - 1		29191	6.2376		-7471B	PPB	20 Jul 2018 22:04:59	U	SMPL
CCV71 - 1	CCV71	23015	4.9063		-7471B	PPB	20 Jul 2018 22:07:17	U	SMPL
CCB71 - 1	CCB71	-5	-0.0555		-7471B	PPB	20 Jul 2018 22:09:35	U	SMPL