



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96807

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	5	27.0	80.6	80.9	07/23/2018	10:20
2	J4102-03		1	6	>100.0	>212.0	>212.0	07/23/2018	10:45
3	J4102-03DUP		1	6	>100.0	>212.0	>212.0	07/23/2018	11:20
4	J4102-04		1	6	>100.0	>212.0	>212.0	07/23/2018	11:50
5	J4102-05		1	7	>100.0	>212.0	>212.0	07/23/2018	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96807

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP72318

WorkList ID : 114773

Date : 7/23/2018 9:36:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/27/2018	Water	J4102-03	Flash Point	Cool 4 deg C	PSEG01	K61	TA-DRUM-2-OILY-WATER	07/20/2018	1010A
07/27/2018	Water	J4102-04	Flash Point	Cool 4 deg C	PSEG01	K61	TA-DRUM-3-OILY-WATER	07/20/2018	1010A
07/27/2018	Water	J4102-05	Flash Point	Cool 4 deg C	PSEG01	K61	TA-DRUM-4-OILY-WATER	07/20/2018	1010A

Date/Time 7/23/18 9:45 AM
 Received by: KS
 Relinquished by: CP

Date/Time 7/23/18 12:25 PM
 Received by: CP
 Relinquished by: KS