


7470A

LB96811
INSTRUMENT ID: CV2

Linear

 μ Abs.:

38618



A= 0.0000e+000

B= 2.5941e-004

C= -2.2852e-002

Rho= 0.9999104

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	101	0.000	101				
0.2	0.200	0.225	0.025	956	0.0 %	956				
2.5	2.500	2.507	0.007	9754	0.0 %	9754				
5.0	5.000	4.903	-0.097	18987	0.0 %	18987				
7.5	7.500	7.566	0.066	29256	0.0 %	29256				
10.0	10.000	9.995	-0.005	38618	0.0 %	38618				

LB96811 INSDTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jul 2018 14:01:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	101	-	0.00007470A	PPB		23 Jul 2018 14:15:36	S	Std
0.2 - 1	50.2	956	-	0.20007470A	PPB		23 Jul 2018 14:17:45	S	Std
2.5 - 1	52.5	9754	-	2.50007470A	PPB		23 Jul 2018 14:19:53	S	Std
5.0 - 1	55	18987	-	5.00007470A	PPB		23 Jul 2018 14:22:01	S	Std
7.5 - 1	57.5	29256	-	7.50007470A	PPB		23 Jul 2018 14:24:10	S	Std
10.0 - 1	59.1	38618	-	10.00007470A	PPB		23 Jul 2018 14:26:18	S	Std
ICV65 - 1	ICV65	15171	3.9127	-7470A	PPB		23 Jul 2018 14:29:01	U	SMPL
ICB65 - 1	ICB65	-21	-0.0283	-7470A	PPB		23 Jul 2018 14:31:07	U	SMPL
CCV72 - 1	CCV72	19469	5.0276	-7470A	PPB		23 Jul 2018 14:33:13	U	SMPL
CCB72 - 1	CCB72	-25	-0.0293	-7470A	PPB		23 Jul 2018 14:35:20	U	SMPL
CRA - 1	CRA	941	0.2213	-7470A	PPB		23 Jul 2018 14:37:27	U	SMPL
HighStd - 1	HighStd	38511	9.9673	-7470A	PPB		23 Jul 2018 14:39:33	U	SMPL
ChkStd - 1	ChkStd	26479	6.8461	-7470A	PPB		23 Jul 2018 14:41:40	U	SMPL
PB111346BL - 1	PBW	115	0.0070	-7470A	PPB		23 Jul 2018 14:43:48	U	SMPL
PB111346BS - 1	LCSW	14411	3.7155	-7470A	PPB		23 Jul 2018 14:45:59	U	SMPL
PB111346TB - 1	PB111288TB	189	0.0262	-7470A	PPB		23 Jul 2018 14:48:06	U	SMPL
J3820-06 - 1	BU-01-071918	277	0.0490	-7470A	PPB		23 Jul 2018 14:50:15	U	SMPL
J3827-04 - 1	OR-01-071918	235	0.0381	-7470A	PPB		23 Jul 2018 14:52:23	U	SMPL
J3827-04DUP - 1	OR-01-071918DUP	260	0.0446	-7470A	PPB		23 Jul 2018 14:54:31	U	SMPL
J3827-04MS - 1	OR-01-071918MS	14687	3.7871	-7470A	PPB		23 Jul 2018 14:56:38	U	SMPL
CCV73 - 1	CCV73	19597	5.0608	-7470A	PPB		23 Jul 2018 14:58:45	U	SMPL
CCB73 - 1	CCB73	-37	-0.0325	-7470A	PPB		23 Jul 2018 15:00:52	U	SMPL
J3827-04MSD - 1	OR-01-071918MSD	13941	3.5936	-7470A	PPB		23 Jul 2018 15:03:00	U	SMPL
J4062-06 - 1	TP-3	482	0.1022	-7470A	PPB		23 Jul 2018 15:05:06	U	SMPL
J4062-12 - 1	TP-4	440	0.0913	-7470A	PPB		23 Jul 2018 15:07:13	U	SMPL
J4062-18 - 1	TP-5	439	0.0910	-7470A	PPB		23 Jul 2018 15:09:20	U	SMPL
J4062-24 - 1	TP-8	386	0.0773	-7470A	PPB		23 Jul 2018 15:11:27	U	SMPL
J4062-30 - 1	TP-9	216	0.0332	-7470A	PPB		23 Jul 2018 15:13:35	U	SMPL
J4062-36 - 1	TP-11	207	0.0308	-7470A	PPB		23 Jul 2018 15:15:43	U	SMPL
J4062-42 - 1	TP-10	231	0.0371	-7470A	PPB		23 Jul 2018 15:17:52	U	SMPL
J4068-01 - 1	GREENBROOK	262	0.0451	-7470A	PPB		23 Jul 2018 15:20:00	U	SMPL
J4079-01 - 1	WHITE-BAG-CHIPS	420	0.0861	-7470A	PPB		23 Jul 2018 15:22:06	U	SMPL
CCV74 - 1	CCV74	19257	4.9726	-7470A	PPB		23 Jul 2018 15:24:13	U	SMPL
CCB74 - 1	CCB74	76	-0.0031	-7470A	PPB		23 Jul 2018 15:26:20	U	SMPL
J4079-02 - 1	BLUE-BAG-DEBRIS	239	0.0391	-7470A	PPB		23 Jul 2018 15:28:28	U	SMPL
J4079-03 - 1	OIL-DRUM	143	0.0142	-7470A	PPB		23 Jul 2018 15:30:35	U	SMPL
J4081-03 - 1	RBR200029	135	0.0122	-7470A	PPB		23 Jul 2018 15:32:41	U	SMPL
J4087-53 - 1	ES-PILLAR-3-4	101	0.0033	-7470A	PPB		23 Jul 2018 15:34:49	U	SMPL
J3827-04LX5 - 1		140	0.0135	-7470A	PPB		23 Jul 2018 15:36:56	U	SMPL
J3827-04A - 1		15259	3.9355	-7470A	PPB		23 Jul 2018 15:39:04	U	SMPL
CCV75 - 1	CCV75	18905	4.8813	-7470A	PPB		23 Jul 2018 15:41:12	U	SMPL
CCB75 - 1	CCB75	5	-0.0216	-7470A	PPB		23 Jul 2018 15:43:20	U	SMPL