



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96816

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	07/23/2018	15:30
2	CAL2	1	Water			22.7	7.00	07/23/2018	15:32
3	CAL3	1	Water			22.9	10.03	07/23/2018	15:34
4	ICV	1	Water			22.7	7.00	07/23/2018	15:36
5	CCV1	1	Water			22.8	2.00	07/23/2018	15:38
6	J4110-03	1	Solid	20.02	20	24.2	7.51	07/23/2018	15:40
7	J4110-03DUP	1	Solid	20.02	20	24.3	7.52	07/23/2018	15:42
8	J4119-05	1	Solid	20.04	20	24.7	5.96	07/23/2018	15:44
9	CCV2	1	Water			22.7	12.00	07/23/2018	15:46

WORKLIST(Hardcopy Internal Chain)

LB96816

WorkList Name : corrosivity072318

WorkList ID : 114786

Date : 7/23/2018 2:59:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2018	Solid	J4110-03	Corrosivity	Cool 4 deg C	PSEG01	K63	GS-RO-338-RO-147-RO-278	07/23/2018	9045C
07/23/2018	Solid	J4119-05	Corrosivity	Cool 4 deg C	PSEG01	K63	GB-SOIL-PILE	07/23/2018	9045C

Date/Time 07/23/18 3:20 PM
 Received by: SP
 Relinquished by: CP

Date/Time 07/23/18 4:00
 Received by: CP
 Relinquished by: SP