

LB 96820


7471B

INSTRUMENT ID: CV4

Linear

μ Abs.:

43034



A= 0.0000e+000

B= 2.3389e-004

C= -3.7813e-002

Rho= 0.9998961

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	95	0.000	95				
0.2	0.200	0.180	-0.020	932	0.0 %	932				
2.5	2.500	2.602	0.102	11288	0.0 %	11288				
5.0	5.000	4.934	-0.066	21258	0.0 %	21258				
7.5	7.500	7.472	-0.028	32108	0.0 %	32108				
10.0	10.000	10.027	0.027	43034	0.0 %	43034				

LB96820 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Jul 2018 16:45:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	95	-	0.0000	7471B	PPB	23 Jul 2018 16:48:14	S	Std
0.2 - 1	50.2	932	-	0.2000	7471B	PPB	23 Jul 2018 16:50:31	S	Std
2.5 - 1	52.5	11288	-	2.5000	7471B	PPB	23 Jul 2018 16:52:49	S	Std
5.0 - 1	55	21258	-	5.0000	7471B	PPB	23 Jul 2018 16:55:06	S	Std
7.5 - 1	57.5	32108	-	7.5000	7471B	PPB	23 Jul 2018 16:57:24	S	Std
10.0 - 1	510	43034	-	10.0000	7471B	PPB	23 Jul 2018 16:59:40	S	Std
ICV66 - 1	ICV66	16849	3.9029	-	7471B	PPB	23 Jul 2018 17:02:27	U	SMPL
ICB66 - 1	ICB66	-51	-0.0497	-	7471B	PPB	23 Jul 2018 17:04:43	U	SMPL
CCV76 - 1	CCV76	22594	5.2466	-	7471B	PPB	23 Jul 2018 17:06:58	U	SMPL
CCB76 - 1	CCB76	-47	-0.0488	-	7471B	PPB	23 Jul 2018 17:09:14	U	SMPL
CRA - 1	CRA	879	0.1678	-	7471B	PPB	23 Jul 2018 17:11:30	U	SMPL
HighStd - 1	HighStd	44091	10.2744	-	7471B	PPB	23 Jul 2018 17:13:46	U	SMPL
ChkStd - 1	ChkStd	30169	7.0183	-	7471B	PPB	23 Jul 2018 17:16:02	U	SMPL
PB111348BL - 1	PBS	-84	-0.0575	-	7471B	PPB	23 Jul 2018 17:18:20	U	SMPL
PB111348BS - 1	LCSS	15111	3.4964	-	7471B	PPB	23 Jul 2018 17:20:39	U	SMPL
J4101-01 - 1	TP-1	502	0.0796	-	7471B	PPB	23 Jul 2018 17:22:55	U	SMPL
J4101-01DUP - 1	TP-1DUP	451	0.0677	-	7471B	PPB	23 Jul 2018 17:25:13	U	SMPL
J4101-01MS - 1	TP-1MS	16386	3.7946	-	7471B	PPB	23 Jul 2018 17:27:30	U	SMPL
J4101-01MSD - 1	TP-1MSD	15962	3.6955	-	7471B	PPB	23 Jul 2018 17:29:48	U	SMPL
J4101-03 - 1	TP-2	771	0.1425	-	7471B	PPB	23 Jul 2018 17:32:05	U	SMPL
CCV77 - 1	CCV77	23021	5.3465	-	7471B	PPB	23 Jul 2018 17:34:21	U	SMPL
CCB77 - 1	CCB77	-81	-0.0568	-	7471B	PPB	23 Jul 2018 17:36:36	U	SMPL
J4104-01 - 1	LA119SW005-S-180719	1240	0.2522	-	7471B	PPB	23 Jul 2018 17:38:53	U	SMPL
J4104-02 - 1	KY030LC026-S-180719	1232	0.2503	-	7471B	PPB	23 Jul 2018 17:41:09	U	SMPL
J4104-03 - 1	DUP-0493-180719	794	0.1479	-	7471B	PPB	23 Jul 2018 17:43:25	U	SMPL
J4106-01 - 1	MOLA-BOULEVARD	2357	0.5135	-	7471B	PPB	23 Jul 2018 17:45:41	U	SMPL
J4106-02 - 1	MARKET-ST	8020	1.8379	-	7471B	PPB	23 Jul 2018 17:47:58	U	SMPL
J4109-01 - 1	GW-04(3-5)	37195	8.6616	-	7471B	PPB	23 Jul 2018 17:50:14	U	SMPL
J4109-02 - 1	GW-04(10-12)	19266	4.4682	-	7471B	PPB	23 Jul 2018 17:52:31	U	SMPL
J4109-03 - 1	SB-12(3-5)	23900	5.5521	-	7471B	PPB	23 Jul 2018 17:54:50	U	SMPL
J4109-04 - 1	SB-12(8-10)	3658	0.8177	-	7471B	PPB	23 Jul 2018 17:57:08	U	SMPL
J4109-05 - 1	SB-11(3-5)	14780	3.4190	-	7471B	PPB	23 Jul 2018 17:59:25	U	SMPL
CCV78 - 1	CCV78	23245	5.3989	-	7471B	PPB	23 Jul 2018 18:01:41	U	SMPL
CCB78 - 1	CCB78	-90	-0.0589	-	7471B	PPB	23 Jul 2018 18:03:57	U	SMPL
J4109-06 - 1	SB-11(8-10)	20279	4.7052	-	7471B	PPB	23 Jul 2018 18:06:15	U	SMPL
J4109-07 - 1	SB-10(3-5)	8857	2.0337	-	7471B	PPB	23 Jul 2018 18:08:31	U	SMPL
J4109-08 - 1	SB-10(10-12)	4122	0.9263	-	7471B	PPB	23 Jul 2018 18:10:48	U	SMPL
J4101-01LX5 - 1		83	-0.0184	-	7471B	PPB	23 Jul 2018 18:13:05	U	SMPL
J4101-01A - 1		17841	4.1349	-	7471B	PPB	23 Jul 2018 18:15:22	U	SMPL
CCV79 - 1	CCV79	23304	5.4127	-	7471B	PPB	23 Jul 2018 18:17:39	U	SMPL
CCB79 - 1	CCB79	-93	-0.0596	-	7471B	PPB	23 Jul 2018 18:19:56	U	SMPL