


ISM02.4

LB96834
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42953



A= 0.0000e+000

B= 2.3400e-004

C= -7.5658e-002

Rho= 0.9999571

Accept=Accepted

Slope
 $y = 1.7125 \times 10^{-4}x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	-0.002	-0.002	316	0.000	316					0
0.2	0.200	0.212	0.012	1231	0.0 %	1231					6
2.5	2.500	2.497	-0.003	10993	0.0 %	10993					0
5.0	5.000	4.954	-0.046	21493	0.0 %	21493					-1
7.5	7.500	7.564	0.064	32648	0.0 %	32648					1
10.0	10.000	9.975	-0.025	42953	0.0 %	42953					0

n63
7/24/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	316	-		0 ISM02.4	PPB	7/24/2018 13:55	Std
	0.2 S0.2	1231	-		0.2 ISM02.4	PPB	7/24/2018 13:58	Std
	2.5 S2.5	10993	-		2.5 ISM02.4	PPB	7/24/2018 14:00	Std
	5 S5	21493	-		5 ISM02.4	PPB	7/24/2018 14:02	Std
	7.5 S7.5	32648	-		7.5 ISM02.4	PPB	7/24/2018 14:05	Std
	10 S10.0	42953	-		10 ISM02.4	PPB	7/24/2018 14:07	Std
ICV	ICV	17312	3.9753	-	ISM02.4	PPB	7/24/2018 14:10	SMPL
ICB	ICB	172	-0.0354	-	ISM02.4	PPB	7/24/2018 14:12	SMPL
CCV091	CCV091	21490	4.9529	-	ISM02.4	PPB	7/24/2018 14:14	SMPL
CCB091	CCB091	217	-0.0249	-	ISM02.4	PPB	7/24/2018 14:17	SMPL
PB111379BL	PBW010	339	0.0037	-	ISM02.4	PPB	7/24/2018 14:21	SMPL
PB111379TB	LEB010	145	-0.0417	-	ISM02.4	PPB	7/24/2018 14:24	SMPL
J4075-01	MJJ1T5	29	-0.0689	-	ISM02.4	PPB	7/24/2018 14:26	SMPL
J4075-02	MJJ1T6	25	-0.0698	-	ISM02.4	PPB	7/24/2018 14:28	SMPL
J4075-03	MJJ1T7	-37	-0.0843	-	ISM02.4	PPB	7/24/2018 14:30	SMPL
J4075-04	MJJ1T8	265	-0.0136	-	ISM02.4	PPB	7/24/2018 14:33	SMPL
J4075-05	MJJ1T9	-105	-0.1002	-	ISM02.4	PPB	7/24/2018 14:35	SMPL
J4075-06	MJJ1T9D	-61	-0.0899	-	ISM02.4	PPB	7/24/2018 14:37	SMPL
J4075-07	MJJ1T9S	4784	1.0438	-	ISM02.4	PPB	7/24/2018 14:40	SMPL
CCV092	CCV092	21460	4.9459	-	ISM02.4	PPB	7/24/2018 14:42	SMPL
CCB092	CCB092	152	-0.0401	-	ISM02.4	PPB	7/24/2018 14:44	SMPL