


L396836

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	34	0.000	34				
0.2	0.200	0.201	0.001	990	0.0 %	990				
2.5	2.500	2.472	-0.028	12069	0.0 %	12069				
5.0	5.000	5.048	0.048	24635	0.0 %	24635				
7.5	7.500	7.468	-0.032	36443	0.0 %	36443				
10.0	10.000	10.006	0.006	48823	0.0 %	48823				

LB96836 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jul 2018 15:59:34

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	24 Jul 2018 16:03:34	S	Std
0.2 - 1	50.2	990	-	0.2000	7471B	PPB	24 Jul 2018 16:05:52	S	Std
2.5 - 1	52.5	12069	-	2.5000	7471B	PPB	24 Jul 2018 16:08:09	S	Std
5.0 - 1	55	24635	-	5.0000	7471B	PPB	24 Jul 2018 16:10:26	S	Std
7.5 - 1	57.5	36443	-	7.5000	7471B	PPB	24 Jul 2018 16:12:43	S	Std
10.0 - 1	59	48823	-	10.0000	7471B	PPB	24 Jul 2018 16:15:02	S	Std
ICV68 - 1	ICV68	19188	3.9312	-	7471B	PPB	24 Jul 2018 16:17:41	U	SMPL
ICB68 - 1	ICB68	-117	-0.0262	-	7471B	PPB	24 Jul 2018 16:19:56	U	SMPL
CCV93 - 1	CCV93	24701	5.0614	-	7471B	PPB	24 Jul 2018 16:22:12	U	SMPL
CCB93 - 1	CCB93	-87	-0.0201	-	7471B	PPB	24 Jul 2018 16:24:28	U	SMPL
CRA - 1	CRA	933	0.1890	-	7471B	PPB	24 Jul 2018 16:26:43	U	SMPL
HighStd - 1	HighStd	48459	9.9317	-	7471B	PPB	24 Jul 2018 16:28:59	U	SMPL
ChkStd - 1	ChkStd	35297	7.2335	-	7471B	PPB	24 Jul 2018 16:31:15	U	SMPL
PB111395BL - 1	PBS	-180	-0.0391	-	7471B	PPB	24 Jul 2018 16:33:33	U	SMPL
PB111395BS - 1	LCSS	18161	3.7207	-	7471B	PPB	24 Jul 2018 16:35:50	U	SMPL
J4098-01 - 1	GL-TOPSOIL-01	3599	0.7356	-	7471B	PPB	24 Jul 2018 16:38:07	U	SMPL
J4098-02 - 1	GL-TOPSOIL-02	2201	0.4490	-	7471B	PPB	24 Jul 2018 16:40:25	U	SMPL
J4098-03 - 1	GL-TOPSOIL-03	3896	0.7964	-	7471B	PPB	24 Jul 2018 16:42:43	U	SMPL
J4098-03DUP - 1	GL-TOPSOIL-03DUP	3880	0.7932	-	7471B	PPB	24 Jul 2018 16:47:41	U	SMPL
J4098-04 - 1	J4098-03MS	21448	4.3945	-	7471B	PPB	24 Jul 2018 16:49:57	U	SMPL
CCV94 - 1	CCV94	24449	5.0097	-	7471B	PPB	24 Jul 2018 16:52:12	U	SMPL
CCB94 - 1	CCB94	-138	-0.0305	-	7471B	PPB	24 Jul 2018 16:54:29	U	SMPL
J4098-05 - 1	J4098-03MSD	20766	4.2547	-	7471B	PPB	24 Jul 2018 16:56:45	U	SMPL
J4110-01 - 1	GS-RO-338-RO-147-RO-278	1370	0.2786	-	7471B	PPB	24 Jul 2018 16:59:01	U	SMPL
J4112-02 - 1	CON-SB-08-09-0-11	3631	0.7421	-	7471B	PPB	24 Jul 2018 17:01:19	U	SMPL
J4113-01 - 1	KY030LC058-S-180720	1076	0.2184	-	7471B	PPB	24 Jul 2018 17:03:35	U	SMPL
J4113-02 - 1	KY030LC066-S-180720	587	0.1181	-	7471B	PPB	24 Jul 2018 17:05:52	U	SMPL
J4119-01 - 1	GB-SOIL-PILE	375	0.0747	-	7471B	PPB	24 Jul 2018 17:08:09	U	SMPL
J4121-01 - 1	SB-9(3-5)	21862	4.4794	-	7471B	PPB	24 Jul 2018 17:10:26	U	SMPL
J4121-02 - 1	SB-9(13-15)	8922	1.8268	-	7471B	PPB	24 Jul 2018 17:12:44	U	SMPL
J4121-03 - 1	SB-8(3-5)	21002	4.3031	-	7471B	PPB	24 Jul 2018 17:15:02	U	SMPL
J4121-04 - 1	SB-8(10-12)	12452	2.5504	-	7471B	PPB	24 Jul 2018 17:17:17	U	SMPL
CCV95 - 1	CCV95	24650	5.0509	-	7471B	PPB	24 Jul 2018 17:19:34	U	SMPL
CCB95 - 1	CCB95	-117	-0.0262	-	7471B	PPB	24 Jul 2018 17:21:50	U	SMPL
J4121-05 - 1	GW-03(3-5)	13423	2.7494	-	7471B	PPB	24 Jul 2018 17:24:06	U	SMPL
J4121-06 - 1	GW-03(8-10)	74007	15.1690	-	7471B	PPB	24 Jul 2018 17:26:23	U	SMPL
J4121-07 - 1	SB-7(3-5)	14443	2.9585	-	7471B	PPB	24 Jul 2018 17:28:39	U	SMPL
J4121-08 - 1	SB-7(8-10)	6814	1.3946	-	7471B	PPB	24 Jul 2018 17:30:57	U	SMPL
J4121-09 - 1	DUP-072318	14766	3.0248	-	7471B	PPB	24 Jul 2018 17:33:14	U	SMPL
J4098-03LX5 - 1		834	0.1687	-	7471B	PPB	24 Jul 2018 17:35:31	U	SMPL
J4098-03A - 1		22542	4.6188	-	7471B	PPB	24 Jul 2018 17:37:48	U	SMPL
J4121-06DLX2 - 1		37037	7.5902	-	7471B	PPB	24 Jul 2018 17:40:06	U	SMPL
CCV96 - 1	CCV96	24456	5.0112	-	7471B	PPB	24 Jul 2018 17:42:23	U	SMPL
CCB96 - 1	CCB96	-156	-0.0342	-	7471B	PPB	24 Jul 2018 17:44:41	U	SMPL