

21396839

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/24/2018 16:28

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.228	0.0	0.091	
ICB1	0.072	0.0	0.004	
CCV1	241.361	0.0	0.222	
CCB1	-0.080	0.0	0.004	
PB111413BL	-0.605	0.0	0.003	
PB111413BS	98.003	0.0	0.093	
MIDPB111413	246.816	0.0	0.227	
J4080-30	392.160	0.0	0.358	
CCV2	246.826	0.0	0.227	
CCB2	0.021	0.0	0.004	

N	10
Mean	132.080
SD	140.7515
CV%	106.57

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Jul 24 16:24:37 2018

Tue Jul 24 16:28:01 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.7386	µg/l	7/24/2018 11:34:48	
5.0PPBCN	A	Total CN	P	4.138	µg/l	7/24/2018 11:34:49	
10PPBCN	A	Total CN	P	9.1804	µg/l	7/24/2018 11:34:50	
50PPBCN	A	Total CN	P	49.321	µg/l	7/24/2018 11:34:51	
100PPBCN	A	Total CN	P	101.1707	µg/l	7/24/2018 11:34:52	
250PPBCN	A	Total CN	P	254.1396	µg/l	7/24/2018 11:34:53	
500PPBCN	A	Total CN	P	497.789	µg/l	7/24/2018 11:34:54	
ICV1	S	Total CN	P	96.2284	µg/l	7/24/2018 16:24:37	
ICB1	S	Total CN	P	0.0716	µg/l	7/24/2018 16:24:38	
CCV1	S	Total CN	P	241.3614	µg/l	7/24/2018 16:24:39	
CCB1	S	Total CN	P	-0.0799	µg/l	7/24/2018 16:24:40	
PB111413BL	S	Total CN	P	-0.6051	µg/l	7/24/2018 16:24:41	
PB111413BS	S	Total CN	P	98.0027	µg/l	7/24/2018 16:24:42	
MIDPB111413	S	Total CN	P	246.8158	µg/l	7/24/2018 16:24:43	
J4080-30	S	Total CN	P	392.1601	µg/l	7/24/2018 16:24:44	
J4080-30	S	Total CN	P	391.8676	µg/l	7/24/2018 16:24:45	
CCV2	S	Total CN	P	246.8263	µg/l	7/24/2018 16:24:46	
CCB2	S	Total CN	P	0.0215	µg/l	7/24/2018 16:24:47	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/24/2018 16:35


Test Total CN

Accepted 7/24/2018 11:36

Factor 1107
Bias 0.004

Coeff. of det. 0.999874

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7386	0.0000	
2	5.0PPBCN	0.008	4.1380	5.0000	
3	10PPBCN	0.012	9.1804	10.0000	
4	50PPBCN	0.049	49.3210	50.0000	
5	100PPBCN	0.095	101.1707	100.0000	
6	250PPBCN	0.234	254.1396	250.0000	
7	500PPBCN	0.454	497.7890	500.0000	