



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96840

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	07/24/2018	15:20
2	CAL2	1	Water			22.0	7.00	07/24/2018	15:22
3	CAL3	1	Water			21.9	10.04	07/24/2018	15:24
4	ICV	1	Water			22.3	7.00	07/24/2018	15:26
5	CCV1	1	Water			21.7	2.00	07/24/2018	16:00
6	J4080-06	1	Solid	20.02	20	23.9	11.20	07/24/2018	16:02
7	CCV2	1	Water			21.8	12.00	07/24/2018	16:18