



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96841

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	07/24/2018	15:20
2	CAL2	1	Water			22.0	7.00	07/24/2018	15:22
3	CAL3	1	Water			21.9	10.04	07/24/2018	15:24
4	ICV	1	Water			22.3	7.00	07/24/2018	15:26
5	CCV1	1	Water			21.7	2.00	07/24/2018	16:00
6	J4080-06	1	Solid	20.02	20	23.9	11.20	07/24/2018	16:02
7	J4123-04	1	Solid	20.03	20	24.3	6.71	07/24/2018	16:04
8	J4124-04	1	Solid	20.02	20	23.9	6.00	07/24/2018	16:06
9	J4124-08	1	Solid	20.04	20	23.6	6.19	07/24/2018	16:08
10	J4135-01	1	Solid	20.02	20	24.2	8.70	07/24/2018	16:10
11	J4135-02	1	Solid	20.04	20	24.3	8.41	07/24/2018	16:12
12	J4135-03	1	Solid	20.02	20	24.1	8.56	07/24/2018	16:14
13	J4135-03DUP	1	Solid	20.03	20	24.2	8.57	07/24/2018	16:16
14	CCV2	1	Water			21.8	12.00	07/24/2018	16:18

LB 96841

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity072418

WorkList ID : 114831

Date : 7/24/2018 3:31:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Solid	J4123-04	Corrosivity	Cool 4 deg C	PSEG01	K63	RO-1-3	07/24/2018	9045C
07/24/2018	Solid	J4124-04	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-11-S	07/24/2018	9045C
07/24/2018	Solid	J4124-08	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-11-D	07/24/2018	9045C
07/24/2018	Solid	J4135-01	Corrosivity	Cool 4 deg C	PARS15	N12	CON-WC-S-1-0-25	07/24/2018	9045C
07/24/2018	Solid	J4135-02	Corrosivity	Cool 4 deg C	PARS15	N12	CON-WC-S-2-0-11	07/24/2018	9045C
07/24/2018	Solid	J4135-03	Corrosivity	Cool 4 deg C	PARS15	N12	CON-WC-S-3-0-11	07/24/2018	9045C

Date/Time 07/24/18 3:50 PM

Received by: SP

Relinquished by: CP

Date/Time 07/24/18 4:30

Received by: CP

Relinquished by: SP