



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96843

Slope : 97.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	07/24/2018	15:20
2	CAL2	1	Water			22.0	7.00	07/24/2018	15:22
3	CAL3	1	Water			21.9	10.04	07/24/2018	15:24
4	ICV	1	Water			22.3	7.00	07/24/2018	15:26
5	CCV1	1	Water			21.8	2.00	07/24/2018	15:28
6	J4128-01	1	Water			27.9	6.74	07/24/2018	15:30
7	J4128-01DUP	1	Water			28.0	6.75	07/24/2018	15:32
8	CCV2	1	Water			21.9	12.00	07/24/2018	15:34

LB96843

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH072418

WorkList ID : 114832

Date : 7/24/2018 3:31:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Water	J4128-01	pH	Cool 4 deg C	PSEG01	K63	OILY-WATER	07/24/2018	9040C

Date/Time 07/24/18 3:10 PM
Received by: SP
Relinquished by: CP

Date/Time 07/24/18 4:30
Received by: CP
Relinquished by: SP