


7470A

L396848
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37790



A= 0.0000e+000

B= 2.6409e-004

C= -2.6470e-003

Rho= 0.9998072

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-9	0.000	-9				
0.2	0.200	0.214	0.014	819	0.0 %	819				
2.5	2.500	2.550	0.050	9667	0.0 %	9667				
5.0	5.000	4.864	-0.136	18430	0.0 %	18430				
7.5	7.500	7.599	0.099	28786	0.0 %	28786				
10.0	10.000	9.977	-0.023	37790	0.0 %	37790				

LB96848 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jul 2018 09:08:41

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.0000	7470A	PPB	25 Jul 2018 09:10:51	S	Std
0.2 - 1	50.2	819	-	0.2000	7470A	PPB	25 Jul 2018 09:12:59	S	Std
2.5 - 1	52.5	9667	-	2.5000	7470A	PPB	25 Jul 2018 09:15:07	S	Std
5.0 - 1	55	18430	-	5.0000	7470A	PPB	25 Jul 2018 09:17:16	S	Std
7.5 - 1	57.5	28786	-	7.5000	7470A	PPB	25 Jul 2018 09:19:24	S	Std
10.0 - 1	510	37790	-	10.0000	7470A	PPB	25 Jul 2018 09:23:57	S	Std
ICV69 - 1	ICV69	14915	3.9362	-	7470A	PPB	25 Jul 2018 09:26:31	U	SMPL
ICB69 - 1	ICB69	-106	-0.0306	-	7470A	PPB	25 Jul 2018 09:28:38	U	SMPL
CCV97 - 1	CCV97	18516	4.8872	-	7470A	PPB	25 Jul 2018 09:30:45	U	SMPL
CCB97 - 1	CCB97	-80	-0.0238	-	7470A	PPB	25 Jul 2018 09:32:52	U	SMPL
CRA - 1	CRA	801	0.2089	-	7470A	PPB	25 Jul 2018 09:35:00	U	SMPL
HighStd - 1	HighStd	37910	10.0089	-	7470A	PPB	25 Jul 2018 09:37:06	U	SMPL
ChkStd - 1	ChkStd	27196	7.1795	-	7470A	PPB	25 Jul 2018 09:39:13	U	SMPL
PB111436BL - 1	PBW	-220	-0.0607	-	7470A	PPB	25 Jul 2018 09:41:22	U	SMPL
PB111436BS - 1	LCSW	14085	3.7170	-	7470A	PPB	25 Jul 2018 09:43:31	U	SMPL
PB111436TB - 1	PB111389TB	-153	-0.0431	-	7470A	PPB	25 Jul 2018 09:45:39	U	SMPL
J4110-03 - 1	GS-RO-338-RO-147-RO-278	-23	-0.0087	-	7470A	PPB	25 Jul 2018 09:47:47	U	SMPL
J4110-03DUP - 1	GS-RO-338-RO-147-RO-278DUP	-29	-0.0103	-	7470A	PPB	25 Jul 2018 09:49:55	U	SMPL
J4110-03MS - 1	GS-RO-338-RO-147-RO-278MS	13782	3.6370	-	7470A	PPB	25 Jul 2018 09:52:03	U	SMPL
J4110-03MSD - 1	GS-RO-338-RO-147-RO-278MSD	13551	3.5760	-	7470A	PPB	25 Jul 2018 09:54:10	U	SMPL
CCV98 - 1	CCV98	18665	4.9265	-	7470A	PPB	25 Jul 2018 09:56:18	U	SMPL
CCB98 - 1	CCB98	-53	-0.0166	-	7470A	PPB	25 Jul 2018 09:58:25	U	SMPL
J4118-01 - 1	BPCXRAY01	1906	0.5007	-	7470A	PPB	25 Jul 2018 10:00:32	U	SMPL
J4119-05 - 1	GB-SOIL-PILE	-78	-0.0232	-	7470A	PPB	25 Jul 2018 10:02:39	U	SMPL
J4110-03LX5 - 1		8	-0.0005	-	7470A	PPB	25 Jul 2018 10:04:45	U	SMPL
J4110-03A - 1		14373	3.7931	-	7470A	PPB	25 Jul 2018 10:06:52	U	SMPL
CCV99 - 1	CCV99	18249	4.8167	-	7470A	PPB	25 Jul 2018 10:08:59	U	SMPL
CCB99 - 1	CCB99	-52	-0.0164	-	7470A	PPB	25 Jul 2018 10:11:06	U	SMPL