

245.1


LB96851

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34742



A= 0.0000e+000

B= 2.8826e-004

C= -9.7514e-003

Rho= 0.9999904

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	20	0.000	20				
0.05	0.050	0.052	0.002	213	0.0 %	213				
0.2	0.200	0.209	0.009	758	0.0 %	758				
2.5	2.500	2.479	-0.021	8633	0.0 %	8633				
5.0	5.000	5.030	0.030	17482	0.0 %	17482				
7.5	7.500	7.480	-0.020	25982	0.0 %	25982				
10.0	10.000	10.005	0.005	34742	0.0 %	34742				

LB96851 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Jul 2018 11:26:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	20	-	0.0000	245.1	PPB	25 Jul 2018 11:29:48	S	Std
0.05 - 1	50.05	213	-	0.0500	245.1	PPB	25 Jul 2018 11:31:56	S	Std
0.2 - 1	50.2	758	-	0.2000	245.1	PPB	25 Jul 2018 11:34:05	S	Std
2.5 - 1	52.5	8633	-	2.5000	245.1	PPB	25 Jul 2018 11:36:13	S	Std
5.0 - 1	55	17482	-	5.0000	245.1	PPB	25 Jul 2018 11:38:21	S	Std
7.5 - 1	57.5	25982	-	7.5000	245.1	PPB	25 Jul 2018 11:40:28	S	Std
10.0 - 1	510	34742	-	10.0000	245.1	PPB	25 Jul 2018 11:45:02	S	Std
ICV70 - 1	ICV70	13733	3.9490	-245.1		PPB	25 Jul 2018 11:47:34	U	SMPL
ICB70 - 1	ICB70	-178	-0.0611	-245.1		PPB	25 Jul 2018 11:49:42	U	SMPL
CCV01 - 1	CCV01	17401	5.0063	-245.1		PPB	25 Jul 2018 11:51:50	U	SMPL
CCB01 - 1	CCB01	-193	-0.0654	-245.1		PPB	25 Jul 2018 11:53:56	U	SMPL
CRA - 1	CRA	240	0.0594	-245.1		PPB	25 Jul 2018 11:56:05	U	SMPL
HighStd - 1	HighStd	34997	10.0786	-245.1		PPB	25 Jul 2018 11:58:11	U	SMPL
ChkStd - 1	ChkStd	23828	6.8590	-245.1		PPB	25 Jul 2018 12:00:18	U	SMPL
PB111393BL - 1	PBW	-135	-0.0487	-245.1		PPB	25 Jul 2018 12:02:27	U	SMPL
PB111393BS - 1	LCSW	12917	3.7138	-245.1		PPB	25 Jul 2018 12:04:36	U	SMPL
J4107-01 - 1	071918-DBRI-E-D-B	-163	-0.0567	-245.1		PPB	25 Jul 2018 12:06:44	U	SMPL
J4107-02 - 1	071918-DBRI-E-D-M	-50	-0.0242	-245.1		PPB	25 Jul 2018 12:08:52	U	SMPL
J4107-03 - 1	071918-DBRI-E-D-S	-92	-0.0363	-245.1		PPB	25 Jul 2018 12:11:00	U	SMPL
J4107-04 - 1	071918-DBRI-E-U-B	-97	-0.0377	-245.1		PPB	25 Jul 2018 12:13:09	U	SMPL
J4107-05 - 1	071918-DBRI-E-U-M	-19	-0.0152	-245.1		PPB	25 Jul 2018 12:15:16	U	SMPL
CCV02 - 1	CCV02	17920	5.1559	-245.1		PPB	25 Jul 2018 12:17:23	U	SMPL
CCB02 - 1	CCB02	-149	-0.0527	-245.1		PPB	25 Jul 2018 12:19:30	U	SMPL
J4107-06 - 1	071918-DBRI-E-U-S	-20	-0.0155	-245.1		PPB	25 Jul 2018 12:21:38	U	SMPL
J4107-07 - 1	071918-TIPC-E-D-M	3	-0.0089	-245.1		PPB	25 Jul 2018 12:23:45	U	SMPL
J4107-08 - 1	071918-TIPC-E-U-M	-19	-0.0152	-245.1		PPB	25 Jul 2018 12:25:51	U	SMPL
J4107-08DUP - 1	071918-TIPC-E-U-MDUP	10	-0.0069	-245.1		PPB	25 Jul 2018 12:27:58	U	SMPL
J4107-08MS - 1	071918-TIPC-E-U-MMS	15922	4.5800	-245.1		PPB	25 Jul 2018 12:30:05	U	SMPL
J4107-08MSD - 1	071918-TIPC-E-U-MMSD	14744	4.2404	-245.1		PPB	25 Jul 2018 12:32:13	U	SMPL
J4133-01 - 1	072318-DBRI-F-D-B	-193	-0.0654	-245.1		PPB	25 Jul 2018 12:34:23	U	SMPL
J4133-02 - 1	072318-DBRI-F-D-M	-61	-0.0273	-245.1		PPB	25 Jul 2018 12:36:31	U	SMPL
J4133-03 - 1	072318-DBRI-F-D-S	-126	-0.0461	-245.1		PPB	25 Jul 2018 12:38:40	U	SMPL
J4133-04 - 1	072318-DBRI-F-U-B	-56	-0.0259	-245.1		PPB	25 Jul 2018 12:40:47	U	SMPL
CCV03 - 1	CCV03	17259	4.9654	-245.1		PPB	25 Jul 2018 12:42:53	U	SMPL
CCB03 - 1	CCB03	-202	-0.0680	-245.1		PPB	25 Jul 2018 12:45:00	U	SMPL
J4133-05 - 1	072318-DBRI-F-U-M	-126	-0.0461	-245.1		PPB	25 Jul 2018 12:47:08	U	SMPL
J4133-06 - 1	072318-DBRI-F-U-S	-78	-0.0322	-245.1		PPB	25 Jul 2018 12:49:15	U	SMPL
J4133-07 - 1	072318-TIPC-E-D-M	-86	-0.0345	-245.1		PPB	25 Jul 2018 12:51:22	U	SMPL
J4133-08 - 1	072318-TIPC-E-U-M	-78	-0.0322	-245.1		PPB	25 Jul 2018 12:53:28	U	SMPL
J4107-08LX5 - 1		-49	-0.0239	-245.1		PPB	25 Jul 2018 12:55:36	U	SMPL
J4107-08A - 1		16980	4.8850	-245.1		PPB	25 Jul 2018 12:57:43	U	SMPL
CCV04 - 1	CCV04	17739	5.1038	-245.1		PPB	25 Jul 2018 12:59:51	U	SMPL
CCB04 - 1	CCB04	-249	-0.0815	-245.1		PPB	25 Jul 2018 13:01:59	U	SMPL