

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB96858

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP66402
calibration std. phosphate 0.5 ppm	WP66401
calibration std. phosphate 0.3 ppm	WP66400
calibration std. 0 ppm	WP66397
phosphate ICV std.	WP66404
phosphate CCV std.	WP66403
calibration std. phosphate 0.1 ppm	WP66399
calibration std. phosphate 0.05 ppm	WP66398
Combined reagent	WP66407
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0009Slope: 0.652Regression: 0.999828

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/25/2018	11:51
2	CAL2	0.05	1	50	50	0.033	07/25/2018	11:51
3	CAL3	0.10	1	50	50	0.066	07/25/2018	11:52
4	CAL4	0.30	1	50	50	0.186	07/25/2018	11:52
5	CAL5	0.50	1	50	50	0.330	07/25/2018	11:53
6	CAL6	1.00	1	50	50	0.651	07/25/2018	11:53



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB96858

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.501	07/25/2018	11:54
2	ICB		1	50	50	0.000	0.001	07/25/2018	11:54
3	CCV1	0.50	1	50	50	0.329	0.506	07/25/2018	11:55
4	CCB1		1	50	50	0.000	0.001	07/25/2018	11:55
5	PB111468BL		1	50	50	0.000	0.001	07/25/2018	11:56
6	PB111468BS	0.50	1	50	50	0.324	0.498	07/25/2018	11:56
7	J4059-01		1	50	50	0.166	0.256	07/25/2018	11:57
8	J4059-01DUP		1	50	50	0.172	0.265	07/25/2018	11:57
9	J4059-01MS	0.50	1	50	50	0.500	0.768	07/25/2018	11:58
10	J4059-01MSD	0.50	1	50	50	0.501	0.770	07/25/2018	11:59
11	J4080-21		1	50	50	0.801	1.230	07/25/2018	11:59
12	J4134-01		1	50	50	0.058	0.090	07/25/2018	12:00
13	J4134-02		1	50	50	0.053	0.083	07/25/2018	12:01
14	J4146-01		1	50	50	0.398	0.612	07/25/2018	12:01
15	CCV2	0.50	1	50	50	0.327	0.503	07/25/2018	12:02
16	CCB2		1	50	50	0.000	0.001	07/25/2018	12:02
17	J4080-21		5	50	50	0.285	0.438	07/25/2018	12:10
18	CCV3	0.50	1	50	50	0.328	0.504	07/25/2018	12:11
19	CCB3		1	50	50	0.000	0.001	07/25/2018	12:11