

L996860


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

4752E



A= 0.0000e+000

B= 2.0998e-004

C= -9.0199e-003

Rho= 0.9997888

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	47	0.000	47				
0.2	0.200	0.190	-0.010	947	0.0 %	947				
2.5	2.500	2.573	0.073	12296	0.0 %	12296				
5.0	5.000	4.866	-0.134	23215	0.0 %	23215				
7.5	7.500	7.601	0.101	36241	0.0 %	36241				
10.0	10.000	9.970	-0.030	47526	0.0 %	47526				

LB96860 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Jul 2018 15:07:23

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471B	PPB		25 Jul 2018 15:09:38	S	Std
0.2 - 1	50.2	947	-	0.20007471B	PPB		25 Jul 2018 15:11:55	S	Std
2.5 - 1	52.5	12296	-	2.50007471B	PPB		25 Jul 2018 15:14:13	S	Std
5.0 - 1	55	23215	-	5.00007471B	PPB		25 Jul 2018 15:16:30	S	Std
7.5 - 1	57.5	36241	-	7.50007471B	PPB		25 Jul 2018 15:18:47	S	Std
10.0 - 1	59	47526	-	10.00007471B	PPB		25 Jul 2018 15:24:13	S	Std
ICV71 - 1	ICV71	18179	3.8081	-7471B	PPB		25 Jul 2018 15:26:59	U	SMPL
ICB71 - 1	ICB71	-127	-0.0357	-7471B	PPB		25 Jul 2018 15:29:15	U	SMPL
CCV05 - 1	CCV05	23408	4.9061	-7471B	PPB		25 Jul 2018 15:31:30	U	SMPL
CCB05 - 1	CCB05	-155	-0.0416	-7471B	PPB		25 Jul 2018 15:33:46	U	SMPL
CRA - 1	CRA	907	0.1814	-7471B	PPB		25 Jul 2018 15:36:02	U	SMPL
HighStd - 1	HighStd	48048	10.0799	-7471B	PPB		25 Jul 2018 15:38:18	U	SMPL
ChkStd - 1	ChkStd	34003	7.1308	-7471B	PPB		25 Jul 2018 15:46:10	U	SMPL
PB111439BL - 1	PBS	-98	-0.0296	-7471B	PPB		25 Jul 2018 15:48:27	U	SMPL
PB111439BS - 1	LCSS	15703	3.2882	-7471B	PPB		25 Jul 2018 15:50:45	U	SMPL
J3821-07 - 1	HR-06-072418-A	16948	3.5496	-7471B	PPB		25 Jul 2018 15:53:01	U	SMPL
J3821-09 - 1	HR-06-072418-B	31792	6.6665	-7471B	PPB		25 Jul 2018 15:55:19	U	SMPL
J3821-11 - 1	HR-06-072418-C	22948	4.8095	-7471B	PPB		25 Jul 2018 15:57:37	U	SMPL
J4123-01 - 1	RO-1-3	1324	0.2690	-7471B	PPB		25 Jul 2018 15:59:56	U	SMPL
J4124-03 - 1	TP-11-S	7992	1.6691	-7471B	PPB		25 Jul 2018 16:02:13	U	SMPL
CCV06 - 1	CCV06	23221	4.8668	-7471B	PPB		25 Jul 2018 16:04:29	U	SMPL
CCB06 - 1	CCB06	-166	-0.0439	-7471B	PPB		25 Jul 2018 16:06:46	U	SMPL
J4124-07 - 1	TP-11-D	879	0.1755	-7471B	PPB		25 Jul 2018 16:09:01	U	SMPL
J4124-07DUP - 1	TP-11-DDUP	840	0.1674	-7471B	PPB		25 Jul 2018 16:14:00	U	SMPL
J4124-07MS - 1	TP-11-DMS	18164	3.8050	-7471B	PPB		25 Jul 2018 16:16:16	U	SMPL
J4124-07MSD - 1	TP-11-DMSD	18487	3.8728	-7471B	PPB		25 Jul 2018 16:18:33	U	SMPL
J4126-01 - 1	2018-NYSEG-CLARK-AREA-13	11974	2.5052	-7471B	PPB		25 Jul 2018 16:20:49	U	SMPL
J4126-03 - 1	2018-NYSEG-CLARK-AREA-12	13535	2.8330	-7471B	PPB		25 Jul 2018 16:23:08	U	SMPL
J4126-05 - 1	2018-NYSEG-CLARK-AREA-11	89839	18.8549	-7471B	PPB		25 Jul 2018 16:25:25	U	SMPL
J4126-07 - 1	2018-NYSEG-CLARK-AREA-10	54177	11.3668	-7471B	PPB		25 Jul 2018 16:27:42	U	SMPL
J4126-09 - 1	2018-NYSEG-CLARK-AREA-9	9977	2.0859	-7471B	PPB		25 Jul 2018 16:30:03	U	SMPL
J4126-11 - 1	2018-NYSEG-CLARK-AREA-8	7268	1.5171	-7471B	PPB		25 Jul 2018 16:32:22	U	SMPL
CCV07 - 1	CCV07	23510	4.9275	-7471B	PPB		25 Jul 2018 16:34:38	U	SMPL
CCB07 - 1	CCB07	-119	-0.0340	-7471B	PPB		25 Jul 2018 16:36:54	U	SMPL
J4136-02 - 1	CON-SB-01-02-03-0-25	3108	0.6436	-7471B	PPB		25 Jul 2018 16:41:28	U	SMPL
J4144-01 - 1	MCMASER-SEDIMENT-7-23-18	8456	1.7665	-7471B	PPB		25 Jul 2018 16:43:44	U	SMPL
PB111440BL - 1	PBS	-37	-0.0168	-7471B	PPB		25 Jul 2018 16:46:01	U	SMPL
PB111440BS - 1	LCSS	16323	3.4184	-7471B	PPB		25 Jul 2018 16:48:17	U	SMPL
J4140-01 - 1	EPE-F-8(75-77)	1163	0.2352	-7471B	PPB		25 Jul 2018 16:50:34	U	SMPL
J4140-02 - 1	EPE-F-8(85-87)	746	0.1476	-7471B	PPB		25 Jul 2018 16:52:51	U	SMPL
J4140-03 - 1	EPE-F-8(90-92)	695	0.1369	-7471B	PPB		25 Jul 2018 16:55:08	U	SMPL
J4140-04 - 1	EPE-F-8(95-97)	560	0.1086	-7471B	PPB		25 Jul 2018 16:57:26	U	SMPL
J4140-05 - 1	EPE-F-8(105-107)	207	0.0344	-7471B	PPB		25 Jul 2018 16:59:42	U	SMPL
CCV08 - 1	CCV08	23544	4.9346	-7471B	PPB		25 Jul 2018 17:01:59	U	SMPL
CCB08 - 1	CCB08	-93	-0.0285	-7471B	PPB		25 Jul 2018 17:04:15	U	SMPL
J4140-06 - 1	EPE-F-8(115-117)	392	0.0733	-7471B	PPB		25 Jul 2018 17:06:31	U	SMPL
J4140-07 - 1	EPE-F-8(125-127)	1068	0.2152	-7471B	PPB		25 Jul 2018 17:08:48	U	SMPL
J4140-08 - 1	EPE-F-8(130-132)	878	0.1753	-7471B	PPB		25 Jul 2018 17:11:04	U	SMPL
J4140-09 - 1	EPE-F-8(135-137)	1082	0.2182	-7471B	PPB		25 Jul 2018 17:13:21	U	SMPL
J4140-10 - 1	EPE-F-8(140-142)	721	0.1424	-7471B	PPB		25 Jul 2018 17:15:37	U	SMPL
J4140-11 - 1	EPE-F-8(145-147)	566	0.1098	-7471B	PPB		25 Jul 2018 17:17:54	U	SMPL
J4141-01 - 1	FK-BPM-03-001-ABCD	1187	0.2402	-7471B	PPB		25 Jul 2018 17:20:11	U	SMPL
J4143-01 - 1	SB-1(3-5)	61111	12.8228	-7471B	PPB		25 Jul 2018 17:22:29	U	SMPL
J4143-01DUP - 1	SB-1(3-5)DUP	59044	12.3887	-7471B	PPB		25 Jul 2018 17:24:47	U	SMPL
J4143-02 - 1	J4143-01MS	73499	15.4239	-7471B	PPB		25 Jul 2018 17:27:05	U	SMPL
CCV09 - 1	CCV09	22868	4.7927	-7471B	PPB		25 Jul 2018 17:29:24	U	SMPL
CCB09 - 1	CCB09	-182	-0.0472	-7471B	PPB		25 Jul 2018 17:31:43	U	SMPL
J4143-03 - 1	J4143-01MSD	73668	15.4594	-7471B	PPB		25 Jul 2018 17:34:00	U	SMPL
J4143-04 - 1	SB-1(11-13)	169293	35.5383	-7471B	PPB		25 Jul 2018 17:36:16	U	SMPL
J4143-05 - 1	GW-01(3-5)	21841	4.5770	-7471B	PPB		25 Jul 2018 17:38:35	U	SMPL
J4143-06 - 1	GW-01(10-12)	66716	13.9997	-7471B	PPB		25 Jul 2018 17:41:00	U	SMPL
J4143-07 - 1	GW-02(3-5)	31860	6.6808	-7471B	PPB		25 Jul 2018 17:43:17	U	SMPL
J4143-08 - 1	GW-02(8-10)	1256	0.2547	-7471B	PPB		25 Jul 2018 17:45:37	U	SMPL
J4143-09 - 1	SB-3(3-5)	16667	3.4906	-7471B	PPB		25 Jul 2018 17:47:56	U	SMPL
J4143-10 - 1	SB-3(9-11)	24754	5.1887	-7471B	PPB		25 Jul 2018 17:50:13	U	SMPL
CCV10 - 1	CCV10	23576	4.9414	-7471B	PPB		25 Jul 2018 17:52:31	U	SMPL
CCB10 - 1	CCB10	-124	-0.0351	-7471B	PPB		25 Jul 2018 17:54:47	U	SMPL
J4124-07LX5 - 1		124	0.0170	-7471B	PPB		25 Jul 2018 17:57:05	U	SMPL
J4124-07A - 1		17913	3.7523	-7471B	PPB		25 Jul 2018 17:59:21	U	SMPL
J4143-01LX5 - 1		11740	2.4561	-7471B	PPB		25 Jul 2018 18:01:38	U	SMPL
J4143-01A - 1		76742	16.1049	-7471B	PPB		25 Jul 2018 18:03:54	U	SMPL
J4126-05DLX5 - 1		17643	3.6956	-7471B	PPB		25 Jul 2018 18:06:11	U	SMPL
J4126-07DLX2 - 1		26779	5.6139	-7471B	PPB		25 Jul 2018 18:08:30	U	SMPL
J4143-01DLX2 - 1		30802	6.4586	-7471B	PPB		25 Jul 2018 18:10:47	U	SMPL
J4143-01DUPDLX2 - 1		30091	6.3093	-7471B	PPB		25 Jul 2018 18:13:05	U	SMPL
J4143-02DLX2 - 1		37399	7.8438	-7471B	PPB		25 Jul 2018 18:15:23	U	SMPL
J4143-03DLX2 - 1		37468	7.8583	-7471B	PPB		25 Jul 2018 18:17:41	U	SMPL
CCV11 - 1	CCV11	23184	4.8590	-7471B	PPB		25 Jul 2018 18:20:00	U	SMPL
CCB11 - 1	CCB11	-192	-0.0493	-7471B	PPB		25 Jul 2018 18:22:18	U	SMPL
J4143-01LDLX10 - 1		6345	1.3233	-7471B	PPB		25 Jul 2018 18:24:36	U	SMPL
J4143-01ADLX2 - 1		40807	8.5594	-7471B	PPB		25 Jul 2018 18:26:53	U	SMPL
J4143-04DLX10 - 1		18284	3.8302	-7471B	PPB		25 Jul 2018 18:29:10	U	SMPL
J4143-05DLX2 - 1		33602	7.0466	-7471B	PPB		25 Jul 2018 18:31:28	U	SMPL
CCV12 - 1	CCV12	22770	4.7721	-7471B	PPB		25 Jul 2018 18:33:46	U	SMPL
CCB12 - 1	CCB12	-171	-0.0449	-7471B	PPB		25 Jul 2018 18:36:04	U	SMPL