



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96861

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/25/2018	12:40
2	CAL2	1	Water			21.8	7.00	07/25/2018	12:42
3	CAL3	1	Water			21.6	10.03	07/25/2018	12:44
4	ICV	1	Water			21.8	7.00	07/25/2018	12:46
5	CCV1	1	Water			21.7	2.00	07/25/2018	12:48
6	J4146-01	1	Water			18.7	7.72	07/25/2018	12:50
7	J4146-01DUP	1	Water			18.8	7.73	07/25/2018	12:52
8	CCV2	1	Water			21.8	12.00	07/25/2018	12:54

WORKLIST(Hardcopy Internal Chain)

LB96861

WorkList Name : pH072518

WorkList ID : 114876

Date : 7/25/2018 12:28:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Water	J4146-01 C	pH	Cool 4 deg C	PARS02	N33	SES-CS-PH4-COA-01	07/24/2018	SM 4500-PH B

Date/Time 07/25/18 12:30
 Received by: SP
 Relinquished by: CP

Date/Time 07/25/18 1:00
 Received by: CP
 Relinquished by: SP