



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96868

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	07/25/2018	16:45
2	CAL2	1	Water			22.3	7.01	07/25/2018	16:47
3	CAL3	1	Water			22.2	10.04	07/25/2018	16:49
4	ICV	1	Water			22.3	7.00	07/25/2018	16:51
5	CCV1	1	Water			22.1	2.00	07/25/2018	16:53
6	J4158-01	1	Water			25.4	6.51	07/25/2018	16:55
7	J4158-01DUP	1	Water			25.5	6.52	07/25/2018	16:57
8	J4167-01	1	Water			26.6	7.08	07/25/2018	16:59
9	CCV2	1	Water			22.1	12.00	07/25/2018	17:01

WORKLIST(Hardcopy Internal Chain)

LB 96868

WorkList Name : pH072518-1

WorkList ID : 114883

Date : 7/25/2018 4:42:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/25/2018	Water	J4158-01	pH	Cool 4 deg C	PSEG01	K63	OILY-WATER-TOTES	07/25/2018	9040C
07/25/2018	Water	J4167-01	pH	Cool 4 deg C	PSEG01	K62	GS-OILY-WATER	07/25/2018	9040C

Date/Time 07/25/18 4:30
 Received by: SP
 Relinquished by: CP

Date/Time 07/25/18 5:45
 Received by: CP
 Relinquished by: SP