

LB96870

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/25/2018 16:47

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.030	0.0	0.111	
ICB1	0.003	0.0	0.004	
CCV1	0.975	0.0	0.105	
CCB1	0.001	0.0	0.004	
PB111421BL	-0.002	0.0	0.003	
PB111421BS	1.026	0.0	0.110	
J4080-17	0.484	0.0	0.054	
CCV2	0.969	0.0	0.105	
CCB2	0.000	0.0	0.003	

N	9
Mean	0.498
SD	0.5002
CV%	100.35

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/25/2018 16:21


Test Phenolics-

Accepted 7/25/2018 16:21

Factor 9.578
Bias 0.003

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0513	0.0500	
2	Phenol-2	0.014	0.0989	0.1000	
3	Phenol-2	0.030	0.2508	0.2500	
4	Phenol-2	0.056	0.5005	0.5000	
5	Phenol-2	0.108	0.9974	1.0000	
6	Phenol-2	0.212	2.0011	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 25 16:21:19 2018

Wed Jul 25 16:47:15 2018

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0513	mg/l	7/25/2018 16:21:19	
0.1PPM	A	Phenolics-	P	0.0989	mg/l	7/25/2018 16:21:20	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	7/25/2018 16:21:21	
0.50PPM	A	Phenolics-	P	0.5005	mg/l	7/25/2018 16:21:22	
1.0PPM	A	Phenolics-	P	0.9974	mg/l	7/25/2018 16:21:23	
2.0PPM	A	Phenolics-	P	2.0011	mg/l	7/25/2018 16:21:24	
ICV1	S	Phenolics-	P	1.03	mg/l	7/25/2018 16:43:16	
ICB1	S	Phenolics-	P	0.003	mg/l	7/25/2018 16:43:17	
CCV1	S	Phenolics-	P	0.9752	mg/l	7/25/2018 16:43:18	
CCB1	S	Phenolics-	P	0.0009	mg/l	7/25/2018 16:43:19	
PB111421BL	S	Phenolics-	P	-0.0023	mg/l	7/25/2018 16:43:20	
PB111421BS	S	Phenolics-	P	1.0257	mg/l	7/25/2018 16:43:21	
J4080-17	S	Phenolics-	P	0.4843	mg/l	7/25/2018 16:43:24	
CCV2	S	Phenolics-	P	0.9691	mg/l	7/25/2018 16:43:26	
CCB2	S	Phenolics-	P	0.0003	mg/l	7/25/2018 16:43:27	