

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/26/2018

TEMP1 IN: 103 07/26/2018 07:10 TEMP OUT: 104 07/26/2018 08:15

TEMP2 IN: 104 07/26/2018 09:00 TEMP OUT: 104 07/26/2018 10:10

TEMP3 IN: 105 07/26/2018 11:30 TEMP OUT: 105 07/26/2018 13:10

TEMP4 IN: 105 07/26/2018 14:00 TEMP OUT: 105 07/26/2018 15:35

Run Number: LB96879

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96879BL	LB96879BL	1.3945	1.3945	100	1.3945	1.3945	1.3945	0.0000	0
2	LB96879BS	LB96879BS	1.3993	1.3993	100	1.4492	1.4492	1.4492	0.0499	499
3	J4146-01	SES-CS-PH4-COA-01	1.4085	1.4085	800	1.4098	1.4098	1.4098	0.0013	1.6
4	J4150-01	UDS-MF-02A-B	1.3990	1.3990	400	1.4021	1.4021	1.4021	0.0031	7.8
5	J4151-01	UDS-CH-4-01	1.3944	1.3944	100	1.4085	1.4085	1.4085	0.0141	141
6	J4153-03	OUTFALL001-180724	1.4004	1.4004	900	1.4007	1.4007	1.4007	0.0003	0.3
7	J4155-01	UDS-ER-22-01	1.3974	1.3974	400	1.4009	1.4009	1.4009	0.0035	8.8
8	J4157-01	DRAIN-WATER-TANK-1	1.3979	1.3979	10	1.4210	1.4210	1.4210	0.0231	2310
9	J4157-01DUP	DRAIN-WATER-TANK-1DUP	1.4095	1.4095	10	1.4322	1.4322	1.4322	0.0227	2270
10	J4158-01	OILY-WATER-TOTES	1.3986	1.3986	5	1.5128	1.5128	1.5128	0.1142	22840
11	J4160-01	072418-TIPC-F-D-B	1.3971	1.3971	400	1.4124	1.4124	1.4124	0.0153	38.3
12	J4160-02	072418-TIPC-F-D-S	1.3940	1.3940	400	1.4079	1.4079	1.4079	0.0139	34.8
13	J4160-03	072418-TIPC-F-U-B	1.3933	1.3933	400	1.4082	1.4082	1.4082	0.0149	37.3
14	J4160-04	072418-TIPC-F-U-S	1.3963	1.3963	400	1.4095	1.4095	1.4095	0.0132	33
15	J4166-01	COMP	1.3950	1.3950	50	1.4069	1.4069	1.4069	0.0119	238



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 07/26/2018TEMP1 IN: 103 07/26/2018 07:10 TEMP OUT: 104 07/26/2018 08:15TEMP2 IN: 104 07/26/2018 09:00 TEMP OUT: 104 07/26/2018 10:10TEMP3 IN: 105 07/26/2018 11:30 TEMP OUT: 105 07/26/2018 13:10TEMP4 IN: 105 07/26/2018 14:00 TEMP OUT: 105 07/26/2018 15:35Run Number: LB96879BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4166-01DUP	COMPDUP	1.3931	1.3931	50	1.4053	1.4053	1.4053	0.0122	244
17	J4167-01	GS-OILY-WATER	1.3998	1.3998	300	1.4132	1.4132	1.4132	0.0134	44.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS072618

WorkList ID : 114894

Date : 7/26/2018 7:51:05 AM

196879

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2018	Water	J4146-01	TSS	Cool 4 deg C	PARS02	N33	SES-CS-PH4-COA-01	07/24/2018	SM2540 D
07/27/2018	Water	J4150-01	TSS	Cool 4 deg C	LIR001	M31	UDS-MF-02A-B	07/24/2018	SM2540 D
07/27/2018	Water	J4151-01	TSS	Cool 4 deg C	LIR001	M31	UDS-CH-4-01	07/24/2018	SM2540 D
07/31/2018	Water	J4153-03	TSS	Cool 4 deg C	SHAW22	M33	OUTFALL001-180724	07/24/2018	SM2540 D
07/27/2018	Water	J4155-01	TSS	Cool 4 deg C	LIR001	M41	UDS-ER-22-01	07/24/2018	SM2540 D
08/01/2018	Water	-J4157-01	TSS	Cool 4 deg C	MAJO01	M42	DRAIN-WATER-TANK-1	07/25/2018	SM2540 D
07/30/2018	Water	J4158-01	TSS	Cool 4 deg C	PSEG01	K63	OILY-WATER-TOTES	07/25/2018	SM2540 D
07/27/2018	Water	J4160-01	TSS	Cool 4 deg C	HDR103	M42	072418-TIPC-F-D-B	07/24/2018	SM2540 D
07/27/2018	Water	J4160-02	TSS	Cool 4 deg C	HDR103	M42	072418-TIPC-F-D-S	07/24/2018	SM2540 D
07/27/2018	Water	J4160-03	TSS	Cool 4 deg C	HDR103	M42	072418-TIPC-F-U-B	07/24/2018	SM2540 D
07/27/2018	Water	J4160-04	TSS	Cool 4 deg C	HDR103	M42	072418-TIPC-F-U-S	07/24/2018	SM2540 D
08/01/2018	Water	-J4166-01	TSS	Cool 4 deg C	ARAM01	M52	COMP	07/25/2018	SM2540 D
07/30/2018	Water	J4167-01	TSS	Cool 4 deg C	PSEG01	K62	GS-OILY-WATER	07/25/2018	SM2540 D

FP-ID-9825888

Date/Time 07/26/18, 08:00
 Received by: AS
 Relinquished by: AS

Date/Time 07/26/18, 11:45
 Received by: AS
 Relinquished by: AS