



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96880

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP66443
Turbidity std, 40NTU	WP66437
turbidity std 5 NTU	WP66440
turbidity std 20 NTU	WP66438
Turbidity std 1 NTU	WP66441
turbidity std 0 NTU	WP66442
turbidity std 10 NTU	WP66439
10 NTU Standard 500 ml	W2242
Turbidity Std, 200NTU	WP66436

Intercept: -0.1586

Slope: 0.8992

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.201	07/26/2018	07:15
2	CAL2	1	1	1.030	07/26/2018	07:19
3	CAL3	5	1	4.450	07/26/2018	07:22
4	CAL4	10	1	8.680	07/26/2018	07:26
5	CAL5	20	1	16.900	07/26/2018	07:30
6	CAL6	40	1	33.900	07/26/2018	07:34
7	CAL7	200	1	179.000	07/26/2018	07:36
8	CAL8	80	1	74.700	07/26/2018	07:38



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96880

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.320	10.541	07/26/2018	07:40
2	ICB		1	0.210	0.410	07/26/2018	07:43
3	CCV1	10	1	9.450	10.686	07/26/2018	07:47
4	CCB1		1	0.201	0.400	07/26/2018	07:50
5	LB96880BL		1	0.213	0.413	07/26/2018	07:53
6	J4045-02		1	3.030	3.546	07/26/2018	07:56
7	J4080-10		1	13.400	15.079	07/26/2018	07:59
8	J4150-01		1	8.350	9.462	07/26/2018	08:02
9	J4150-01DUP		1	8.360	9.474	07/26/2018	08:05
10	J4151-01		1	162.000	180.337	07/26/2018	08:07
11	J4155-01		1	21.300	23.864	07/26/2018	08:10
12	CCV2	10	1	9.160	10.363	07/26/2018	08:13
13	CCB2		1	0.206	0.405	07/26/2018	08:16