



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96881

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP66443
Turbidity std, 40NTU	WP66437
turbidity std 5 NTU	WP66440
turbidity std 20 NTU	WP66438
Turbidity std 1 NTU	WP66441
turbidity std 0 NTU	WP66442
turbidity std 10 NTU	WP66439
10 NTU Standard 500 ml	W2242

Intercept: 0.2096

Slope: 0.8411

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.201	07/26/2018	08:40
2	CAL2	1	1	1.050	07/26/2018	08:42
3	CAL3	5	1	4.450	07/26/2018	08:45
4	CAL4	10	1	8.680	07/26/2018	08:47
5	CAL5	20	1	16.900	07/26/2018	08:50
6	CAL6	40	1	33.900	07/26/2018	08:53



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96881

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.300	10.808	07/26/2018	08:55
2	ICB		1	0.199	-0.013	07/26/2018	08:57
3	CCV1	10	1	9.400	10.927	07/26/2018	08:59
4	CCB1		1	0.216	0.008	07/26/2018	09:02
5	LB96881BL		1	0.211	0.002	07/26/2018	09:06
6	J4045-01		1	3.040	3.365	07/26/2018	09:08
7	J4045-01DUP		1	3.050	3.377	07/26/2018	09:10
8	J4080-09		1	13.500	15.801	07/26/2018	09:13
9	CCV2	10	1	9.260	10.760	07/26/2018	09:16
10	CCB2		1	0.211	0.002	07/26/2018	09:18

WORKLIST(Hardcopy Internal Chain)

1396881

WorkList Name : turbidity072518-2

WorkList ID : 114870

Date : 7/25/2018 12:06:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2018	Water	J4045-01	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS	07/17/2018	180.1

Date/Time 7/26/18 8:30
 Received by: AK
 Relinquished by: AK

Date/Time 7/26/18 10:00
 Received by: W
 Relinquished by: AK