


7470A

LB96887
INSTRUMENT ID: CVD

Linear

 μ Abs.:

39240



A= 0.0000e+000

B= 2.5847e-004

C= -1.1549e-001

Rho= 0.9998710

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.051	-0.051	250	0.000	250				
0.2	0.200	0.202	0.002	1227	0.0 %	1227				
2.5	2.500	2.616	0.116	10567	0.0 %	10567				
5.0	5.000	4.952	-0.048	19604	0.0 %	19604				
7.5	7.500	7.455	-0.045	29290	0.0 %	29290				
10.0	10.000	10.027	0.027	39240	0.0 %	39240				

LB96887 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jul 2018 10:31:11

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	250	-	0.0000	7470A	PPB	26 Jul 2018 10:35:06	S	Std
0.2 - 1	50.2	1227	-	0.2000	7470A	PPB	26 Jul 2018 10:37:15	S	Std
2.5 - 1	52.5	10567	-	2.5000	7470A	PPB	26 Jul 2018 10:39:23	S	Std
5.0 - 1	55	19604	-	5.0000	7470A	PPB	26 Jul 2018 10:41:31	S	Std
7.5 - 1	57.5	29290	-	7.5000	7470A	PPB	26 Jul 2018 10:43:39	S	Std
10.0 - 1	510	39240	-	10.0000	7470A	PPB	26 Jul 2018 10:45:48	S	Std
ICV72 - 1	ICV72	15519	3.8957	-	7470A	PPB	26 Jul 2018 10:48:30	U	SMPL
ICB72 - 1	ICB72	209	-0.0615	-	7470A	PPB	26 Jul 2018 10:53:00	U	SMPL
CCV13 - 1	CCV13	19189	4.8443	-	7470A	PPB	26 Jul 2018 10:55:06	U	SMPL
CCB13 - 1	CCB13	278	-0.0436	-	7470A	PPB	26 Jul 2018 10:59:35	U	SMPL
CRA - 1	CRA	1246	0.2066	-	7470A	PPB	26 Jul 2018 11:01:41	U	SMPL
HighStd - 1	HighStd	39238	10.0263	-	7470A	PPB	26 Jul 2018 11:03:48	U	SMPL
ChkStd - 1	ChkStd	26426	6.7148	-	7470A	PPB	26 Jul 2018 11:05:55	U	SMPL
PB111471BL - 1	PBW	146	-0.0778	-	7470A	PPB	26 Jul 2018 11:08:03	U	SMPL
PB111471BS - 1	LCSW	14489	3.6295	-	7470A	PPB	26 Jul 2018 11:10:12	U	SMPL
PB111471TB - 1	PB111423TB	-76	-0.1351	-	7470A	PPB	26 Jul 2018 11:12:19	U	SMPL
J3821-08 - 1	HR-06-072418-A	12	-0.1124	-	7470A	PPB	26 Jul 2018 11:14:27	U	SMPL
J3821-08DUP - 1	HR-06-072418-ADUP	17	-0.1111	-	7470A	PPB	26 Jul 2018 11:16:35	U	SMPL
J3821-08MS - 1	HR-06-072418-AMS	15400	3.8649	-	7470A	PPB	26 Jul 2018 11:18:44	U	SMPL
J3821-08MSD - 1	HR-06-072418-AMSD	14637	3.6677	-	7470A	PPB	26 Jul 2018 11:20:51	U	SMPL
CCV14 - 1	CCV14	19537	4.9342	-	7470A	PPB	26 Jul 2018 11:22:58	U	SMPL
CCB14 - 1	CCB14	132	-0.0814	-	7470A	PPB	26 Jul 2018 11:25:05	U	SMPL
J3821-10 - 1	HR-06-072418-B	15	-0.1116	-	7470A	PPB	26 Jul 2018 11:27:13	U	SMPL
J3821-12 - 1	HR-06-072418-C	54	-0.1015	-	7470A	PPB	26 Jul 2018 11:29:20	U	SMPL
J4123-04 - 1	RO-1-3	26	-0.1088	-	7470A	PPB	26 Jul 2018 11:31:26	U	SMPL
J4124-04 - 1	TP-11-S	30	-0.1077	-	7470A	PPB	26 Jul 2018 11:33:34	U	SMPL
J4124-08 - 1	TP-11-D	26	-0.1088	-	7470A	PPB	26 Jul 2018 11:35:41	U	SMPL
J4126-02 - 1	2018-NYSEG-CLARK-AREA-13	34	-0.1067	-	7470A	PPB	26 Jul 2018 11:37:48	U	SMPL
J4126-04 - 1	2018-NYSEG-CLARK-AREA-12	15	-0.1116	-	7470A	PPB	26 Jul 2018 11:39:57	U	SMPL
J4126-06 - 1	2018-NYSEG-CLARK-AREA-11	37	-0.1059	-	7470A	PPB	26 Jul 2018 11:42:05	U	SMPL
J4126-08 - 1	2018-NYSEG-CLARK-AREA-10	20	-0.1103	-	7470A	PPB	26 Jul 2018 11:44:13	U	SMPL
J4126-10 - 1	2018-NYSEG-CLARK-AREA-9	6	-0.1139	-	7470A	PPB	26 Jul 2018 11:46:20	U	SMPL
CCV15 - 1	CCV15	19287	4.8696	-	7470A	PPB	26 Jul 2018 11:48:27	U	SMPL
CCB15 - 1	CCB15	239	-0.0537	-	7470A	PPB	26 Jul 2018 11:50:33	U	SMPL
J4126-12 - 1	2018-NYSEG-CLARK-AREA-8	12	-0.1124	-	7470A	PPB	26 Jul 2018 11:52:41	U	SMPL
J4135-01 - 1	CON-WC-S-1-0-25	43	-0.1044	-	7470A	PPB	26 Jul 2018 11:54:48	U	SMPL
J4135-02 - 1	CON-WC-S-2-0-11	51	-0.1023	-	7470A	PPB	26 Jul 2018 11:56:55	U	SMPL
J4135-03 - 1	CON-WC-S-3-0-11	57	-0.1008	-	7470A	PPB	26 Jul 2018 11:59:02	U	SMPL
J4144-02 - 1	MCMaster-SEDIMENT-7-23-18	44	-0.1041	-	7470A	PPB	26 Jul 2018 12:01:10	U	SMPL
J3821-08LX5 - 1		117	-0.0852	-	7470A	PPB	26 Jul 2018 12:03:17	U	SMPL
J3821-08A - 1		14918	3.7404	-	7470A	PPB	26 Jul 2018 12:05:26	U	SMPL
CCV16 - 1	CCV16	19392	4.8967	-	7470A	PPB	26 Jul 2018 12:07:34	U	SMPL
CCB16 - 1	CCB16	145	-0.0780	-	7470A	PPB	26 Jul 2018 12:09:42	U	SMPL