



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96891

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	5	27.0	80.6	80.9	07/26/2018	10:05
2	J4158-02		1	7	>100.0	>212.0	>212.0	07/26/2018	10:40
3	J4158-02DUP		1	7	>100.0	>212.0	>212.0	07/26/2018	11:15
4	J4167-02		1	6	>100.0	>212.0	>212.0	07/26/2018	11:46

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP72618

WorkList ID : 114919

Date : 7/26/2018 9:31:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2018	Water	J4158-02	Flash Point	Cool 4 deg C	PSEG01	K63	OILY-WATER-TOTES	07/25/2018	1010A
07/30/2018	Water	J4167-02	Flash Point	Cool 4 deg C	PSEG01	K62	GS-OILY-WATER	07/25/2018	1010A

LB 96891

Date/Time 07/26/18 9:40 AM
 Received by: KS
 Relinquished by: CP

Date/Time 07/26/18 11:50 AM
 Received by: CP
 Relinquished by: KS