



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96897

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP66477
COD calibration std. 100 ppm	WP66474
COD calibration std. 10 ppm	WP66473
COD calibration std. 150 ppm	WP66476
COD calibration std. 50 ppm	WP66475
COD calibration std. 0 ppm	WP66472
COD CCV std, 50ppm	WP66478
COD Digestion Vials Low Level 0-150Mg/L	W2375

Temp In (C): 150	Date In: 07/26/2018	Time In: 09:40
Temp Out (C): 150	Date Out: 07/26/2018	Time Out: 11:40

Intercept: 0.1033

Slope: 1.0048

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/26/2018	12:42
2	CAL2	10	1	Water	10.000	07/26/2018	12:42
3	CAL3	50	1	Water	51.000	07/26/2018	12:42
4	CAL4	100	1	Water	100.000	07/26/2018	12:43
5	CAL5	150	1	Water	151.000	07/26/2018	12:44



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96897

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	50.654	07/26/2018	12:44
2	ICB				1	Water	0.000	-0.103	07/26/2018	12:45
3	CCV1	50			1	Water	51.000	50.654	07/26/2018	12:45
4	CCB1				1	Water	0.000	-0.103	07/26/2018	12:46
5	LB96897BL				1	Water	0.000	-0.103	07/26/2018	12:46
6	LB96897BS	50			1	Water	50.000	49.658	07/26/2018	12:47
7	J4080-18				1	Water	92.000	91.458	07/26/2018	12:47
8	J4134-01				1	Water	22.000	21.792	07/26/2018	12:48
9	J4134-02				1	Water	4.000	3.878	07/26/2018	12:48
10	J4146-01				1	Water	1.000	0.892	07/26/2018	12:49
11	J4153-02				1	Water	2.000	1.888	07/26/2018	12:49
12	J4153-03				1	Water	0.000	-0.103	07/26/2018	12:50
13	J4153-03DUP				1	Water	0.000	-0.103	07/26/2018	12:50
14	J4153-03MS	50			1	Water	51.000	50.654	07/26/2018	12:51
15	CCV2	50			1	Water	51.000	50.654	07/26/2018	12:51
16	CCB2				1	Water	0.000	-0.103	07/26/2018	12:51
17	J4153-03MSD	50			1	Water	52.000	51.649	07/26/2018	12:52
18	J4158-01				20	Water	85.000	84.491	07/26/2018	12:52
19	J4167-01				1	Water	78.000	77.525	07/26/2018	12:53
20	CCV3	50			1	Water	52.000	51.649	07/26/2018	12:53
21	CCB3				1	Water	0.000	-0.103	07/26/2018	12:54

LB96897

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD072518

WorkList ID : 114862

Date : 7/25/2018 9:54:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2018	Water	J4134-01	COD	Conc H2SO4 tc	KWAR01	M63	STORM-DRAIN-1	07/24/2018	SM5220 D
08/06/2018	Water	J4134-02	COD	Conc H2SO4 tc	KWAR01	M63	STORM-DRAIN-2	07/24/2018	SM5220 D
07/30/2018	Water	J4146-01	COD	Conc H2SO4 tc	PARS02	N33	SES-CS-PH4-COA-01	07/24/2018	SM5220 D
08/07/2018	Water	J4153-02	COD	Conc H2SO4 tc	SHAW22	M33	OUTFALL001-180717	07/17/2018	SM5220 D
08/07/2018	Water	J4153-03	COD	Conc H2SO4 tc	SHAW22	M33	OUTFALL001-180724	07/24/2018	SM5220 D
07/30/2018	Water	J4158-01	COD	Conc H2SO4 tc	PSEG01	K63	OILY-WATER-TOTES	07/25/2018	SM5220 D
07/30/2018	Water	J4167-01	COD	Conc H2SO4 tc	PSEG01	K62	GS-OILY-WATER	07/25/2018	SM5220 D

Date/Time

Received by:

Relinquished by:

07/26/18 9:00

ISB

CS

Date/Time

Received by:

Relinquished by:

07/26/18 10:00

CS

ISB