


ISM02.4

LB 96899
INSTRUMENT ID: CV4

Linear

 μ Abs.:

48568



A= 0.0000e+000

B= 2.0741e-004

C= -7.2318e-002

Rho= 0.9999392

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.036	-0.036	174	0.000	174					-9
0.2	0.200	0.183	-0.017	1233	0.0 %	1233					3
2.5	2.500	2.587	0.087	12824	0.0 %	12824					0
5.0	5.000	4.987	-0.013	24394	0.0 %	24394					0
7.5	7.500	7.477	-0.023	36398	0.0 %	36398					0
10.0	10.000	10.001	0.001	48568	0.0 %	48568					0

NB
7/26/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	174	-		0 ISM02.4	PPB	7/26/2018 14:52	Std
	0.2 S0.2	1233	-		0.2 ISM02.4	PPB	7/26/2018 14:54	Std
	2.5 S2.5	12824	-		2.5 ISM02.4	PPB	7/26/2018 14:57	Std
	5 S5	24394	-		5 ISM02.4	PPB	7/26/2018 14:59	Std
	7.5 S7.5	36398	-		7.5 ISM02.4	PPB	7/26/2018 15:06	Std
	10 S10.0	48568	-		10 ISM02.4	PPB	7/26/2018 15:09	Std
ICV	ICV	19780	4.0302 -		ISM02.4	PPB	7/26/2018 15:11	SMPL
ICB	ICB	308	-0.0084 -		ISM02.4	PPB	7/26/2018 15:14	SMPL
CCV017	CCV017	24549	5.0194 -		ISM02.4	PPB	7/26/2018 15:16	SMPL
CCB017	CCB017	237	-0.0232 -		ISM02.4	PPB	7/26/2018 15:18	SMPL
PB111476BL	PBS004	23	-0.0675 -		ISM02.4	PPB	7/26/2018 15:20	SMPL
J4077-01	MJJ1T5	491	0.0295 -		ISM02.4	PPB	7/26/2018 15:23	SMPL
J4077-02	MJJ1T6	716	0.0762 -		ISM02.4	PPB	7/26/2018 15:25	SMPL
J4077-03	MJJ1T7	991	0.1332 -		ISM02.4	PPB	7/26/2018 15:27	SMPL
J4077-04	MJJ1T8	5123	0.9902 -		ISM02.4	PPB	7/26/2018 15:30	SMPL
J4077-05	MJJ1T9	3940	0.7449 -		ISM02.4	PPB	7/26/2018 15:32	SMPL
J4077-06	MJJ1T9D	3946	0.7461 -		ISM02.4	PPB	7/26/2018 15:38	SMPL
J4077-07	MJJ1T9S	15432	3.1284 -		ISM02.4	PPB	7/26/2018 15:43	SMPL
CCV018	CCV018	24666	5.0436 -		ISM02.4	PPB	7/26/2018 15:45	SMPL
CCB018	CCB018	17	-0.0688 -		ISM02.4	PPB	7/26/2018 15:47	SMPL