



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96900

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/26/2018	13:20
2	CAL2	1	Water			21.7	7.01	07/26/2018	13:22
3	CAL3	1	Water			21.9	10.04	07/26/2018	13:24
4	ICV	1	Water			21.8	7.00	07/26/2018	13:26
5	CCV1	1	Water			21.7	2.00	07/26/2018	13:28
6	J4181-01	1	Water			22.2	4.50	07/26/2018	13:30
7	J4181-01DUP	1	Water			22.3	4.51	07/26/2018	13:32
8	J4182-01	1	Water			22.1	4.48	07/26/2018	13:34
9	CCV2	1	Water			21.9	12.00	07/26/2018	13:36

WORKLIST(Hardcopy Internal Chain)

LB 96900

WorkList Name : pH072618

WorkList ID : 114928

Date : 7/26/2018 12:59:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Water	J4181-01	A pH	Cool 4 deg C	STER16	N12	MAIN-CELL	07/24/2018	SM 4500-PH B
07/24/2018	Water	J4182-01	A pH	Cool 4 deg C	STER16	N12	EX-CELL	07/24/2018	SM 4500-PH B

7:10

Date/Time 07/26/18 7:00
 Received by: SP
 Relinquished by: SP