



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96903

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	07/26/2018	16:40
2	CAL2	1	Water			21.7	7.00	07/26/2018	16:42
3	CAL3	1	Water			22.0	10.04	07/26/2018	16:44
4	ICV	1	Water			21.8	7.01	07/26/2018	16:46
5	CCV1	1	Water			21.6	2.00	07/26/2018	16:48
6	J4176-05	1	Solid	20.02	20	23.2	5.12	07/26/2018	16:50
7	J4176-05DUP	1	Solid	20.04	20	23.3	5.13	07/26/2018	16:52
8	J4180-04	1	Solid	20.02	20	24.4	8.69	07/26/2018	16:54
9	J4180-08	1	Solid	20.03	20	24.2	8.78	07/26/2018	16:56
10	J4189-04	1	Solid	20.02	20	23.2	8.87	07/26/2018	16:58
11	J4189-08	1	Solid	20.02	20	23.9	6.60	07/26/2018	17:00
12	J4189-12	1	Solid	20.03	20	24.0	8.76	07/26/2018	17:02
13	J4189-16	1	Solid	20.02	20	23.6	7.61	07/26/2018	17:04
14	J4189-20	1	Solid	20.04	20	24.0	8.10	07/26/2018	17:06
15	CCV2	1	Water			21.8	12.00	07/26/2018	17:08

WORKLIST(Hardcopy Internal Chain)

LB 96903

WorkList Name : corrosivity072618

WorkList ID : 114934

Date : 7/26/2018 4:09:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/26/2018	Solid	J4176-05	Corrosivity	Cool 4 deg C	PSEG01	K63	PLASTIC-COAT	07/26/2018	9045C
07/26/2018	Solid	J4180-04	Corrosivity	Cool 4 deg C	PSEG01	K63	GS-PIT-1	07/26/2018	9045C
07/26/2018	Solid	J4180-08	Corrosivity	Cool 4 deg C	PSEG01	K63	GS-PIT-2	07/26/2018	9045C
07/26/2018	Solid	J4189-04	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-12-S	07/26/2018	9045C
07/26/2018	Solid	J4189-08	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-12-D	07/26/2018	9045C
07/26/2018	Solid	J4189-12	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-13-S	07/26/2018	9045C
07/26/2018	Solid	J4189-16	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-14-S	07/26/2018	9045C
07/26/2018	Solid	J4189-20	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-14-D	07/26/2018	9045C

Date/Time 6/26/18 4:30

Received by: SP

Relinquished by: CP

Date/Time 07/26/18 5:30

Received by: CP

Relinquished by: SP