


7471B

LB 96905
INSTRUMENT ID: CVJ

 μ Abs.:

45879



Linear

A= 0.0000e+000

B= 2.1556e-004

C= -5.8173e-002

Rho= 0.9993426

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	195	0.000	195				
0.2	0.200	0.228	0.028	1328	0.0 %	1328				
2.5	2.500	2.337	-0.163	11109	0.0 %	11109				
5.0	5.000	5.168	0.168	24244	0.0 %	24244				
7.5	7.500	7.652	0.152	35767	0.0 %	35767				
10.0	10.000	9.832	-0.168	45879	0.0 %	45879				

LB96905 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jul 2018 16:37:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	195	-	0.00007471B	PPB		26 Jul 2018 16:40:32	S	Std
0.2 - 1	50.2	1328	-	0.20007471B	PPB		26 Jul 2018 16:42:50	S	Std
2.5 - 1	52.5	11109	-	2.50007471B	PPB		26 Jul 2018 16:45:07	S	Std
5.0 - 1	55	24244	-	5.00007471B	PPB		26 Jul 2018 16:47:24	S	Std
7.5 - 1	57.5	35767	-	7.50007471B	PPB		26 Jul 2018 16:49:42	S	Std
10.0 - 1	510	45879	-	10.00007471B	PPB		26 Jul 2018 16:54:36	S	Std
ICV73 - 1	ICV73	18862	4.0078	-7471B	PPB		26 Jul 2018 16:57:14	U	SMPL
ICB73 - 1	ICB73	281	0.0024	-7471B	PPB		26 Jul 2018 16:59:30	U	SMPL
CCV19 - 1	CCV19	24127	5.1427	-7471B	PPB		26 Jul 2018 17:01:46	U	SMPL
CCB19 - 1	CCB19	241	-0.0062	-7471B	PPB		26 Jul 2018 17:04:01	U	SMPL
CRA - 1	CRA	1290	0.2199	-7471B	PPB		26 Jul 2018 17:06:18	U	SMPL
HighStd - 1	HighStd	46385	9.9407	-7471B	PPB		26 Jul 2018 17:08:33	U	SMPL
ChkStd - 1	ChkStd	32714	6.9938	-7471B	PPB		26 Jul 2018 17:10:50	U	SMPL
PB111477BL - 1	PBS	27	-0.0524	-7471B	PPB		26 Jul 2018 17:13:08	U	SMPL
PB111477BS - 1	LCSS	18507	3.9313	-7471B	PPB		26 Jul 2018 17:15:25	U	SMPL
J4152-01 - 1	LA-SOIL-1	1861	0.3430	-7471B	PPB		26 Jul 2018 17:17:42	U	SMPL
J4152-05 - 1	LA-SOIL-2	2070	0.3880	-7471B	PPB		26 Jul 2018 17:19:59	U	SMPL
J4164-01 - 1	SB-6(3-5)	8593	1.7942	-7471B	PPB		26 Jul 2018 17:22:16	U	SMPL
J4164-02 - 1	SB-6(7-9)	13844	2.9261	-7471B	PPB		26 Jul 2018 17:24:34	U	SMPL
J4164-02DUP - 1	SB-6(7-9)DUP	15050	3.1861	-7471B	PPB		26 Jul 2018 17:31:01	U	SMPL
CCV20 - 1	CCV20	23931	5.1005	-7471B	PPB		26 Jul 2018 17:33:17	U	SMPL
CCB20 - 1	CCB20	242	-0.0060	-7471B	PPB		26 Jul 2018 17:35:33	U	SMPL
J4164-03 - 1	J4164-02MS	34429	7.3635	-7471B	PPB		26 Jul 2018 17:37:49	U	SMPL
J4164-04 - 1	J4164-02MSD	34050	7.2818	-7471B	PPB		26 Jul 2018 17:44:07	U	SMPL
J4164-05 - 1	SB-5(3-5)	4698	0.9545	-7471B	PPB		26 Jul 2018 17:46:23	U	SMPL
J4164-06 - 1	SB-5(8-10)	16582	3.5163	-7471B	PPB		26 Jul 2018 17:48:41	U	SMPL
J4164-07 - 1	SB-4(3-5)	33356	7.1322	-7471B	PPB		26 Jul 2018 17:50:58	U	SMPL
J4164-08 - 1	SB-4(7-9)	9214	1.9280	-7471B	PPB		26 Jul 2018 17:53:15	U	SMPL
J4164-09 - 1	SB-2(3-5)	52506	11.2602	-7471B	PPB		26 Jul 2018 17:55:34	U	SMPL
J4164-10 - 1	SB-2	179295	38.5913	-7471B	PPB		26 Jul 2018 17:57:51	U	SMPL
J4164-11 - 1	DUP-072518	19204	4.0815	-7471B	PPB		26 Jul 2018 18:00:11	U	SMPL
CCV21 - 1	CCV21	23894	5.0925	-7471B	PPB		26 Jul 2018 18:02:38	U	SMPL
CCB21 - 1	CCB21	160	-0.0237	-7471B	PPB		26 Jul 2018 18:04:55	U	SMPL
J4164-02LX5 - 1		2978	0.5838	-7471B	PPB		26 Jul 2018 18:07:11	U	SMPL
J4164-02A - 1		32259	6.8957	-7471B	PPB		26 Jul 2018 18:09:27	U	SMPL
J4164-09DLX2 - 1		27314	5.8297	-7471B	PPB		26 Jul 2018 18:11:43	U	SMPL
J4164-10DLX10 - 1		19628	4.1729	-7471B	PPB		26 Jul 2018 18:14:01	U	SMPL
CCV22 - 1	CCV22	23593	5.0276	-7471B	PPB		26 Jul 2018 18:16:19	U	SMPL
CCB22 - 1	CCB22	244	-0.0056	-7471B	PPB		26 Jul 2018 18:18:35	U	SMPL